


7470A

LB91972  
INSTRUMENT ID: CX1

Linear

 $\mu$  Abs.:

34933



A= 0.0000e+000

B= 2.8497e-004

C= 1.4695e-002

Rho= 0.9999558

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.015 | 0.015  | 0     | 0.000      | 0     |       |       |       |       |
| 0.2    | 0.200  | 0.210 | 0.010  | 687   | 0.0 %      | 687   |       |       |       |       |
| 2.5    | 2.500  | 2.442 | -0.058 | 8518  | 0.0 %      | 8518  |       |       |       |       |
| 5.0    | 5.000  | 5.014 | 0.014  | 17542 | 0.0 %      | 17542 |       |       |       |       |
| 7.5    | 7.500  | 7.549 | 0.049  | 26440 | 0.0 %      | 26440 |       |       |       |       |
| 10.0   | 10.000 | 9.970 | -0.030 | 34933 | 0.0 %      | 34933 |       |       |       |       |

## LB91972 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 06 Dec 2017 11:11:52

| Sample ID       | Extended ID                | µ Abs | Conc.   | Std ConcMethod | Units | Date                 | Type | Type |
|-----------------|----------------------------|-------|---------|----------------|-------|----------------------|------|------|
| 0.0 - 1         | 50                         | 0     | -       | 0.00007470A    | PPB   | 06 Dec 2017 11:16:43 | S    | Std  |
| 0.2 - 1         | 50.2                       | 687   | -       | 0.20007470A    | PPB   | 06 Dec 2017 11:18:51 | S    | Std  |
| 2.5 - 1         | 52.5                       | 8518  | -       | 2.50007470A    | PPB   | 06 Dec 2017 11:21:00 | S    | Std  |
| 5.0 - 1         | 55                         | 17542 | -       | 5.00007470A    | PPB   | 06 Dec 2017 11:23:08 | S    | Std  |
| 7.5 - 1         | 57.5                       | 26440 | -       | 7.50007470A    | PPB   | 06 Dec 2017 11:25:16 | S    | Std  |
| 10.0 - 1        | 50                         | 34933 | -       | 10.00007470A   | PPB   | 06 Dec 2017 11:27:24 | S    | Std  |
| ICV92 - 1       | ICV92                      | 13612 | 3.8937  | -7470A         | PPB   | 06 Dec 2017 11:29:59 | U    | SMPL |
| ICB92 - 1       | ICB92                      | -114  | -0.0178 | -7470A         | PPB   | 06 Dec 2017 11:32:05 | U    | SMPL |
| CCV60 - 1       | CCV60                      | 16330 | 4.6683  | -7470A         | PPB   | 06 Dec 2017 11:34:12 | U    | SMPL |
| CCB60 - 1       | CCB60                      | -122  | -0.0201 | -7470A         | PPB   | 06 Dec 2017 11:36:18 | U    | SMPL |
| CRA - 1         | CRA                        | 690   | 0.2113  | -7470A         | PPB   | 06 Dec 2017 11:38:26 | U    | SMPL |
| HighStd - 1     | HighStd                    | 35169 | 10.0369 | -7470A         | PPB   | 06 Dec 2017 11:40:32 | U    | SMPL |
| ChkStd - 1      | ChkStd                     | 23680 | 6.7629  | -7470A         | PPB   | 06 Dec 2017 11:42:39 | U    | SMPL |
| PB104754BL - 1  | PBW                        | -171  | -0.0340 | -7470A         | PPB   | 06 Dec 2017 11:44:48 | U    | SMPL |
| PB104754BS - 1  | LCSW                       | 13184 | 3.7718  | -7470A         | PPB   | 06 Dec 2017 11:46:58 | U    | SMPL |
| PB104754TB - 1  | PB104703TB                 | -109  | -0.0164 | -7470A         | PPB   | 06 Dec 2017 11:49:05 | U    | SMPL |
| I6676-02 - 1    | OK-02-120417               | -23   | 0.0081  | -7470A         | PPB   | 06 Dec 2017 11:51:13 | U    | SMPL |
| I6705-03 - 1    | B12(0-10)COMP              | 0     | 0.0147  | -7470A         | PPB   | 06 Dec 2017 11:53:22 | U    | SMPL |
| I6717-05 - 1    | PB-3(0-10)                 | 21    | 0.0207  | -7470A         | PPB   | 06 Dec 2017 11:55:30 | U    | SMPL |
| I6717-06 - 1    | PB-3(10-20)                | -3    | 0.0138  | -7470A         | PPB   | 06 Dec 2017 11:57:37 | U    | SMPL |
| CCV61 - 1       | CCV61                      | 16970 | 4.8507  | -7470A         | PPB   | 06 Dec 2017 11:59:44 | U    | SMPL |
| CCB61 - 1       | CCB61                      | -130  | -0.0224 | -7470A         | PPB   | 06 Dec 2017 12:01:51 | U    | SMPL |
| I6718-03 - 1    | 2041                       | -22   | 0.0084  | -7470A         | PPB   | 06 Dec 2017 12:03:58 | U    | SMPL |
| I6718-06 - 1    | RT-3992                    | 9     | 0.0173  | -7470A         | PPB   | 06 Dec 2017 12:06:05 | U    | SMPL |
| I6718-09 - 1    | RT-4651                    | 55    | 0.0304  | -7470A         | PPB   | 06 Dec 2017 12:08:11 | U    | SMPL |
| I6718-12 - 1    | RT-1500                    | 80    | 0.0375  | -7470A         | PPB   | 06 Dec 2017 12:10:18 | U    | SMPL |
| I6718-15 - 1    | RT-3026                    | 63    | 0.0326  | -7470A         | PPB   | 06 Dec 2017 12:12:26 | U    | SMPL |
| I6718-18 - 1    | RT-2898                    | 66    | 0.0335  | -7470A         | PPB   | 06 Dec 2017 12:14:33 | U    | SMPL |
| I6726-02 - 1    | 256434                     | 54    | 0.0301  | -7470A         | PPB   | 06 Dec 2017 12:16:41 | U    | SMPL |
| I6733-01 - 1    | OILY-SPEEDY-DRY            | 59    | 0.0315  | -7470A         | PPB   | 06 Dec 2017 12:18:49 | U    | SMPL |
| I6734-02 - 1    | Driveway-Paving-1(0-12)    | 76    | 0.0364  | -7470A         | PPB   | 06 Dec 2017 12:20:58 | U    | SMPL |
| I6734-04 - 1    | Driveway-Paving-1(0-12)    | 61    | 0.0321  | -7470A         | PPB   | 06 Dec 2017 12:23:05 | U    | SMPL |
| CCV62 - 1       | CCV62                      | 17259 | 4.9330  | -7470A         | PPB   | 06 Dec 2017 12:25:12 | U    | SMPL |
| CCB62 - 1       | CCB62                      | -54   | -0.0007 | -7470A         | PPB   | 06 Dec 2017 12:27:19 | U    | SMPL |
| I6734-04DUP - 1 | Driveway-Paving-1(0-12)DUP | 100   | 0.0432  | -7470A         | PPB   | 06 Dec 2017 12:29:27 | U    | SMPL |
| I6734-04MS - 1  | Driveway-Paving-1(0-12)MS  | 12803 | 3.6632  | -7470A         | PPB   | 06 Dec 2017 12:31:33 | U    | SMPL |
| I6734-04MSD - 1 | Driveway-Paving-1(0-12)MSD | 12751 | 3.6484  | -7470A         | PPB   | 06 Dec 2017 12:33:40 | U    | SMPL |
| PB104755BL - 1  | PBW                        | -39   | 0.0036  | -7470A         | PPB   | 06 Dec 2017 12:35:48 | U    | SMPL |
| PB104755BS - 1  | LCSW                       | 13124 | 3.7547  | -7470A         | PPB   | 06 Dec 2017 12:37:55 | U    | SMPL |
| PB104755TB - 1  | PB104724TB                 | -15   | 0.0104  | -7470A         | PPB   | 06 Dec 2017 12:40:03 | U    | SMPL |
| 6724-10 - 1     | P001-BC001-01              | 141   | 0.0549  | -7470A         | PPB   | 06 Dec 2017 12:42:11 | U    | SMPL |
| 6724-11 - 1     | P001-BC002-01              | 82    | 0.0381  | -7470A         | PPB   | 06 Dec 2017 12:44:19 | U    | SMPL |
| 6724-12 - 1     | P001-BC003-01              | 121   | 0.0492  | -7470A         | PPB   | 06 Dec 2017 12:46:28 | U    | SMPL |
| 6724-13 - 1     | P001-BC004-01              | 134   | 0.0529  | -7470A         | PPB   | 06 Dec 2017 12:48:35 | U    | SMPL |
| CCV63 - 1       | CCV63                      | 17066 | 4.8780  | -7470A         | PPB   | 06 Dec 2017 12:50:42 | U    | SMPL |
| CCB63 - 1       | CCB63                      | -46   | 0.0016  | -7470A         | PPB   | 06 Dec 2017 12:52:49 | U    | SMPL |
| 6724-14 - 1     | P001-BC005-01              | 136   | 0.0535  | -7470A         | PPB   | 06 Dec 2017 12:54:57 | U    | SMPL |
| 6724-15 - 1     | P001-BC006-01              | 134   | 0.0529  | -7470A         | PPB   | 06 Dec 2017 12:57:04 | U    | SMPL |
| 6724-16 - 1     | P001-BC007-01              | 79    | 0.0372  | -7470A         | PPB   | 06 Dec 2017 12:59:12 | U    | SMPL |
| 6724-17 - 1     | P001-BC008-01              | 73    | 0.0355  | -7470A         | PPB   | 06 Dec 2017 13:01:19 | U    | SMPL |
| 6724-18 - 1     | P001-BC009-01              | 90    | 0.0403  | -7470A         | PPB   | 06 Dec 2017 13:03:27 | U    | SMPL |
| 6725-04 - 1     | P001-TPI1-1218-01          | 105   | 0.0446  | -7470A         | PPB   | 06 Dec 2017 13:05:34 | U    | SMPL |
| 6725-05 - 1     | P001-TPI1-SW01             | 78    | 0.0369  | -7470A         | PPB   | 06 Dec 2017 13:07:43 | U    | SMPL |
| 6725-06 - 1     | P001-TPI1-SW02             | 74    | 0.0358  | -7470A         | PPB   | 06 Dec 2017 13:09:51 | U    | SMPL |
| 6725-07 - 1     | P001-TPI1-SW03             | 84    | 0.0386  | -7470A         | PPB   | 06 Dec 2017 13:11:59 | U    | SMPL |
| 6725-07DUP - 1  | P001-TPI1-SW03DUP          | 76    | 0.0364  | -7470A         | PPB   | 06 Dec 2017 13:14:07 | U    | SMPL |
| CCV64 - 1       | CCV64                      | 17102 | 4.8883  | -7470A         | PPB   | 06 Dec 2017 13:16:14 | U    | SMPL |
| CCB64 - 1       | CCB64                      | -41   | 0.0030  | -7470A         | PPB   | 06 Dec 2017 13:18:21 | U    | SMPL |
| 6725-07MS - 1   | P001-TPI1-SW03MS           | 12815 | 3.6666  | -7470A         | PPB   | 06 Dec 2017 13:20:29 | U    | SMPL |
| 6725-07MSD - 1  | P001-TPI1-SW03MSD          | 12857 | 3.6786  | -7470A         | PPB   | 06 Dec 2017 13:22:36 | U    | SMPL |
| PB104756BL - 1  | PBW                        | -21   | 0.0087  | -7470A         | PPB   | 06 Dec 2017 13:24:43 | U    | SMPL |
| PB104756BS - 1  | LCSW                       | 12903 | 3.6917  | -7470A         | PPB   | 06 Dec 2017 13:26:50 | U    | SMPL |
| 6720-01 - 1     | MW-01                      | 196   | 0.0706  | -7470A         | PPB   | 06 Dec 2017 13:28:58 | U    | SMPL |
| 6720-01DUP - 1  | MW-01DUP                   | 78    | 0.0369  | -7470A         | PPB   | 06 Dec 2017 13:31:05 | U    | SMPL |
| 6720-01MS - 1   | MW-01MS                    | 14043 | 4.0166  | -7470A         | PPB   | 06 Dec 2017 13:33:13 | U    | SMPL |
| 6720-01MSD - 1  | MW-01MSD                   | 13699 | 3.9185  | -7470A         | PPB   | 06 Dec 2017 13:35:21 | U    | SMPL |
| 6720-02 - 1     | MW-10                      | -46   | 0.0016  | -7470A         | PPB   | 06 Dec 2017 13:37:30 | U    | SMPL |
| 6720-03 - 1     | MW-DUPLICATE               | 27    | 0.0224  | -7470A         | PPB   | 06 Dec 2017 13:39:37 | U    | SMPL |
| CCV65 - 1       | CCV65                      | 17116 | 4.8923  | -7470A         | PPB   | 06 Dec 2017 13:41:45 | U    | SMPL |
| CCB65 - 1       | CCB65                      | -38   | 0.0039  | -7470A         | PPB   | 06 Dec 2017 13:43:52 | U    | SMPL |
| 6726-01 - 1     | 256434                     | 5395  | 1.5521  | -7470A         | PPB   | 06 Dec 2017 13:46:01 | U    | SMPL |
| 6742-01 - 1     | TOTES-1-4                  | 7434  | 2.1332  | -7470A         | PPB   | 06 Dec 2017 13:48:07 | U    | SMPL |
| 6734-04LX5 - 1  |                            | 7     | 0.0167  | -7470A         | PPB   | 06 Dec 2017 13:50:15 | U    | SMPL |
| 6734-04A - 1    |                            | 14201 | 4.0616  | -7470A         | PPB   | 06 Dec 2017 13:52:22 | U    | SMPL |
| 6725-07LX5 - 1  |                            | -24   | 0.0079  | -7470A         | PPB   | 06 Dec 2017 13:54:30 | U    | SMPL |
| 6725-07A - 1    |                            | 13674 | 3.9114  | -7470A         | PPB   | 06 Dec 2017 13:56:37 | U    | SMPL |
| 6720-01LX5 - 1  |                            | -30   | 0.0061  | -7470A         | PPB   | 06 Dec 2017 13:58:45 | U    | SMPL |
| 6720-01A - 1    |                            | 14292 | 4.0875  | -7470A         | PPB   | 06 Dec 2017 14:00:53 | U    | SMPL |
| CV66 - 1        | CCV66                      | 17281 | 4.9393  | -7470A         | PPB   | 06 Dec 2017 14:03:02 | U    | SMPL |
| CB66 - 1        | CCB66                      | -60   | -0.0024 | -7470A         | PPB   | 06 Dec 2017 14:05:10 | U    | SMPL |