


245.1

LB91979
INSTRUMENT ID: CV1

Linear

 μ Abs.:

35407



A= 0.0000e+000

B= 2.8148e-004

C= -6.0506e-003

Rho= 0.9999505

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	5	0.000	5				
0.2	0.200	0.200	-0.000	731	0.0 %	731				
2.5	2.500	2.506	0.006	8925	0.0 %	8925				
5.0	5.000	4.966	-0.034	17664	0.0 %	17664				
7.5	7.500	7.573	0.073	26925	0.0 %	26925				
10.0	10.000	9.960	-0.040	35407	0.0 %	35407				

LB91979 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 06 Dec 2017 14:22:28

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	5	-	0.0000	245.1	PPB	06 Dec 2017 14:24:34	S	Std
0.2 - 1	50.2	731	-	0.2000	245.1	PPB	06 Dec 2017 14:26:43	S	Std
2.5 - 1	52.5	8925	-	2.5000	245.1	PPB	06 Dec 2017 14:28:51	S	Std
5.0 - 1	55	17664	-	5.0000	245.1	PPB	06 Dec 2017 14:30:59	S	Std
7.5 - 1	57.5	26925	-	7.5000	245.1	PPB	06 Dec 2017 14:33:07	S	Std
10.0 - 1	510	35407	-	10.0000	245.1	PPB	06 Dec 2017 14:37:50	S	Std
ICV93 - 1	ICV93	13696	3.8490	-245.1		PPB	06 Dec 2017 14:40:23	U	SMPL
ICB93 - 1	ICB93	-106	-0.0359	-245.1		PPB	06 Dec 2017 14:42:29	U	SMPL
CCV67 - 1	CCV67	17685	4.9719	-245.1		PPB	06 Dec 2017 14:44:36	U	SMPL
CCB67 - 1	CCB67	-103	-0.0350	-245.1		PPB	06 Dec 2017 14:46:42	U	SMPL
CRA - 1	CRA	728	0.1989	-245.1		PPB	06 Dec 2017 14:48:49	U	SMPL
HighStd - 1	HighStd	34597	9.7322	-245.1		PPB	06 Dec 2017 14:50:56	U	SMPL
ChkStd - 1	ChkStd	24972	7.0230	-245.1		PPB	06 Dec 2017 14:53:03	U	SMPL
PB104757BL - 1	PBW	-102	-0.0348	-245.1		PPB	06 Dec 2017 14:55:11	U	SMPL
PB104757BS - 1	LCSW	13322	3.7438	-245.1		PPB	06 Dec 2017 14:57:20	U	SMPL
I6731-01 - 1	113017-TIDO-E-D-B	-23	-0.0125	-245.1		PPB	06 Dec 2017 14:59:27	U	SMPL
I6731-02 - 1	113017-TIDO-E-D-S	72	0.0142	-245.1		PPB	06 Dec 2017 15:01:35	U	SMPL
I6731-03 - 1	113017-TIDO-E-U-B	62	0.0114	-245.1		PPB	06 Dec 2017 15:03:43	U	SMPL
I6731-04 - 1	113017-TIDO-E-U-S	47	0.0072	-245.1		PPB	06 Dec 2017 15:05:52	U	SMPL
I6732-01 - 1	120117-TIDO-F-D-B	52	0.0086	-245.1		PPB	06 Dec 2017 15:07:58	U	SMPL
CCV68 - 1	CCV68	17498	4.9192	-245.1		PPB	06 Dec 2017 15:10:05	U	SMPL
CCB68 - 1	CCB68	-95	-0.0328	-245.1		PPB	06 Dec 2017 15:12:12	U	SMPL
I6732-02 - 1	120117-TIDO-F-D-S	45	0.0066	-245.1		PPB	06 Dec 2017 15:14:19	U	SMPL
I6732-03 - 1	120117-TIDO-F-U-B	89	0.0190	-245.1		PPB	06 Dec 2017 15:16:26	U	SMPL
I6732-04 - 1	120117-TIDO-F-U-S	72	0.0142	-245.1		PPB	06 Dec 2017 15:18:33	U	SMPL
I6732-05 - 1	120117-FOPI-E-D-M	81	0.0167	-245.1		PPB	06 Dec 2017 15:20:40	U	SMPL
I6732-06 - 1	120117-FOPI-E-U-M	65	0.0122	-245.1		PPB	06 Dec 2017 15:22:48	U	SMPL
I6732-06DUP - 1	120117-FOPI-E-U-MDUP	59	0.0106	-245.1		PPB	06 Dec 2017 15:24:55	U	SMPL
I6732-06MS - 1	120117-FOPI-E-U-MMS	14865	4.1781	-245.1		PPB	06 Dec 2017 15:27:04	U	SMPL
I6732-06MSD - 1	120117-FOPI-E-U-MMSD	14777	4.1533	-245.1		PPB	06 Dec 2017 15:29:12	U	SMPL
I6732-06LX5 - 1		-67	-0.0249	-245.1		PPB	06 Dec 2017 15:31:20	U	SMPL
I6732-06A - 1		15949	4.4832	-245.1		PPB	06 Dec 2017 15:33:27	U	SMPL
CCV69 - 1	CCV69	17491	4.9172	-245.1		PPB	06 Dec 2017 15:35:34	U	SMPL
CCB69 - 1	CCB69	-107	-0.0362	-245.1		PPB	06 Dec 2017 15:37:41	U	SMPL