


7471A

LB 91985
INSTRUMENT 105 CV1

Linear

 μ Abs.:

4094E



A= 0.0000e+000

B= 2.4324e-004

C= 8.4399e-003

Rho= 0.9998252

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	10	0.000	10				
0.2	0.200	0.221	0.021	874	0.0 %	874				
2.5	2.500	2.384	-0.116	9768	0.0 %	9768				
5.0	5.000	5.117	0.117	21001	0.0 %	21001				
7.5	7.500	7.499	-0.001	30794	0.0 %	30794	0			
10.0	10.000	9.968	-0.032	40946	0.0 %	40946				

LB91985 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 06 Dec 2017 16:26:10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	10	-	0.0000	7471A	PPB	06 Dec 2017 16:29:47	S	Std
0.2 - 1	50.2	874	-	0.2000	7471A	PPB	06 Dec 2017 16:32:05	S	Std
2.5 - 1	52.5	9768	-	2.5000	7471A	PPB	06 Dec 2017 16:34:22	S	Std
5.0 - 1	55	21001	-	5.0000	7471A	PPB	06 Dec 2017 16:36:40	S	Std
7.5 - 1	57.5	30794	-	7.5000	7471A	PPB	06 Dec 2017 16:38:57	S	Std
10.0 - 1	510	40946	-	10.0000	7471A	PPB	06 Dec 2017 16:43:53	S	Std
ICV94 - 1	ICV94	15441	3.7643	-7471A	PPB	06 Dec 2017 16:46:34	U	SMPL	
ICB94 - 1	ICB94	-130	-0.0232	-7471A	PPB	06 Dec 2017 16:48:49	U	SMPL	
CCV70 - 1	CCV70	21044	5.1272	-7471A	PPB	06 Dec 2017 16:51:05	U	SMPL	
CCB70 - 1	CCB70	-137	-0.0249	-7471A	PPB	06 Dec 2017 16:53:21	U	SMPL	
CRA - 1	CRA	843	0.2135	-7471A	PPB	06 Dec 2017 16:55:36	U	SMPL	
HighStd - 1	HighStd	41476	10.0971	-7471A	PPB	06 Dec 2017 16:57:52	U	SMPL	
ChkStd - 1	ChkStd	29922	7.2867	-7471A	PPB	06 Dec 2017 17:00:08	U	SMPL	
PB104758BL - 1	PBS	-143	-0.0263	-7471A	PPB	06 Dec 2017 17:02:27	U	SMPL	
PB104758BS - 1	LCSS	15468	3.7709	-7471A	PPB	06 Dec 2017 17:04:44	U	SMPL	
I6676-01 - 1	OK-02-120417	1646	0.4088	-7471A	PPB	06 Dec 2017 17:07:01	U	SMPL	
I6718-01 - 1	2041	3367	0.8274	-7471A	PPB	06 Dec 2017 17:09:18	U	SMPL	
I6718-04 - 1	RT-3992	3262	0.8019	-7471A	PPB	06 Dec 2017 17:11:36	U	SMPL	
I6718-07 - 1	RT-4651	2817	0.6936	-7471A	PPB	06 Dec 2017 17:13:53	U	SMPL	
I6718-10 - 1	RT-1500	1411	0.3517	-7471A	PPB	06 Dec 2017 17:16:09	U	SMPL	
CCV71 - 1	CCV71	20514	4.9983	-7471A	PPB	06 Dec 2017 17:18:25	U	SMPL	
CCB71 - 1	CCB71	-122	-0.0212	-7471A	PPB	06 Dec 2017 17:20:41	U	SMPL	
I6718-13 - 1	RT-3026	6284	1.5370	-7471A	PPB	06 Dec 2017 17:22:57	U	SMPL	
I6718-16 - 1	RT-2898	1130	0.2833	-7471A	PPB	06 Dec 2017 17:25:13	U	SMPL	
I6721-01 - 1	TOTE-1-2	31	0.0160	-7471A	PPB	06 Dec 2017 17:27:29	U	SMPL	
I6733-01 - 1	OILY-SPEEDY-DRY	957	0.2412	-7471A	PPB	06 Dec 2017 17:29:45	U	SMPL	
I6733-02 - 1	OILY-DEBRIS	64	0.0240	-7471A	PPB	06 Dec 2017 17:32:02	U	SMPL	
I6733-03 - 1	OILY-DRUM	34	0.0167	-7471A	PPB	06 Dec 2017 17:34:19	U	SMPL	
I6733-04 - 1	OIL-TOTES	45	0.0194	-7471A	PPB	06 Dec 2017 17:36:36	U	SMPL	
I6733-05 - 1	LTS-OIL	52	0.0211	-7471A	PPB	06 Dec 2017 17:38:53	U	SMPL	
6734-01 - 1	Driveway-Paving-1(0-12)	3007	0.7399	-7471A	PPB	06 Dec 2017 17:41:11	U	SMPL	
6734-03 - 1	Driveway-Paving-2(0-12)	3060	0.7528	-7471A	PPB	06 Dec 2017 17:43:27	U	SMPL	
CCV72 - 1	CCV72	20754	5.0567	-7471A	PPB	06 Dec 2017 17:45:43	U	SMPL	
CCB72 - 1	CCB72	-111	-0.0186	-7471A	PPB	06 Dec 2017 17:47:59	U	SMPL	
6738-01 - 1	88-20-90-20	2245	0.5545	-7471A	PPB	06 Dec 2017 17:50:16	U	SMPL	
6738-01DUP - 1	88-20-90-20DUP	2004	0.4959	-7471A	PPB	06 Dec 2017 17:52:32	U	SMPL	
6738-01MS - 1	88-20-90-20MS	18354	4.4729	-7471A	PPB	06 Dec 2017 17:54:49	U	SMPL	
6738-01MSD - 1	88-20-90-20MSD	17329	4.2236	-7471A	PPB	06 Dec 2017 17:57:05	U	SMPL	
6741-01 - 1	DRUM-1-7	1038	0.2609	-7471A	PPB	06 Dec 2017 17:59:22	U	SMPL	
6738-01LX5 - 1		392	0.1038	-7471A	PPB	06 Dec 2017 18:01:39	U	SMPL	
6738-01A - 1		18378	4.4787	-7471A	PPB	06 Dec 2017 18:03:56	U	SMPL	
CCV73 - 1	CCV73	19794	4.8231	-7471A	PPB	06 Dec 2017 18:06:14	U	SMPL	
CCB73 - 1	CCB73	-125	-0.0220	-7471A	PPB	06 Dec 2017 18:08:32	U	SMPL	