



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: apatel

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: AMANDEEP

Run Number: LB91993

Reagent/Standard	Lot/Log #
Residual chlorine std, 0.8PPM	WP60164
Residual chlorine std, 0PPM	WP60159
Residual chlorine std, 0.1PPM	WP60161
Residual chlorine std, 0.2PPM	WP60162
Residual chlorine std, and CCV , 0.4PPM	WP60163
Residual chlorine std, 1.6PPM	WP60165
Residual chlorine ICV-LCS, 0.4PPM	WP60166
Residual chlorine LOD, 0.05PPM	WP60160
Total Chlorine Powder Pillows	W1959

Intercept: 0.0066

Slope: 0.9776

Regression: 0.999814

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	12/06/2017	15:40
2	CAL2	0.1	1	0.000	0.100	0.100	12/06/2017	15:44
3	CAL3	0.2	1	0.000	0.200	0.200	12/06/2017	15:48
4	CAL4	0.4	1	0.000	0.400	0.400	12/06/2017	15:52
5	CAL5	0.8	1	0.000	0.810	0.810	12/06/2017	15:57
6	CAL6	1.6	1	0.000	1.560	1.560	12/06/2017	16:02



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: apatel

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: AMANDEEP

Run Number: LB91993

Seq	Lab ID	Initial Weight	Final Vol	True Value	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.000	0.400	0.400	0.402	12/06/2017	16:07
2	ICB				1	0.000	0.000	0.000	-0.007	12/06/2017	16:11
3	CCV1			0.4	1	0.000	0.400	0.400	0.402	12/06/2017	16:14
4	CCB1				1	0.000	0.000	0.000	-0.007	12/06/2017	16:18
5	LB91993BL	3	3		1	0.000	0.000	0.000	-0.007	12/06/2017	16:22
6	LB91993BS	3	3	0.4	1	0.000	0.380	0.380	0.382	12/06/2017	16:26
7	I6767-04	3	3		1	0.000	0.080	0.080	0.075	12/06/2017	16:32
8	I6767-05	3	3		1	0.000	0.010	0.010	0.003	12/06/2017	16:37
9	CCV2			0.4	1	0.000	0.410	0.410	0.413	12/06/2017	16:41
10	CCB2				1	0.000	0.000	0.000	-0.007	12/06/2017	16:45

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ResCL120617

WorkList ID : 106210

Date : 12/6/2017 3:11:01 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/06/2017	Water	I6767-04	Residual Chlorine	Cool 4 deg C	THEL01	P11	FLOOR	12/06/2017	SM4500 Cl G
12/06/2017	Water	I6767-05	Residual Chlorine	Cool 4 deg C	THEL01	P11	TAP-WATER	12/06/2017	SM4500 Cl G

Date/Time 12/6/17 3:15 PM

Received by: [Signature]

Relinquished by: [Signature]

Date/Time 12/6/17 3:00 PM

Received by: [Signature]

Relinquished by: [Signature]