

7471A


LB91998

INSTRUMENT ID: CV2

Linear

 μ Abs.:

43803



A= 0.0000e+000

B= 2.3127e-004

C= 5.2130e-002

Rho= 0.9994763

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.059	0.059	31	0.000	31				
0.2	0.200	0.243	0.043	826	0.0 %	826				
2.5	2.500	2.498	-0.002	10577	0.0 %	10577				
5.0	5.000	4.916	-0.084	21032	0.0 %	21032				
7.5	7.500	7.300	-0.200	31341	0.0 %	31341	0			
10.0	10.000	10.183	0.183	43803	0.0 %	43803				

LB91998 INSTRUMENT ID : CV1

Method: 7471A		Operator: Admin		Date of Analysis: 07 Dec 2017 08:19:04				
Sample ID	Extended ID	µ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50	31	-	0.00007471A	PPB	07 Dec 2017 08:31:28	S	Std
0.2 - 1	50.2	826	-	0.20007471A	PPB	07 Dec 2017 08:33:46	S	Std
2.5 - 1	52.5	10577	-	2.50007471A	PPB	07 Dec 2017 08:36:04	S	Std
5.0 - 1	55	21032	-	5.00007471A	PPB	07 Dec 2017 08:38:21	S	Std
7.5 - 1	57.5	31341	-	7.50007471A	PPB	07 Dec 2017 08:40:38	S	Std
10.0 - 1	510	43803	-	10.00007471A	PPB	07 Dec 2017 08:45:29	S	Std
ICV95 - 1	ICV95	15973	3.7462	-7471A	PPB	07 Dec 2017 08:49:48	U	SMPL
ICB95 - 1	ICB95	-87	0.0320	-7471A	PPB	07 Dec 2017 08:52:04	U	SMPL
CCV74 - 1	CCV74	21284	4.9745	-7471A	PPB	07 Dec 2017 08:54:19	U	SMPL
CCB74 - 1	CCB74	-118	0.0248	-7471A	PPB	07 Dec 2017 08:56:36	U	SMPL
CRA - 1	CRA	816	0.2408	-7471A	PPB	07 Dec 2017 08:58:53	U	SMPL
HighStd - 1	HighStd	44330	10.3044	-7471A	PPB	07 Dec 2017 09:01:08	U	SMPL
ChkStd - 1	ChkStd	28313	6.6001	-7471A	PPB	07 Dec 2017 09:03:25	U	SMPL
PB104793BL - 1	PBS	-188	0.0087	-7471A	PPB	07 Dec 2017 09:05:43	U	SMPL
PB104793BS - 1	LCSS	15718	3.6873	-7471A	PPB	07 Dec 2017 09:10:33	U	SMPL
I6303-01 - 1	HD-02-120417	2787	0.6967	-7471A	PPB	07 Dec 2017 09:12:50	U	SMPL
I6675-01 - 1	SU-04-120517	2400	0.6072	-7471A	PPB	07 Dec 2017 09:15:07	U	SMPL
I6683-01 - 1	EO-01-120517-A	1668	0.4379	-7471A	PPB	07 Dec 2017 09:17:25	U	SMPL
I6683-03 - 1	EO-01-120517-B	4638	1.1248	-7471A	PPB	07 Dec 2017 09:19:43	U	SMPL
CCV75 - 1	CCV75	21268	4.9708	-7471A	PPB	07 Dec 2017 09:24:15	U	SMPL
CCB75 - 1	CCB75	-121	0.0241	-7471A	PPB	07 Dec 2017 09:26:31	U	SMPL
I6683-03MS - 1	EO-01-120517-BMS	20294	4.7456	-7471A	PPB	07 Dec 2017 09:28:48	U	SMPL
I6683-03MSD - 1	EO-01-120517-BMSD	19268	4.5083	-7471A	PPB	07 Dec 2017 09:31:04	U	SMPL
I6683-05 - 1	EO-01-120517-C	14983	3.5173	-7471A	PPB	07 Dec 2017 09:33:21	U	SMPL
I6685-01 - 1	AU-06-120517-A	4952	1.1974	-7471A	PPB	07 Dec 2017 09:35:38	U	SMPL
I6685-03 - 1	AU-06-120517-B	9271	2.1963	-7471A	PPB	07 Dec 2017 09:37:54	U	SMPL
I6744-01 - 1	PROS-3-E(0-1)	15383	3.6098	-7471A	PPB	07 Dec 2017 09:40:11	U	SMPL
I6744-02 - 1	PROS-3-E(4-5)	7044	1.6812	-7471A	PPB	07 Dec 2017 09:42:29	U	SMPL
I6744-03 - 1	PROS-3-E(7-8)	3970	0.9703	-7471A	PPB	07 Dec 2017 09:44:46	U	SMPL
I6744-04 - 1	PROS-3-E(11-12)	140	0.0845	-7471A	PPB	07 Dec 2017 09:47:04	U	SMPL
I6744-06 - 1	PROS-2-E(0-1)	10624	2.5092	-7471A	PPB	07 Dec 2017 09:49:20	U	SMPL
CCV76 - 1	CCV76	21428	5.0078	-7471A	PPB	07 Dec 2017 09:51:36	U	SMPL
CCB76 - 1	CCB76	-119	0.0246	-7471A	PPB	07 Dec 2017 09:53:52	U	SMPL
I6744-07 - 1	PROS-2-E(4-5)	277	0.1162	-7471A	PPB	07 Dec 2017 09:56:10	U	SMPL
I6744-08 - 1	PROS-2-E(5-6)	1451	0.3877	-7471A	PPB	07 Dec 2017 09:58:26	U	SMPL
I6744-09 - 1	PROS-1-E(0-1)	4147	1.0112	-7471A	PPB	07 Dec 2017 10:00:43	U	SMPL
I6744-10 - 1	PROS-1-E(3-4)	7115	1.6976	-7471A	PPB	07 Dec 2017 10:02:59	U	SMPL
I6747-01 - 1	EPE-106-6(0-5)	2124	0.5434	-7471A	PPB	07 Dec 2017 10:05:16	U	SMPL
I6747-02 - 1	ENV-109-2(0-5)DUP	2788	0.6969	-7471A	PPB	07 Dec 2017 10:07:33	U	SMPL
I6747-03 - 1	ENV-103-1(0-5)	14967	3.5136	-7471A	PPB	07 Dec 2017 10:09:50	U	SMPL
I6747-04 - 1	ENV-103-1(0-5)DUP	12747	3.0002	-7471A	PPB	07 Dec 2017 10:12:08	U	SMPL
PB104794BL - 1	PBS	-63	0.0376	-7471A	PPB	07 Dec 2017 10:14:26	U	SMPL
PB104794BS - 1	LCSS	14395	3.3813	-7471A	PPB	07 Dec 2017 10:16:42	U	SMPL
CCV77 - 1	CCV77	21149	4.9433	-7471A	PPB	07 Dec 2017 10:18:58	U	SMPL
CCB77 - 1	CCB77	-142	0.0193	-7471A	PPB	07 Dec 2017 10:21:15	U	SMPL
I6676-03 - 1	OK-03-120617	766	0.2293	-7471A	PPB	07 Dec 2017 10:23:32	U	SMPL
I6680-01 - 1	CL-01-120617-A	1056	0.2964	-7471A	PPB	07 Dec 2017 10:25:48	U	SMPL
I6680-03 - 1	CL-01-120617-B	762	0.2284	-7471A	PPB	07 Dec 2017 10:28:05	U	SMPL
I6680-05 - 1	CL-01-120617-C	1629	0.4289	-7471A	PPB	07 Dec 2017 10:30:21	U	SMPL
I6680-07 - 1	CL-01-120617-D	940	0.2695	-7471A	PPB	07 Dec 2017 10:32:38	U	SMPL
I6680-09 - 1	CL-01-120617-E	1416	0.3796	-7471A	PPB	07 Dec 2017 10:34:55	U	SMPL
I6684-01 - 1	BU-01-120617	4346	1.0572	-7471A	PPB	07 Dec 2017 10:37:12	U	SMPL
I6686-01 - 1	TR-05-120617	1299	0.3526	-7471A	PPB	07 Dec 2017 10:39:29	U	SMPL
I6746-01 - 1	ENV-109-4(0-5)	465	0.1597	-7471A	PPB	07 Dec 2017 10:41:47	U	SMPL
I6746-01DUP - 1	ENV-109-4(0-5)DUP	377	0.1393	-7471A	PPB	07 Dec 2017 10:44:04	U	SMPL
CCV78 - 1	CCV78	21375	4.9956	-7471A	PPB	07 Dec 2017 10:46:20	U	SMPL
CCB78 - 1	CCB78	-130	0.0221	-7471A	PPB	07 Dec 2017 10:48:37	U	SMPL
I6746-01MS - 1	ENV-109-4(0-5)MS	17296	4.0522	-7471A	PPB	07 Dec 2017 10:50:53	U	SMPL
I6746-01MSD - 1	ENV-109-4(0-5)MSD	16142	3.7853	-7471A	PPB	07 Dec 2017 10:53:09	U	SMPL
I6746-02 - 1	ENV-109-4(5-10)	1040	0.2927	-7471A	PPB	07 Dec 2017 10:55:26	U	SMPL
I6746-03 - 1	ENV-109-4(10-15)	136	0.0836	-7471A	PPB	07 Dec 2017 10:57:42	U	SMPL
I6746-04 - 1	ENV-109-4(15-20)	85	0.0718	-7471A	PPB	07 Dec 2017 10:59:59	U	SMPL
I6746-05 - 1	ENV-109-4(20-30)	97	0.0746	-7471A	PPB	07 Dec 2017 11:02:15	U	SMPL
I6746-06 - 1	ENV-109-4(30-40)	106	0.0766	-7471A	PPB	07 Dec 2017 11:04:33	U	SMPL
I6746-07 - 1	ENV-109-4(40-50)	92	0.0734	-7471A	PPB	07 Dec 2017 11:06:50	U	SMPL
I6746-08 - 1	ENV-109-4(50-60)	111	0.0778	-7471A	PPB	07 Dec 2017 11:09:08	U	SMPL
I6746-09 - 1	ENV-109-7(0-5)	1631	0.4293	-7471A	PPB	07 Dec 2017 11:11:25	U	SMPL
CCV79 - 1	CCV79	20993	4.9072	-7471A	PPB	07 Dec 2017 11:13:42	U	SMPL
CCB79 - 1	CCB79	-103	0.0283	-7471A	PPB	07 Dec 2017 11:15:59	U	SMPL
I6746-10 - 1	ENV-109-2(0-5)	326	0.1275	-7471A	PPB	07 Dec 2017 11:18:16	U	SMPL
PB104795BL - 1	PBS	4	0.0531	-7471A	PPB	07 Dec 2017 11:20:33	U	SMPL
PB104795BS - 1	LCSS	14998	3.5208	-7471A	PPB	07 Dec 2017 11:22:50	U	SMPL
I6765-01 - 1	SOIL-SB001A	1864	0.4832	-7471A	PPB	07 Dec 2017 11:25:07	U	SMPL
I6765-02 - 1	SOIL-SB001B	1542	0.4088	-7471A	PPB	07 Dec 2017 11:27:23	U	SMPL
I6765-03 - 1	SOIL-SB002A	2013	0.5177	-7471A	PPB	07 Dec 2017 11:29:40	U	SMPL
I6765-04 - 1	SOIL-SB002B	959	0.2739	-7471A	PPB	07 Dec 2017 11:31:58	U	SMPL
I6765-05 - 1	SOIL-SB003A	10085	2.3845	-7471A	PPB	07 Dec 2017 11:34:15	U	SMPL
I6765-06 - 1	SOIL-SB003B	1116	0.3102	-7471A	PPB	07 Dec 2017 11:36:33	U	SMPL
I6765-07 - 1	SOIL-SB004A	7232	1.7247	-7471A	PPB	07 Dec 2017 11:38:49	U	SMPL
CCV80 - 1	CCV80	20768	4.8552	-7471A	PPB	07 Dec 2017 11:41:06	U	SMPL
CCB80 - 1	CCB80	-137	0.0204	-7471A	PPB	07 Dec 2017 11:43:23	U	SMPL

LB91998 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 07 Dec 2017 08:19:04

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
I6765-08 - 1	SOIL-SB004B	1349	0.3641		-7471A	PPB	07 Dec 2017 11:45:41	U	SMPL
I6765-09 - 1	SOIL-SB005A	10823	2.5552		-7471A	PPB	07 Dec 2017 11:47:58	U	SMPL
I6765-10 - 1	SOIL-SB005B	1191	0.3276		-7471A	PPB	07 Dec 2017 11:50:15	U	SMPL
I6765-11 - 1	SOIL-SB006A	1555	0.4118		-7471A	PPB	07 Dec 2017 11:52:32	U	SMPL
I6765-12 - 1	SOIL-SB006B	1470	0.3921		-7471A	PPB	07 Dec 2017 11:54:49	U	SMPL
I6765-13 - 1	SOIL-SB007A	2652	0.6655		-7471A	PPB	07 Dec 2017 11:57:06	U	SMPL
I6765-14 - 1	SOIL-SB007B	1176	0.3241		-7471A	PPB	07 Dec 2017 11:59:23	U	SMPL
I6765-15 - 1	SOIL-SB008A	2339	0.5931		-7471A	PPB	07 Dec 2017 12:01:41	U	SMPL
I6765-15DUP - 1	SOIL-SB008ADUP	2468	0.6229		-7471A	PPB	07 Dec 2017 12:03:59	U	SMPL
I6765-15MS - 1	SOIL-SB008AMS	17653	4.1348		-7471A	PPB	07 Dec 2017 12:06:16	U	SMPL
CCV81 - 1	CCV81	20758	4.8529		-7471A	PPB	07 Dec 2017 12:08:33	U	SMPL
CCB81 - 1	CCB81	-137	0.0204		-7471A	PPB	07 Dec 2017 12:10:49	U	SMPL
I6765-15MSD - 1	SOIL-SB008AMSD	17110	4.0092		-7471A	PPB	07 Dec 2017 12:13:07	U	SMPL
I6765-16 - 1	SOIL-SB008B	1298	0.3523		-7471A	PPB	07 Dec 2017 12:15:24	U	SMPL
I6765-17 - 1	SOIL-SB009A	4733	1.1467		-7471A	PPB	07 Dec 2017 12:17:41	U	SMPL
I6765-18 - 1	SOIL-SB009B	1354	0.3653		-7471A	PPB	07 Dec 2017 12:19:58	U	SMPL
I6765-19 - 1	SOIL-SB010A	2613	0.6564		-7471A	PPB	07 Dec 2017 12:22:15	U	SMPL
I6765-20 - 1	SOIL-SB010B	1735	0.4534		-7471A	PPB	07 Dec 2017 12:24:32	U	SMPL
I6765-21 - 1	DUP-CHA001	6024	1.4453		-7471A	PPB	07 Dec 2017 12:26:49	U	SMPL
I6683-03DUP - 1	EO-01-120517-BDUP	4638	1.1248		-7471A	PPB	07 Dec 2017 12:29:07	U	SMPL
CCV82 - 1	CCV82	21464	5.0162		-7471A	PPB	07 Dec 2017 12:31:25	U	SMPL
CCB82 - 1	CCB82	-124	0.0235		-7471A	PPB	07 Dec 2017 12:33:42	U	SMPL
I6683-03LX5 - 1		880	0.2556		-7471A	PPB	07 Dec 2017 12:35:59	U	SMPL
I6683-03A - 1		21342	4.9879		-7471A	PPB	07 Dec 2017 12:38:16	U	SMPL
I6746-01LX5 - 1		-30	0.0452		-7471A	PPB	07 Dec 2017 12:40:33	U	SMPL
I6746-01A - 1		17396	4.0753		-7471A	PPB	07 Dec 2017 12:42:51	U	SMPL
I6765-15LX5 - 1		364	0.1363		-7471A	PPB	07 Dec 2017 12:45:08	U	SMPL
I6765-15A - 1		18708	4.3788		-7471A	PPB	07 Dec 2017 12:47:25	U	SMPL
CCV83 - 1	CCV83	20937	4.8943		-7471A	PPB	07 Dec 2017 12:49:42	U	SMPL
CCB83 - 1	CCB83	-125	0.0232		-7471A	PPB	07 Dec 2017 12:51:59	U	SMPL