


7470A

LB92006
INSTRUMENT ID: CY2

Linear

 μ Abs.:

35685



A= 0.0000e+000

B= 2.8350e-004

C= -1.7311e-003

Rho= 0.9996851

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.010	0.010	40	0.000	40				
0.2	0.200	0.204	0.004	726	0.0 %	726				
2.5	2.500	2.533	0.033	8941	0.0 %	8941				
5.0	5.000	5.029	0.029	17744	0.0 %	17744				
7.5	7.500	7.310	-0.190	25790	0.0 %	25790				
10.0	10.000	10.115	0.115	35685	0.0 %	35685				

LB92006 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 07 Dec 2017 14:39:00

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	40	-	0.0000	7470A	PPB	07 Dec 2017 14:42:06	S	Std
0.2 - 1	502	726	-	0.2000	7470A	PPB	07 Dec 2017 14:44:14	S	Std
2.5 - 1	52.5	8941	-	2.5000	7470A	PPB	07 Dec 2017 14:46:22	S	Std
5.0 - 1	55	17744	-	5.0000	7470A	PPB	07 Dec 2017 14:48:31	S	Std
7.5 - 1	57.5	25790	-	7.5000	7470A	PPB	07 Dec 2017 14:50:38	S	Std
10.0 - 1	510	35685	-	10.0000	7470A	PPB	07 Dec 2017 14:55:13	S	Std
ICV96 - 1	ICV96	14277	4.0458	-	7470A	PPB	07 Dec 2017 14:57:41	U	SMPL
ICB96 - 1	ICB96	-46	-0.0148	-	7470A	PPB	07 Dec 2017 14:59:48	U	SMPL
CCV84 - 1	CCV84	17723	5.0227	-	7470A	PPB	07 Dec 2017 15:01:54	U	SMPL
CCB84 - 1	CCB84	-68	-0.0210	-	7470A	PPB	07 Dec 2017 15:04:01	U	SMPL
CRA - 1	CRA	715	0.2010	-	7470A	PPB	07 Dec 2017 15:06:07	U	SMPL
HighStd - 1	HighStd	35468	10.0534	-	7470A	PPB	07 Dec 2017 15:08:14	U	SMPL
ChkStd - 1	ChkStd	24006	6.8039	-	7470A	PPB	07 Dec 2017 15:10:21	U	SMPL
PB104790BL - 1	PBW	-128	-0.0380	-	7470A	PPB	07 Dec 2017 15:12:29	U	SMPL
PB104790BS - 1	LCSW	13452	3.8119	-	7470A	PPB	07 Dec 2017 15:14:38	U	SMPL
PB104790TB - 1	PB104752TB	-62	-0.0193	-	7470A	PPB	07 Dec 2017 15:16:46	U	SMPL
I6303-02 - 1	HD-02-120417	37	0.0088	-	7470A	PPB	07 Dec 2017 15:18:54	U	SMPL
I6675-02 - 1	SU-04-120517	34	0.0079	-	7470A	PPB	07 Dec 2017 15:21:03	U	SMPL
I6683-02 - 1	EO-01-120517-A	43	0.0105	-	7470A	PPB	07 Dec 2017 15:23:11	U	SMPL
I6683-04 - 1	EO-01-120517-B	29	0.0065	-	7470A	PPB	07 Dec 2017 15:25:18	U	SMPL
CCV85 - 1	CCV85	17943	5.0851	-	7470A	PPB	07 Dec 2017 15:27:25	U	SMPL
CCB85 - 1	CCB85	-40	-0.0131	-	7470A	PPB	07 Dec 2017 15:29:31	U	SMPL
I6683-06 - 1	EO-01-120517-C	28	0.0062	-	7470A	PPB	07 Dec 2017 15:31:39	U	SMPL
I6685-02 - 1	AU-06-120517-A	19	0.0037	-	7470A	PPB	07 Dec 2017 15:33:46	U	SMPL
I6685-04 - 1	AU-06-120517-B	28	0.0062	-	7470A	PPB	07 Dec 2017 15:35:53	U	SMPL
I6733-12 - 1	OILY-SPEEDY-DRY	60	0.0153	-	7470A	PPB	07 Dec 2017 15:38:00	U	SMPL
I6738-03 - 1	88-20-90-20	33	0.0076	-	7470A	PPB	07 Dec 2017 15:40:07	U	SMPL
I6741-02 - 1	DRUM-1-7	32	0.0073	-	7470A	PPB	07 Dec 2017 15:42:15	U	SMPL
I6742-02 - 1	TOTES-1-4	27	0.0059	-	7470A	PPB	07 Dec 2017 15:44:24	U	SMPL
I6742-02DUP - 1	TOTES-1-4DUP	33	0.0076	-	7470A	PPB	07 Dec 2017 15:46:32	U	SMPL
I6742-02MS - 1	TOTES-1-4MS	12936	3.6656	-	7470A	PPB	07 Dec 2017 15:48:40	U	SMPL
I6742-02MSD - 1	TOTES-1-4MSD	13229	3.7487	-	7470A	PPB	07 Dec 2017 15:50:47	U	SMPL
CCV86 - 1	CCV86	17816	5.0491	-	7470A	PPB	07 Dec 2017 15:52:54	U	SMPL
CCB86 - 1	CCB86	-69	-0.0213	-	7470A	PPB	07 Dec 2017 15:55:01	U	SMPL
I6742-02LX5 - 1		33	0.0076	-	7470A	PPB	07 Dec 2017 15:57:08	U	SMPL
I6742-02A - 1		14082	3.9905	-	7470A	PPB	07 Dec 2017 15:59:15	U	SMPL
CCV87 - 1	CCV87	17672	5.0083	-	7470A	PPB	07 Dec 2017 16:01:22	U	SMPL
CCB87 - 1	CCB87	-75	-0.0230	-	7470A	PPB	07 Dec 2017 16:03:29	U	SMPL