



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB92017

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP60075
COD calibration std. 100 ppm	WP60072
COD calibration std. 10 ppm	WP60071
COD calibration std. 150 ppm	WP60073
COD calibration std. 50 ppm	WP60074
COD calibration std. 0 ppm	WP60070
COD Digestion Vials Low Level 0-150Mg/L	W2291
COD ICV-LCS std, 50ppm	WP60201
COD CCV std, 50ppm	WP60202

Temp In (C): <u>150</u>	Date In: <u>12/07/2017</u>	Time In: <u>11:00</u>
Temp Out (C): <u>150</u>	Date Out: <u>12/07/2017</u>	Time Out: <u>13:00</u>

Intercept: -1.029

Slope: 1.0231

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	12/05/2017	12:27
2	CAL2	10	1	Water	9.000	12/05/2017	12:27
3	CAL3	50	1	Water	49.000	12/05/2017	12:28
4	CAL4	100	1	Water	101.000	12/05/2017	12:28
5	CAL5	150	1	Water	153.000	12/05/2017	12:29



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB92017

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.877	12/05/2017	12:29
2	ICB				1	Water	0.000	1.006	12/05/2017	12:30
3	CCV1	50			1	Water	51.000	50.854	12/07/2017	14:10
4	CCB1				1	Water	0.000	1.006	12/07/2017	14:10
5	LB92017BL				1	Water	0.000	1.006	12/07/2017	14:11
6	LB92017BS	50			1	Water	49.000	48.899	12/07/2017	14:11
7	I6742-01				10	Water	70.000	69.425	12/07/2017	14:12
8	I6757-02				1	Water	2.000	2.961	12/07/2017	14:12
9	I6757-02DUP				1	Water	2.000	2.961	12/07/2017	14:13
10	I6757-02MS	50			1	Water	52.000	51.832	12/07/2017	14:13
11	I6757-02MSD	50			1	Water	53.000	52.809	12/07/2017	14:14
12	CCV2	50			1	Water	51.000	50.854	12/07/2017	14:14
13	CCB2				1	Water	0.000	1.006	12/07/2017	14:15