


ISM02.4

LB92018
INSTRUMENT ID: LV2

Linear

 μ Abs.:

40224



A= 0.0000e+000

B= 2.4944e-004

C= -8.0372e-003

Rho= 0.9998662

Accept=Accepted

slope
 $y = 10125 \text{ conc}$

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0 D
0.00	0.000	0.002	0.002	39	0.000	39					4
0.2	0.200	0.209	0.009	869	0.0 %	869					+3
2.5	2.500	2.565	0.065	10317	0.0 %	10317					-3
5.0	5.000	4.873	-0.127	19567	0.0 %	19567					0
7.5	7.500	7.526	0.026	30203	0.0 %	30203					0
10.0	10.000	10.025	0.025	40224	0.0 %	40224					0

103
12/7/17

LB92018
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	39	-		0 ISM02.4	PPB	12/7/2017 17:15	Std
	0.2 S0.2	869	-		0.2 ISM02.4	PPB	12/7/2017 17:17	Std
	2.5 S2.5	10317	-		2.5 ISM02.4	PPB	12/7/2017 17:21	Std
	5 S5	19567	-		5 ISM02.4	PPB	12/7/2017 17:23	Std
	7.5 S7.5	30203	-		7.5 ISM02.4	PPB	12/7/2017 17:25	Std
	10 S10	40224	-		10 ISM02.4	PPB	12/7/2017 17:28	Std
ICV	ICV	15935	3.9668 -		ISM02.4	PPB	12/7/2017 17:31	SMPL
ICB	ICB	-81	-0.0282 -		ISM02.4	PPB	12/7/2017 17:33	SMPL
CCV088	CCV088	19688	4.903 -		ISM02.4	PPB	12/7/2017 17:35	SMPL
CCB088	CCB088	-76	-0.027 -		ISM02.4	PPB	12/7/2017 17:37	SMPL
PB104819BL	PBW003	11	-0.0053 -		ISM02.4	PPB	12/7/2017 17:40	SMPL
16694-18	MA4YX4	-9	-0.0103 -		ISM02.4	PPB	12/7/2017 17:42	SMPL
16694-19	MA4YY0	79	0.0117 -		ISM02.4	PPB	12/7/2017 17:44	SMPL
16694-20	MA4YY1	37	0.0012 -		ISM02.4	PPB	12/7/2017 17:46	SMPL
16694-21	MA4YY2	59	0.0067 -		ISM02.4	PPB	12/7/2017 17:49	SMPL
16694-22	MA4YY3	149	0.0291 -		ISM02.4	PPB	12/7/2017 17:51	SMPL
16694-23	MA4YY4	40	0.0019 -		ISM02.4	PPB	12/7/2017 17:53	SMPL
16694-24	MA4YY5	93	0.0152 -		ISM02.4	PPB	12/7/2017 17:56	SMPL
16694-25	MA4Z09	25	-0.0018 -		ISM02.4	PPB	12/7/2017 17:58	SMPL
16694-26	MA4Z10	45	0.0032 -		ISM02.4	PPB	12/7/2017 18:00	SMPL
16694-27	MA4Z11	199	0.0416 -		ISM02.4	PPB	12/7/2017 18:02	SMPL
16694-28	MA4Z12	166	0.0334 -		ISM02.4	PPB	12/7/2017 18:05	SMPL
16694-29	MA4Z12D	211	0.0446 -		ISM02.4	PPB	12/7/2017 18:07	SMPL
16694-30	MA4Z12S	4415	1.0932 -		ISM02.4	PPB	12/7/2017 18:09	SMPL
16694-31	MA4Z13	29	-0.0008 -		ISM02.4	PPB	12/7/2017 18:11	SMPL
16694-32	MA4Z19	3829	0.9471 -		ISM02.4	PPB	12/7/2017 18:14	SMPL
16694-35	MA4Z37	61	0.0072 -		ISM02.4	PPB	12/7/2017 18:16	SMPL
16694-36	MA4Z38	148	0.0289 -		ISM02.4	PPB	12/7/2017 18:18	SMPL
16694-37	MA4Z57	104	0.0179 -		ISM02.4	PPB	12/7/2017 18:21	SMPL
16694-38	MA4Z58	151	0.0296 -		ISM02.4	PPB	12/7/2017 18:23	SMPL
CCV089	CCV089	19484	4.8521 -		ISM02.4	PPB	12/7/2017 18:25	SMPL
CCB089	CCB089	-71	-0.0257 -		ISM02.4	PPB	12/7/2017 18:27	SMPL
PB104820BL	PBW003	88	0.0139 -		ISM02.4	PPB	12/7/2017 18:30	SMPL
16652-01	MC0286	-9	-0.0103 -		ISM02.4	PPB	12/7/2017 18:32	SMPL
16652-02	MC0287	-2	-0.0085 -		ISM02.4	PPB	12/7/2017 18:34	SMPL
16695-01	MA4YY6	154	0.0304 -		ISM02.4	PPB	12/7/2017 18:37	SMPL
16695-02	MA4YY7	73	0.0102 -		ISM02.4	PPB	12/7/2017 18:39	SMPL
16695-03	MA4YY8	62	0.0074 -		ISM02.4	PPB	12/7/2017 18:41	SMPL
16695-04	MA4YY9	149	0.0291 -		ISM02.4	PPB	12/7/2017 18:43	SMPL
16695-05	MA4YZ0	109	0.0192 -		ISM02.4	PPB	12/7/2017 18:46	SMPL
16695-06	MA4YZ1	77	0.0112 -		ISM02.4	PPB	12/7/2017 18:48	SMPL
16695-07	MA4Z14	52	0.0049 -		ISM02.4	PPB	12/7/2017 18:50	SMPL
16695-08	MA4Z15	64	0.0079 -		ISM02.4	PPB	12/7/2017 18:53	SMPL
16695-09	MA4Z16	130	0.0244 -		ISM02.4	PPB	12/7/2017 18:55	SMPL
16695-10	MA4Z17	121	0.0221 -		ISM02.4	PPB	12/7/2017 18:57	SMPL
16695-11	MA4Z17D	65	0.0082 -		ISM02.4	PPB	12/7/2017 18:59	SMPL

LB92018
INSTRUMENT ID : CV1

I6695-12	MA4Z17S	4239	1.0493 -	ISM02.4	PPB	12/7/2017 19:02 SMPL
I6695-13	MA4Z18	108	0.0189 -	ISM02.4	PPB	12/7/2017 19:04 SMPL
I6695-14	MA4Z39	71	0.0097 -	ISM02.4	PPB	12/7/2017 19:06 SMPL
I6695-15	MA4Z40	87	0.0137 -	ISM02.4	PPB	12/7/2017 19:08 SMPL
I6695-16	MA4Z55	98	0.0164 -	ISM02.4	PPB	12/7/2017 19:11 SMPL
I6695-17	MA4Z56	135	0.0256 -	ISM02.4	PPB	12/7/2017 19:13 SMPL
CCV090	CCV090	19352	4.8192 -	ISM02.4	PPB	12/7/2017 19:15 SMPL
CCB090	CCB090	-80	-0.028 -	ISM02.4	PPB	12/7/2017 19:18 SMPL
I6715-14	MA4Z36	21	-0.0028 -	ISM02.4	PPB	12/7/2017 19:20 SMPL
I6756-21	MA4Z80	24	-0.0021 -	ISM02.4	PPB	12/7/2017 19:22 SMPL
CCV091	CCV091	19230	4.7887 -	ISM02.4	PPB	12/7/2017 19:31 SMPL
CCB091	CCB091	-81	-0.0282 -	ISM02.4	PPB	12/7/2017 19:34 SMPL