



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92025

Temp.: 22.6 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/07/2017	17:40
2	CAL2	1	Water			7.01	12/07/2017	17:41
3	CAL3	1	Water			10.03	12/07/2017	17:43
4	ICV	1	Water			7.00	12/07/2017	17:45
5	CCV1	1	Water			2.01	12/07/2017	17:46
6	I6677-02	1	Solid	20.03	20	8.20	12/07/2017	17:47
7	I6677-02DUP	1	Solid	20.04	20	8.21	12/07/2017	17:48
8	I6677-04	1	Solid	20.02	20	8.81	12/07/2017	17:49
9	I6789-03	1	Solid	20.01	20	8.65	12/07/2017	17:50
10	I6790-02	1	Solid	20.03	20	8.10	12/07/2017	17:52
11	I6791-07	1	Solid	20.04	20	10.02	12/07/2017	18:00
12	I6791-08	1	Solid	20.02	20	11.06	12/07/2017	18:01
13	I6791-09	1	Solid	20.02	20	11.10	12/07/2017	18:03
14	I6791-10	1	Solid	20.03	20	11.01	12/07/2017	18:05
15	I6791-11	1	Solid	20.01	20	11.13	12/07/2017	18:07
16	CCV2	1	Water			2.02	12/07/2017	18:10
17	I6791-12	1	Solid	20.03	20	11.18	12/07/2017	18:11
18	CCV3	1	Water			12.01	12/07/2017	18:13

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6777

WorkList ID : 106268

Date : 12/7/2017 5:31:03 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2017	Solid	I6790-02	Corrosivity	Cool 4 deg C	DVIR01	O11	SOILWASTE-CHARACTERIZ/	12/07/2017	9045C
12/07/2017	Solid	I6791-07	Corrosivity	Cool 4 deg C	TULL02	M11	AREA-1-1-SAMPLE	12/07/2017	9045C
12/07/2017	Solid	I6791-08	Corrosivity	Cool 4 deg C	TULL02	M11	AREA-1-2-SAMPLE	12/07/2017	9045C
12/07/2017	Solid	I6791-09	Corrosivity	Cool 4 deg C	TULL02	M11	AREA-2-1-SAMPLE	12/07/2017	9045C
12/07/2017	Solid	I6791-10	Corrosivity	Cool 4 deg C	TULL02	M11	AREA-2-2-SAMPLE	12/07/2017	9045C
12/07/2017	Solid	I6791-11	Corrosivity	Cool 4 deg C	TULL02	M11	AREA-3-1-SAMPLE	12/07/2017	9045C
12/07/2017	Solid	I6791-12	Corrosivity	Cool 4 deg C	TULL02	M11	AREA-3-2-SAMPLE	12/07/2017	9045C
12/07/2017	Solid	I6789-03	Corrosivity	Cool 4 deg C	EART12	P12	SP-1	12/07/2017	9045C
12/07/2017	Solid	I6677-02	Corrosivity	Cool 4 deg C	PSEG05	M52	OR-03-120717-A	12/07/2017	9045C
12/07/2017	Solid	I6677-04	Corrosivity	Cool 4 deg C	PSEG05	M52	OR-03-120717-B	12/07/2017	9045C

Date/Time 12-07-17 5:30 PM
 Received by: JD
 Relinquished by: CD

Date/Time 12-07-17 6:45 PM
 Received by: CD
 Relinquished by: JD