

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/08/2017

DATE & TIME IN: 12/08/2017 11:00

OVEN TEMP: 103 °C

Run Number: LB92031

DATE & TIME OUT: 12/08/2017 15:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB92031BL	LB92031BL	1.4116	1.4116	100	1.4116	1.4116	1.4116	0	0
2	LB92031BS	LB92031BS	1.4112	1.4112	100	1.4654	1.4654	1.4654	0.0542	542
3	I6762-01	120417-FOPI-F-D-B	1.4073	1.4073	200	1.4433	1.4433	1.4433	0.036	180
4	I6762-02	120417-FOPI-F-D-S	1.4135	1.4135	200	1.4299	1.4299	1.4299	0.0164	82
5	I6762-03	120417-FOPI-F-U-B	1.415	1.415	300	1.4303	1.4303	1.4303	0.0153	51
6	I6762-04	120417-FOPI-F-U-S	1.4099	1.4099	200	1.4197	1.4197	1.4197	0.0098	49
7	I6762-05	120417-TIDO-F-D-B	1.4268	1.4268	200	1.449	1.449	1.449	0.0222	111
8	I6762-05DUP	120417-TIDO-F-D-BDUP	1.4069	1.4069	200	1.4298	1.4298	1.4298	0.0229	114.5
9	I6762-06	120417-TIDO-F-D-S	1.4153	1.4153	300	1.4369	1.4369	1.4369	0.0216	72
10	I6762-07	120417-TIDO-F-U-B	1.4156	1.4156	300	1.4349	1.4349	1.4349	0.0193	64.3
11	I6762-08	120417-TIDO-F-U-S	1.4148	1.4148	300	1.4301	1.4301	1.4301	0.0153	51
12	I6762-09	120417-JETT-E-D-B	1.4128	1.4128	300	1.4285	1.4285	1.4285	0.0157	52.3
13	I6762-10	120417-JETT-E-D-S	1.406	1.406	300	1.4282	1.4282	1.4282	0.0222	74
14	I6762-11	120417-JETT-E-U-B	1.4164	1.4164	300	1.4321	1.4321	1.4321	0.0157	52.3
15	I6762-12	120417-JETT-E-U-S	1.4125	1.4125	300	1.4245	1.4245	1.4245	0.012	40

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/08/2017

DATE & TIME IN: 12/08/2017 11:00

OVEN TEMP: 103 °C

Run Number: LB92031

DATE & TIME OUT: 12/08/2017 15:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	I6763-01	120517-JETT-F-D-B	1.4	1.4	200	1.4403	1.4403	1.4403	0.0403	201.5
17	I6763-02	120517-JETT-F-D-M	1.407	1.407	200	1.4411	1.4411	1.4411	0.0341	170.5
18	I6763-03	120517-JETT-F-D-S	1.4077	1.4077	200	1.4366	1.4366	1.4366	0.0289	144.5
19	I6763-04	120517-JETT-F-U-B	1.411	1.411	200	1.4373	1.4373	1.4373	0.0263	131.5
20	I6763-05	120517-JETT-F-U-M	1.4107	1.4107	200	1.4362	1.4362	1.4362	0.0255	127.5
21	I6763-05DUP	120517-JETT-F-U-MDUP	1.4091	1.4091	200	1.4335	1.4335	1.4335	0.0244	122
22	I6763-06	120517-JETT-F-U-S	1.4165	1.4165	200	1.4398	1.4398	1.4398	0.0233	116.5
23	I6763-07	120517-TIDO-F-D-B	1.4055	1.4055	200	1.4324	1.4324	1.4324	0.0269	134.5
24	I6763-08	120517-TIDO-F-D-S	1.4109	1.4109	200	1.4425	1.4425	1.4425	0.0316	158
25	I6763-09	120517-TIDO-F-U-B	1.416	1.416	200	1.4538	1.4538	1.4538	0.0378	189
26	I6763-10	120517-TIDO-F-U-S	1.4131	1.4131	300	1.4448	1.4448	1.4448	0.0317	105.7
27	I6763-11	120517-FOPI-F-D-B	1.4123	1.4123	200	1.4313	1.4313	1.4313	0.019	95
28	I6763-12	120517-FOPI-F-D-S	1.4116	1.4116	200	1.4304	1.4304	1.4304	0.0188	94
29	I6763-13	120517-FOPI-F-U-B	1.4102	1.4102	200	1.4266	1.4266	1.4266	0.0164	82
30	I6763-14	120517-FOPI-F-U-S	1.4184	1.4184	300	1.4338	1.4338	1.4338	0.0154	51.3

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/08/2017

DATE & TIME IN: 12/08/2017 11:00

OVEN TEMP: 103 °C

Run Number: LB92031

DATE & TIME OUT: 12/08/2017 15:00

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
31	I6772-02	EFE-WW	1.4175	1.4175	400	1.4225	1.4225	1.4225	0.005	12.5
32	I6784-01	32220008-3	1.4124	1.4124	100	1.4282	1.4282	1.4282	0.0158	158
33	I6787-01	SPDES-1	1.4167	1.4167	900	1.4172	1.4172	1.4172	0.0005	0.6
34	I6793-01	15220032-1	1.4179	1.4179	250	1.4246	1.4246	1.4246	0.0067	26.8
35	I6793-01DUP	15220032-1DUP	1.4096	1.4096	250	1.416	1.416	1.416	0.0064	25.6

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS120817

WorkList ID : 106288

Date : 12/8/2017 8:37:45 AM

0392031
2892028

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/11/2017	Water	I6762-01	TSS	Cool 4 deg C	HDR103	M11	120417-FOPI-F-D-B	12/04/2017	SM2540 D
12/11/2017	Water	I6762-02	TSS	Cool 4 deg C	HDR103	M11	120417-FOPI-F-D-S	12/04/2017	SM2540 D
12/11/2017	Water	I6762-03	TSS	Cool 4 deg C	HDR103	M11	120417-FOPI-F-U-B	12/04/2017	SM2540 D
12/11/2017	Water	I6762-04	TSS	Cool 4 deg C	HDR103	M11	120417-FOPI-F-U-S	12/04/2017	SM2540 D
12/11/2017	Water	I6762-05	TSS	Cool 4 deg C	HDR103	M11	120417-TIDO-F-D-B	12/04/2017	SM2540 D
12/11/2017	Water	I6762-06	TSS	Cool 4 deg C	HDR103	M11	120417-TIDO-F-D-S	12/04/2017	SM2540 D
12/11/2017	Water	I6762-07	TSS	Cool 4 deg C	HDR103	M11	120417-TIDO-F-U-B	12/04/2017	SM2540 D
12/11/2017	Water	I6762-08	TSS	Cool 4 deg C	HDR103	M11	120417-TIDO-F-U-S	12/04/2017	SM2540 D
12/11/2017	Water	I6762-09	TSS	Cool 4 deg C	HDR103	M11	120417-JETT-E-D-B	12/04/2017	SM2540 D
12/11/2017	Water	I6762-10	TSS	Cool 4 deg C	HDR103	M11	120417-JETT-E-D-S	12/04/2017	SM2540 D
12/11/2017	Water	I6762-11	TSS	Cool 4 deg C	HDR103	M11	120417-JETT-E-U-B	12/04/2017	SM2540 D
12/11/2017	Water	I6762-12	TSS	Cool 4 deg C	HDR103	M11	120417-JETT-E-U-S	12/04/2017	SM2540 D
12/12/2017	Water	I6763-01	TSS	Cool 4 deg C	HDR103	M11	120517-JETT-F-D-B	12/05/2017	SM2540 D
12/12/2017	Water	I6763-02	TSS	Cool 4 deg C	HDR103	M11	120517-JETT-F-D-M	12/05/2017	SM2540 D
12/12/2017	Water	I6763-03	TSS	Cool 4 deg C	HDR103	M11	120517-JETT-F-D-S	12/05/2017	SM2540 D
12/12/2017	Water	I6763-04	TSS	Cool 4 deg C	HDR103	M11	120517-JETT-F-U-B	12/05/2017	SM2540 D
12/12/2017	Water	I6763-05	TSS	Cool 4 deg C	HDR103	M11	120517-JETT-F-U-M	12/05/2017	SM2540 D
12/12/2017	Water	I6763-06	TSS	Cool 4 deg C	HDR103	M11	120517-JETT-F-U-S	12/05/2017	SM2540 D
12/12/2017	Water	I6763-07	TSS	Cool 4 deg C	HDR103	M11	120517-TIDO-F-D-B	12/05/2017	SM2540 D
12/12/2017	Water	I6763-08	TSS	Cool 4 deg C	HDR103	M11	120517-TIDO-F-D-S	12/05/2017	SM2540 D
12/12/2017	Water	I6763-09	TSS	Cool 4 deg C	HDR103	M11	120517-TIDO-F-U-B	12/05/2017	SM2540 D

Date/Time 12/08/17 AS
Received by: AS
Relinquished by: WV

Date/Time 12/08/17, 12:00
Received by: WV
Relinquished by: AS

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS120817

WorkList ID : 106288

Date : 12/8/2017 8:37:45 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/12/2017	Water	I6763-10	TSS	Cool 4 deg C	HDR103	M11	120517-TIDO-F-U-S	12/05/2017	SM2540 D
12/12/2017	Water	I6763-11	TSS	Cool 4 deg C	HDR103	M11	120517-FOPI-F-D-B	12/05/2017	SM2540 D
12/12/2017	Water	I6763-12	TSS	Cool 4 deg C	HDR103	M11	120517-FOPI-F-D-S	12/05/2017	SM2540 D
12/12/2017	Water	I6763-13	TSS	Cool 4 deg C	HDR103	M11	120517-FOPI-F-U-B	12/05/2017	SM2540 D
12/12/2017	Water	I6763-14	TSS	Cool 4 deg C	HDR103	M11	120517-FOPI-F-U-S	12/05/2017	SM2540 D
12/13/2017	Water	I6772-02	TSS	Cool 4 deg C	ARDM01	P11	EFE-WW	12/06/2017	SM2540 D
12/14/2017	Water	I6784-01	TSS	Cool 4 deg C	PHAR02	P11	32220008-3	12/07/2017	SM2540 D
12/14/2017	Water	I6793-01	TSS	Cool 4 deg C	PHAR02	P11	15220032-1	12/07/2017	SM2540 D

Date/Time 12/08/17, 9:30
 Received by: AS
 Relinquished by: WV

Date/Time 12/08/17
 Received by: WV
 Relinquished by: AS

WORKLIST(Hardcopy Internal Chain)

UB92031
2007029

WorkList Name : TSS120817-2

WorkList ID : 106304

Date : 12/8/2017 10:17:44 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/08/2017	Water	I6787-01	TSS	Cool 4 deg C	SKAN06	M11	SPDES-1	12/07/2017	SM2540 D

Date/Time _____
 Received by: _____
 Relinquished by: _____

Date/Time _____
 Received by: _____
 Relinquished by: _____