


7471B

LB92041
INSTRUMENT ID: CV2

Linear

 μ Abs.:

45622



A= 0.0000e+000

B= 2.1954e-004

C= 3.0253e-002

Rho= 0.9999532

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.037	0.037	29	0.000	29				
0.2	0.200	0.221	0.021	867	0.0 %	867				
2.5	2.500	2.459	-0.041	11065	0.0 %	11065				
5.0	5.000	4.960	-0.040	22456	0.0 %	22456				
7.5	7.500	7.477	-0.023	33920	0.0 %	33920				
10.0	10.000	10.046	0.046	45622	0.0 %	45622				

LB92041 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 08 Dec 2017 14:09:18

Sample ID	Extended ID	μ Abs.	Conc.	Std CondMethod	Units	Date	Type	Type
0.0 - 1	50	29	-	0.00007471B	PPB	08 Dec 2017 14:31:06	S	Std
0.2 - 1	50.2	867	-	0.20007471B	PPB	08 Dec 2017 14:33:24	S	Std
2.5 - 1	52.5	11065	-	2.50007471B	PPB	08 Dec 2017 14:35:41	S	Std
5.0 - 1	55	22456	-	5.00007471B	PPB	08 Dec 2017 14:37:59	S	Std
7.5 - 1	57.5	33920	-	7.50007471B	PPB	08 Dec 2017 14:40:16	S	Std
10.0 - 1	59	45622	-	10.00007471B	PPB	08 Dec 2017 14:48:22	S	Std
ICV99 - 1	ICV99	16346	3.6188	-7471B	PPB	08 Dec 2017 14:51:00	U	SMPL
ICB99 - 1	ICB99	-132	0.0013	-7471B	PPB	08 Dec 2017 14:53:15	U	SMPL
CCV95 - 1	CCV95	21038	4.6489	-7471B	PPB	08 Dec 2017 14:55:31	U	SMPL
CCB95 - 1	CCB95	-142	-0.0009	-7471B	PPB	08 Dec 2017 14:57:47	U	SMPL
CRA - 1	CRA	816	0.2094	-7471B	PPB	08 Dec 2017 15:00:04	U	SMPL
HighStd - 1	HighStd	44898	9.8871	-7471B	PPB	08 Dec 2017 15:02:19	U	SMPL
ChkStd - 1	ChkStd	30337	6.6904	-7471B	PPB	08 Dec 2017 15:07:11	U	SMPL
PB104870BL - 1	PBS	-188	-0.0110	-7471B	PPB	08 Dec 2017 15:09:27	U	SMPL
PB104870BS - 1	LCSS	15539	3.4417	-7471B	PPB	08 Dec 2017 15:11:45	U	SMPL
I6725-01 - 1	P001-TPI1-SW01	178	0.0693	-7471B	PPB	08 Dec 2017 15:14:03	U	SMPL
I6725-02 - 1	P001-TPI1-SW02	671	0.1776	-7471B	PPB	08 Dec 2017 15:16:20	U	SMPL
I6725-03 - 1	P001-TPI1-SW03	1812	0.4281	-7471B	PPB	08 Dec 2017 15:18:38	U	SMPL
I6789-01 - 1	SP-1	15171	3.3609	-7471B	PPB	08 Dec 2017 15:20:56	U	SMPL
I6790-01 - 1	SOILWASTE-CHARACTERIZATION	32422	7.1482	-7471B	PPB	08 Dec 2017 15:23:11	U	SMPL
CCV96 - 1	CCV96	20742	4.5839	-7471B	PPB	08 Dec 2017 15:25:28	U	SMPL
CCB96 - 1	CCB96	-137	0.0002	-7471B	PPB	08 Dec 2017 15:27:45	U	SMPL
I6791-01 - 1	AREA-1-1-SAMPLE	9551	2.1271	-7471B	PPB	08 Dec 2017 15:30:01	U	SMPL
I6791-02 - 1	AREA-1-2-SAMPLE	6245	1.4013	-7471B	PPB	08 Dec 2017 15:32:17	U	SMPL
I6791-03 - 1	AREA-2-1-SAMPLE	6610	1.4814	-7471B	PPB	08 Dec 2017 15:34:33	U	SMPL
I6791-04 - 1	AREA-2-2-SAMPLE	5572	1.2535	-7471B	PPB	08 Dec 2017 15:36:49	U	SMPL
I6791-05 - 1	AREA-3-1-SAMPLE	4291	0.9723	-7471B	PPB	08 Dec 2017 15:39:06	U	SMPL
I6791-06 - 1	AREA-3-2-SAMPLE	4985	1.1247	-7471B	PPB	08 Dec 2017 15:41:23	U	SMPL
I6792-01 - 1	EPE-108-5(0-5)	658	0.1747	-7471B	PPB	08 Dec 2017 15:43:40	U	SMPL
I6792-01DUP - 1	EPE-108-5(0-5)DUP	570	0.1554	-7471B	PPB	08 Dec 2017 15:45:57	U	SMPL
I6792-01MS - 1	EPE-108-5(0-5)MS	16427	3.6366	-7471B	PPB	08 Dec 2017 15:48:15	U	SMPL
I6792-01MSD - 1	EPE-108-5(0-5)MSD	16299	3.6085	-7471B	PPB	08 Dec 2017 15:50:32	U	SMPL
CCV97 - 1	CCV97	20463	4.5227	-7471B	PPB	08 Dec 2017 15:52:48	U	SMPL
CCB97 - 1	CCB97	-107	0.0068	-7471B	PPB	08 Dec 2017 15:55:04	U	SMPL
I6792-02 - 1	EPE-108-5(5-10)	197	0.0735	-7471B	PPB	08 Dec 2017 15:57:22	U	SMPL
I6794-02 - 1	LGA-TA-ZONE-B-A-4-EXEAV-COMP	6705	1.5023	-7471B	PPB	08 Dec 2017 15:59:38	U	SMPL
I6792-01LX5 - 1		101	0.0524	-7471B	PPB	08 Dec 2017 16:01:54	U	SMPL
I6792-01A - 1		17519	3.8764	-7471B	PPB	08 Dec 2017 16:04:11	U	SMPL
CCV98 - 1	CCV98	21525	4.7558	-7471B	PPB	08 Dec 2017 16:09:04	U	SMPL
CCB98 - 1	CCB98	-123	0.0032	-7471B	PPB	08 Dec 2017 16:11:21	U	SMPL