


ISM02.4

LB92047
INSTRUMENT ID: CV1

Linear

 μ Abs.:

44747



A= 0.0000e+000

B= 2.2557e-004

C= 1.1906e-003

Rho= 0.9998336

Accept=Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	Q/D
0.00	0.000	0.009	0.009	33	0.000	33					4
0.2	0.200	0.209	0.009	920	0.0 %	920					2
2.5	2.500	2.547	0.047	11284	0.0 %	11284					1
5.0	5.000	4.961	-0.039	21989	0.0 %	21989					-1
7.5	7.500	7.380	-0.120	32712	0.0 %	32712					-2
10.0	10.000	10.095	0.095	44747	0.0 %	44747					1

NB
12/8/10

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	33	-		0 ISM02.4	PPB	12/8/2017 16:29	Std
	0.2 S0.2	920	-		0.2 ISM02.4	PPB	12/8/2017 16:32	Std
	2.5 S2.5	11284	-		2.5 ISM02.4	PPB	12/8/2017 16:34	Std
	5 S5	21989	-		5 ISM02.4	PPB	12/8/2017 16:36	Std
	7.5 S7.5	32712	-		7.5 ISM02.4	PPB	12/8/2017 16:39	Std
	10 S10	44747	-		10 ISM02.4	PPB	12/8/2017 16:46	Std
ICV	ICV	16790	3.7885 -		ISM02.4	PPB	12/8/2017 16:49	SMPL
ICB	ICB	-79	-0.0166 -		ISM02.4	PPB	12/8/2017 16:51	SMPL
CCV001	CCV001	21996	4.9628 -		ISM02.4	PPB	12/8/2017 16:53	SMPL
CCB001	CCB001	-57	-0.0117 -		ISM02.4	PPB	12/8/2017 16:56	SMPL
PB104833BL	PBS004	86	0.0206 -		ISM02.4	PPB	12/8/2017 16:58	SMPL
I6651-01	MC0269	209	0.0483 -		ISM02.4	PPB	12/8/2017 17:00	SMPL
I6651-02	MC0270	117	0.0276 -		ISM02.4	PPB	12/8/2017 17:02	SMPL
I6651-03	MC0271	411	0.0939 -		ISM02.4	PPB	12/8/2017 17:05	SMPL
I6651-04	MC0272	362	0.0828 -		ISM02.4	PPB	12/8/2017 17:07	SMPL
I6651-05	MC0273	279	0.0641 -		ISM02.4	PPB	12/8/2017 17:09	SMPL
I6651-06	MC0274	350	0.0801 -		ISM02.4	PPB	12/8/2017 17:11	SMPL
I6651-07	MC0275	509	0.116 -		ISM02.4	PPB	12/8/2017 17:14	SMPL
I6651-08	MC0275D	445	0.1016 -		ISM02.4	PPB	12/8/2017 17:16	SMPL
I6651-09	MC0275S	11481	2.591 -		ISM02.4	PPB	12/8/2017 17:18	SMPL
I6651-10	MC0276	606	0.1379 -		ISM02.4	PPB	12/8/2017 17:21	SMPL
I6651-11	MC0277	330	0.0756 -		ISM02.4	PPB	12/8/2017 17:23	SMPL
I6651-12	MC0278	187	0.0434 -		ISM02.4	PPB	12/8/2017 17:25	SMPL
I6651-13	MC0279	251	0.0578 -		ISM02.4	PPB	12/8/2017 17:27	SMPL
I6651-14	MC0280	206	0.0477 -		ISM02.4	PPB	12/8/2017 17:30	SMPL
I6651-15	MC0281	220	0.0508 -		ISM02.4	PPB	12/8/2017 17:32	SMPL
I6651-16	MC0282	282	0.0648 -		ISM02.4	PPB	12/8/2017 17:34	SMPL
PB104834BL	PBS004	120	0.0283 -		ISM02.4	PPB	12/8/2017 17:36	SMPL
I6710-01	MBE186	374	0.0856 -		ISM02.4	PPB	12/8/2017 17:39	SMPL
I6710-02	MBE186D	366	0.0837 -		ISM02.4	PPB	12/8/2017 17:41	SMPL
CCV002	CCV002	21856	4.9313 -		ISM02.4	PPB	12/8/2017 17:43	SMPL
CCB002	CCB002	-67	-0.0139 -		ISM02.4	PPB	12/8/2017 17:46	SMPL
I6710-03	MBE186S	11298	2.5497 -		ISM02.4	PPB	12/8/2017 17:48	SMPL
I6710-04	MBE187	3047	0.6885 -		ISM02.4	PPB	12/8/2017 17:50	SMPL
I6710-05	MBE188	9205	2.0776 -		ISM02.4	PPB	12/8/2017 17:52	SMPL
I6710-06	MBE189	5561	1.2556 -		ISM02.4	PPB	12/8/2017 17:55	SMPL
I6710-07	MBE190	135293	30.5193 -		ISM02.4	PPB	12/8/2017 17:57	SMPL
I6710-08	MBE191	4460	1.0072 -		ISM02.4	PPB	12/8/2017 17:59	SMPL
I6710-09	MBE192	2923	0.6605 -		ISM02.4	PPB	12/8/2017 18:02	SMPL
I6710-10	MBE193	1959	0.4431 -		ISM02.4	PPB	12/8/2017 18:04	SMPL
I6710-11	MBE194	1451	0.3285 -		ISM02.4	PPB	12/8/2017 18:06	SMPL
I6710-12	MBE195	3657	0.8261 -		ISM02.4	PPB	12/8/2017 18:09	SMPL
I6710-13	MBE196	1614	0.3653 -		ISM02.4	PPB	12/8/2017 18:11	SMPL
I6710-14	MBE197	3906	0.8823 -		ISM02.4	PPB	12/8/2017 18:13	SMPL
I6710-15	MBE198	8760	1.9772 -		ISM02.4	PPB	12/8/2017 18:15	SMPL
I6710-16	MBE199	198	0.0459 -		ISM02.4	PPB	12/8/2017 18:18	SMPL

INSTRUMENT ID : CV1

I6710-17	MBE1A0	1985	0.4489 -	ISM02.4	PPB	12/8/2017 18:20	SMPL
I6710-18	MBE1A1	1244	0.2818 -	ISM02.4	PPB	12/8/2017 18:22	SMPL
I6710-19	MBE1A2	7669	1.7311 -	ISM02.4	PPB	12/8/2017 18:24	SMPL
I6710-20	MBE1A3	2736	0.6184 -	ISM02.4	PPB	12/8/2017 18:27	SMPL
I6710-21	MBE1A4	8263	1.8651 -	ISM02.4	PPB	12/8/2017 18:29	SMPL
PB104835BL	PBS004	10	0.0034 -	ISM02.4	PPB	12/8/2017 18:31	SMPL
CCV003	CCV003	21654	4.8857 -	ISM02.4	PPB	12/8/2017 18:34	SMPL
CCB003	CCB003	-69	-0.0144 -	ISM02.4	PPB	12/8/2017 18:36	SMPL
I6711-01	MBE165	821	0.1864 -	ISM02.4	PPB	12/8/2017 18:38	SMPL
I6711-02	MBE166	1911	0.4323 -	ISM02.4	PPB	12/8/2017 18:40	SMPL
I6711-03	MBE167	1403	0.3177 -	ISM02.4	PPB	12/8/2017 18:43	SMPL
I6711-04	MBE167D	1274	0.2886 -	ISM02.4	PPB	12/8/2017 18:45	SMPL
I6711-05	MBE167S	12541	2.8301 -	ISM02.4	PPB	12/8/2017 18:47	SMPL
I6711-06	MBE168	862	0.1956 -	ISM02.4	PPB	12/8/2017 18:50	SMPL
I6711-07	MBE169	2175	0.4918 -	ISM02.4	PPB	12/8/2017 18:52	SMPL
I6711-08	MBE170	1019	0.231 -	ISM02.4	PPB	12/8/2017 18:54	SMPL
I6711-09	MBE171	2628	0.594 -	ISM02.4	PPB	12/8/2017 18:56	SMPL
I6711-10	MBE172	11133	2.5125 -	ISM02.4	PPB	12/8/2017 18:59	SMPL
I6711-11	MBE173	10557	2.3825 -	ISM02.4	PPB	12/8/2017 19:01	SMPL
I6711-12	MBE174	10441	2.3564 -	ISM02.4	PPB	12/8/2017 19:03	SMPL
I6711-13	MBE175	4607	1.0404 -	ISM02.4	PPB	12/8/2017 19:06	SMPL
I6711-14	MBE176	6094	1.3758 -	ISM02.4	PPB	12/8/2017 19:08	SMPL
I6711-15	MBE177	2732	0.6174 -	ISM02.4	PPB	12/8/2017 19:10	SMPL
I6711-16	MBE178	5165	1.1663 -	ISM02.4	PPB	12/8/2017 19:12	SMPL
I6711-17	MBE179	1975	0.4467 -	ISM02.4	PPB	12/8/2017 19:15	SMPL
I6711-18	MBE180	3117	0.7043 -	ISM02.4	PPB	12/8/2017 19:17	SMPL
I6711-19	MBE181	2108	0.4767 -	ISM02.4	PPB	12/8/2017 19:19	SMPL
I6711-20	MBE182	3214	0.7262 -	ISM02.4	PPB	12/8/2017 19:21	SMPL
CCV004	CCV004	21969	4.9568 -	ISM02.4	PPB	12/8/2017 19:24	SMPL
CCB004	CCB004	-72	-0.0151 -	ISM02.4	PPB	12/8/2017 19:26	SMPL
I6711-21	MBE183	3706	0.8372 -	ISM02.4	PPB	12/8/2017 19:28	SMPL
I6711-22	MBE184	3572	0.8069 -	ISM02.4	PPB	12/8/2017 19:31	SMPL
PB104836BL	PBS004	51	0.0127 -	ISM02.4	PPB	12/8/2017 19:33	SMPL
I6712-01	MBE163	1726	0.3905 -	ISM02.4	PPB	12/8/2017 19:40	SMPL
I6712-02	MBE163D	1788	0.4045 -	ISM02.4	PPB	12/8/2017 19:42	SMPL
I6712-03	MBE163S	13280	2.9968 -	ISM02.4	PPB	12/8/2017 19:45	SMPL
I6712-04	MBE1A5	1521	0.3443 -	ISM02.4	PPB	12/8/2017 19:47	SMPL
I6712-05	MBE1A6	255	0.0587 -	ISM02.4	PPB	12/8/2017 19:49	SMPL
I6712-06	MBE1A7	2251	0.509 -	ISM02.4	PPB	12/8/2017 19:52	SMPL
I6712-07	MBE1A8	1663	0.3763 -	ISM02.4	PPB	12/8/2017 19:54	SMPL
I6712-08	MBE1A9	1707	0.3862 -	ISM02.4	PPB	12/8/2017 19:56	SMPL
I6712-09	MBE1B1	338	0.0774 -	ISM02.4	PPB	12/8/2017 19:58	SMPL
I6712-10	MBE1B2	2391	0.5405 -	ISM02.4	PPB	12/8/2017 20:01	SMPL
I6712-11	MBE185	2218	0.5015 -	ISM02.4	PPB	12/8/2017 20:03	SMPL
I6712-12	MBE1B0	1148	0.2601 -	ISM02.4	PPB	12/8/2017 20:05	SMPL
PB104838BL	PBS004	26	0.0071 -	ISM02.4	PPB	12/8/2017 20:08	SMPL
I6715-01	MA4Z25	661	0.1503 -	ISM02.4	PPB	12/8/2017 20:10	SMPL

LB92047
INSTRUMENT ID : CV1

I6715-02	MA4Z26	1453	0.3289 -	ISM02.4	PPB	12/8/2017 20:12 SMPL
I6715-03	MA4Z27	2043	0.462 -	ISM02.4	PPB	12/8/2017 20:14 SMPL
I6715-04	MA4Z28	4320	0.9757 -	ISM02.4	PPB	12/8/2017 20:17 SMPL
CCV005	CCV005	21617	4.8774 -	ISM02.4	PPB	12/8/2017 20:19 SMPL
CCB005	CCB005	-51	-0.0103 -	ISM02.4	PPB	12/8/2017 20:21 SMPL
I6715-05	MA4Z29	3209	0.725 -	ISM02.4	PPB	12/8/2017 20:24 SMPL
I6715-06	MA4Z30	2225	0.5031 -	ISM02.4	PPB	12/8/2017 20:26 SMPL
I6715-07	MA4Z31	3464	0.7826 -	ISM02.4	PPB	12/8/2017 20:28 SMPL
I6715-08	MA4Z32	1090	0.2471 -	ISM02.4	PPB	12/8/2017 20:30 SMPL
I6715-09	MA4Z33	734	0.1668 -	ISM02.4	PPB	12/8/2017 20:33 SMPL
I6715-10	MA4Z34	2214	0.5006 -	ISM02.4	PPB	12/8/2017 20:35 SMPL
I6715-11	MA4Z35	2289	0.5175 -	ISM02.4	PPB	12/8/2017 20:37 SMPL
I6715-12	MA4Z35D	2707	0.6118 -	ISM02.4	PPB	12/8/2017 20:40 SMPL
I6715-13	MA4Z35S	13351	3.0128 -	ISM02.4	PPB	12/8/2017 20:42 SMPL
I6715-15	MA4Z41	4819	1.0882 -	ISM02.4	PPB	12/8/2017 20:44 SMPL
I6715-16	MA4Z42	3285	0.7422 -	ISM02.4	PPB	12/8/2017 20:46 SMPL
I6715-17	MA4Z43	1810	0.4095 -	ISM02.4	PPB	12/8/2017 20:49 SMPL
I6715-18	MA4Z44	1804	0.4081 -	ISM02.4	PPB	12/8/2017 20:51 SMPL
I6715-19	MA4Z45	1466	0.3319 -	ISM02.4	PPB	12/8/2017 20:53 SMPL
I6715-20	MA4Z46	10618	2.3963 -	ISM02.4	PPB	12/8/2017 20:56 SMPL
I6715-21	MA4Z47	5058	1.1421 -	ISM02.4	PPB	12/8/2017 20:58 SMPL
I6715-22	MA4Z48	8428	1.9023 -	ISM02.4	PPB	12/8/2017 21:00 SMPL
I6710-07DLX5	MBE190	25981	5.8617 -	ISM02.4	PPB	12/8/2017 21:02 SMPL
CCV006	CCV006	21508	4.8528 -	ISM02.4	PPB	12/8/2017 21:05 SMPL
CCB006	CCB006	-70	-0.0146 -	ISM02.4	PPB	12/8/2017 21:07 SMPL