



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92052

Temp.: 18.6 °C

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	12/08/2017	15:30
2	CAL2	1	Water			7.01	12/08/2017	15:31
3	CAL3	1	Water			10.03	12/08/2017	15:33
4	ICV	1	Water			7.00	12/08/2017	15:35
5	CCV1	1	Water			2.01	12/08/2017	15:36
6	I6811-01	1	Water			4.12	12/08/2017	15:37
7	I6811-02	1	Water			4.35	12/08/2017	15:39
8	I6811-03	1	Water			4.62	12/08/2017	15:42
9	I6813-03	1	Water			7.66	12/08/2017	15:45
10	I6813-03DUP	1	Water			7.67	12/08/2017	15:46
11	CCV2	1	Water			12.01	12/08/2017	15:47

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6811

WorkList ID : 106326

Date : 12/8/2017 3:18:55 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/07/2017	Water	I6811-01 ~ A	pH	Cool 4 deg C	STER13	P41	CELL-A-POOL-WATER	12/07/2017	SM 4500-PH B
12/07/2017	Water	I6811-02 ~ A	pH	Cool 4 deg C	STER13	P41	CELL-B-POOL-WATER	12/07/2017	SM 4500-PH B
12/07/2017	Water	I6811-03 ~ A	pH	Cool 4 deg C	STER13	P41	CELL-C-POOL-WATER	12/07/2017	SM 4500-PH B
12/08/2017	Water	I6813-03 ~ A	pH	Cool 4 deg C	PSEG01	M52	Drill-Cutting-Spoils	12/08/2017	SM 4500-PH B

Date/Time

Received by:

Relinquished by:

12-08-17 3:20 PM

JL

CP

Date/Time

Received by:

Relinquished by:

12-08-17 4:40 PM

CP

JL