



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92054

Temp.: 18.6 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	12/08/2017	15:30
2	CAL2	1	Water			7.01	12/08/2017	15:31
3	CAL3	1	Water			10.03	12/08/2017	15:33
4	ICV	1	Water			7.00	12/08/2017	15:35
5	CCV1	1	Water			2.02	12/08/2017	16:45
6	I6813-02	1	Solid	20.02	20	5.91	12/08/2017	16:47
7	I6813-02DUP	1	Solid	20.04	20	5.92	12/08/2017	16:48
8	I6814-04	1	Solid	20.05	20	8.98	12/08/2017	17:00
9	CCV2	1	Water			12.01	12/08/2017	17:01

1392054

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY I6814 WorkList ID : 106334 Date : 12/8/2017 4:35:33 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/08/2017	Solid	I6813-02	Corrosivity	Cool 4 deg C	PSEG01	M52	Soil-Cuttings	12/08/2017	9045C
12/08/2017	Solid	I6814-04	Corrosivity	Cool 4 deg C	DEWB01	P33	PHMHC-WC1	12/08/2017	9045C

Date/Time 12-08-17 5:00 PM
Received by: JH
Relinquished by: CP

Date/Time 12-08-17 6:30 PM
Received by: CP
Relinquished by: JH