

7470A


LB92057

INSTRUMENT ID: CV2

Linear

 μ Abs.:

35917



A= 0.0000e+000

B= 2.8105e-004

C= -1.1519e-002

Rho= 0.9998087

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.001	0.001	44	0.000	44				
0.2	0.200	0.203	0.003	762	0.0 %	762				
2.5	2.500	2.592	0.092	9262	0.0 %	9262				
5.0	5.000	4.902	-0.098	17483	0.0 %	17483				
7.5	7.500	7.420	-0.080	26442	0.0 %	26442				
10.0	10.000	10.083	0.083	35917	0.0 %	35917				

LB92057 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 08 Dec 2017 21:20:56

Sample ID	Extended ID	μ Abs	Conc	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	44	-	0.0000	7470A	PPB	08 Dec 2017 21:21:45	S	Std
0.2 - 1	50.2	762	-	0.2000	7470A	PPB	08 Dec 2017 21:23:54	S	Std
2.5 - 1	52.5	9262	-	2.5000	7470A	PPB	08 Dec 2017 21:26:02	S	Std
5.0 - 1	55	17483	-	5.0000	7470A	PPB	08 Dec 2017 21:28:11	S	Std
7.5 - 1	57.5	26442	-	7.5000	7470A	PPB	08 Dec 2017 21:30:18	S	Std
10.0 - 1	510	35917	-	10.0000	7470A	PPB	08 Dec 2017 21:32:26	S	Std
ICV01 - 1	ICV01	14094	3.9496	-	7470A	PPB	08 Dec 2017 21:34:56	U	SMPL
ICB01 - 1	ICB01	-53	-0.0264	-	7470A	PPB	08 Dec 2017 21:37:03	U	SMPL
CCV07 - 1	CCV07	17753	4.9779	-	7470A	PPB	08 Dec 2017 21:39:10	U	SMPL
CCB07 - 1	CCB07	-44	-0.0239	-	7470A	PPB	08 Dec 2017 21:41:17	U	SMPL
CRA - 1	CRA	741	0.1967	-	7470A	PPB	08 Dec 2017 21:43:24	U	SMPL
HighStd - 1	HighStd	35698	10.0214	-	7470A	PPB	08 Dec 2017 21:45:31	U	SMPL
ChkStd - 1	ChkStd	25271	7.0909	-	7470A	PPB	08 Dec 2017 21:47:38	U	SMPL
PB104872BL - 1	PBW	-111	-0.0427	-	7470A	PPB	08 Dec 2017 21:49:47	U	SMPL
PB104872BS - 1	LCSW	13163	3.6879	-	7470A	PPB	08 Dec 2017 21:51:56	U	SMPL
PB104872TB - 1	PB104830TB	-15	-0.0157	-	7470A	PPB	08 Dec 2017 21:54:04	U	SMPL
I6677-02 - 1	OR-03-120717-A	66	0.0070	-	7470A	PPB	08 Dec 2017 21:56:12	U	SMPL
I6677-04 - 1	OR-03-120717-B	78	0.0104	-	7470A	PPB	08 Dec 2017 21:58:20	U	SMPL
I6789-03 - 1	SP-1	83	0.0118	-	7470A	PPB	08 Dec 2017 22:00:29	U	SMPL
I6790-02 - 1	SOILWASTE-CHARACTERIZATION	72	0.0087	-	7470A	PPB	08 Dec 2017 22:02:36	U	SMPL
CCV08 - 1	CCV08	17577	4.9285	-	7470A	PPB	08 Dec 2017 22:04:42	U	SMPL
CCB08 - 1	CCB08	-42	-0.0233	-	7470A	PPB	08 Dec 2017 22:06:49	U	SMPL
I6791-07 - 1	AREA-1-1-SAMPLE	62	0.0059	-	7470A	PPB	08 Dec 2017 22:08:56	U	SMPL
I6791-08 - 1	AREA-1-2-SAMPLE	103	0.0174	-	7470A	PPB	08 Dec 2017 22:11:02	U	SMPL
I6791-09 - 1	AREA-2-1-SAMPLE	123	0.0230	-	7470A	PPB	08 Dec 2017 22:13:09	U	SMPL
I6791-10 - 1	AREA-2-2-SAMPLE	65	0.0067	-	7470A	PPB	08 Dec 2017 22:15:16	U	SMPL
I6791-11 - 1	AREA-3-1-SAMPLE	85	0.0124	-	7470A	PPB	08 Dec 2017 22:17:24	U	SMPL
I6791-12 - 1	AREA-3-2-SAMPLE	45	0.0011	-	7470A	PPB	08 Dec 2017 22:19:31	U	SMPL
I6794-04 - 1	LGA-TA-ZONE-B-A-4-EXEAV-COMP	96	0.0155	-	7470A	PPB	08 Dec 2017 22:21:39	U	SMPL
I6794-04DUP - 1	LGA-TA-ZONE-B-A-4-EXEAV-COMPDUP	84	0.0121	-	7470A	PPB	08 Dec 2017 22:23:48	U	SMPL
I6794-04MS - 1	LGA-TA-ZONE-B-A-4-EXEAV-COMPMMS	12790	3.5831	-	7470A	PPB	08 Dec 2017 22:25:56	U	SMPL
I6794-04MSD - 1	LGA-TA-ZONE-B-A-4-EXEAV-COMPMMSD	12816	3.5904	-	7470A	PPB	08 Dec 2017 22:28:03	U	SMPL
CCV09 - 1	CCV09	17615	4.9392	-	7470A	PPB	08 Dec 2017 22:30:10	U	SMPL
CCB09 - 1	CCB09	-58	-0.0278	-	7470A	PPB	08 Dec 2017 22:32:17	U	SMPL
I6794-04LX5 - 1		50	0.0025	-	7470A	PPB	08 Dec 2017 22:34:25	U	SMPL
I6794-04A - 1		13625	3.8178	-	7470A	PPB	08 Dec 2017 22:36:32	U	SMPL
CCV10 - 1	CCV10	17721	4.9690	-	7470A	PPB	08 Dec 2017 22:38:38	U	SMPL
CCB10 - 1	CCB10	-40	-0.0228	-	7470A	PPB	08 Dec 2017 22:40:46	U	SMPL