

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/11/2017

DATE & TIME IN: 12/11/2017 11:30

OVEN TEMP: 103 °C

Run Number: LB92064

DATE & TIME OUT: 12/11/2017 15:30

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB92064BL	LB92064BL	1.4157	1.4157	100	1.4157	1.4157	1.4157	0	0
2	LB92064BS	LB92064BS	1.4173	1.4173	100	1.47	1.47	1.47	0.0527	527
3	I6813-03	Drill-Cutting-Spoils	1.4206	1.4206	100	1.4588	1.4588	1.4588	0.0382	382
4	I6817-01	120617-TIDO-F-D-B	1.4311	1.4311	300	1.4486	1.4486	1.4486	0.0175	58.3
5	I6817-02	120617-TIDO-F-D-S	1.4321	1.4321	300	1.4468	1.4468	1.4468	0.0147	49
6	I6817-03	120617-TIDO-F-U-B	1.4095	1.4095	200	1.4237	1.4237	1.4237	0.0142	71
7	I6817-04	120617-TIDO-F-U-S	1.3942	1.3942	200	1.4084	1.4084	1.4084	0.0142	71
8	I6817-05	120617-JETT-F-D-B	1.4174	1.4174	200	1.4382	1.4382	1.4382	0.0208	104
9	I6817-06	120617-JETT-F-D-M	1.4049	1.4049	200	1.4211	1.4211	1.4211	0.0162	81
10	I6817-06DUP	120617-JETT-F-D-MDUP	1.4104	1.4104	200	1.4271	1.4271	1.4271	0.0167	83.5
11	I6817-07	120617-JETT-F-D-S	1.3925	1.3925	200	1.4074	1.4074	1.4074	0.0149	74.5
12	I6817-08	120617-JETT-F-U-B	1.4141	1.4141	200	1.4405	1.4405	1.4405	0.0264	132
13	I6817-09	120617-JETT-F-U-M	1.415	1.415	200	1.4295	1.4295	1.4295	0.0145	72.5
14	I6817-10	120617-JETT-F-U-S	1.4242	1.4242	200	1.4381	1.4381	1.4381	0.0139	69.5
15	I6817-11	120617-FOPI-F-D-B	1.4348	1.4348	200	1.4512	1.4512	1.4512	0.0164	82

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/11/2017

DATE & TIME IN: 12/11/2017 11:30

OVEN TEMP: 103 °C

Run Number: LB92064

DATE & TIME OUT: 12/11/2017 15:30

OVEN TEMP: 105 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
16	I6817-12	120617-FOPI-F-D-S	1.4124	1.4124	200	1.4243	1.4243	1.4243	0.0119	59.5
17	I6817-13	120617-FOPI-F-U-B	1.4153	1.4153	300	1.4272	1.4272	1.4272	0.0119	39.7
18	I6817-14	120617-FOPI-F-U-S	1.4088	1.4088	300	1.423	1.423	1.423	0.0142	47.3

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS121117-2 WorkList ID : 106377 Date : 12/11/2017 9:51:56 AM 1292064

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/15/2017	Water	I6813-03	TSS	Cool 4 deg C	PSEG01	M52	Drill-Cutting-Spoils	12/08/2017	SM2540 D
12/13/2017	Water	I6817-01	TSS	Cool 4 deg C	HDR103	M11	120617-TIDO-F-D-B	12/06/2017	SM2540 D
12/13/2017	Water	I6817-02	TSS	Cool 4 deg C	HDR103	M11	120617-TIDO-F-D-S	12/06/2017	SM2540 D
12/13/2017	Water	I6817-03	TSS	Cool 4 deg C	HDR103	M11	120617-TIDO-F-U-B	12/06/2017	SM2540 D
12/13/2017	Water	I6817-04	TSS	Cool 4 deg C	HDR103	M11	120617-TIDO-F-U-S	12/06/2017	SM2540 D
12/13/2017	Water	I6817-05	TSS	Cool 4 deg C	HDR103	M11	120617-JETT-F-D-B	12/06/2017	SM2540 D
12/13/2017	Water	I6817-06	TSS	Cool 4 deg C	HDR103	M11	120617-JETT-F-D-M	12/06/2017	SM2540 D
12/13/2017	Water	I6817-07	TSS	Cool 4 deg C	HDR103	M11	120617-JETT-F-D-S	12/06/2017	SM2540 D
12/13/2017	Water	I6817-08	TSS	Cool 4 deg C	HDR103	M11	120617-JETT-F-U-B	12/06/2017	SM2540 D
12/13/2017	Water	I6817-09	TSS	Cool 4 deg C	HDR103	M11	120617-JETT-F-U-M	12/06/2017	SM2540 D
12/13/2017	Water	I6817-10	TSS	Cool 4 deg C	HDR103	M11	120617-JETT-F-U-S	12/06/2017	SM2540 D
12/13/2017	Water	I6817-11	TSS	Cool 4 deg C	HDR103	M11	120617-FOPI-F-D-B	12/06/2017	SM2540 D
12/13/2017	Water	I6817-12	TSS	Cool 4 deg C	HDR103	M11	120617-FOPI-F-D-S	12/06/2017	SM2540 D
12/13/2017	Water	I6817-13	TSS	Cool 4 deg C	HDR103	M11	120617-FOPI-F-U-B	12/06/2017	SM2540 D
12/13/2017	Water	I6817-14	TSS	Cool 4 deg C	HDR103	M11	120617-FOPI-F-U-S	12/06/2017	SM2540 D

Date/Time 12/11/17 10:00
 Received by: AS
 Relinquished by: CP

Date/Time 12/11/17 12:30
 Received by: AS
 Relinquished by: AS