

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB92067

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP60256
calibration std. phosphate 0.5 ppm	WP60255
calibration std. phosphate 0.3 ppm	WP60254
calibration std. 0 ppm	WP60249
phosphate ICV std.	WP60258
phosphate CCV std.	WP60257
calibration std. phosphate 0.1 ppm	WP60253
calibration std. phosphate 0.05 ppm	WP60252
Combined reagent	WP60260
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0009Slope: 0.6504Regression: 0.999829

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	12/11/2017	11:25
2	CAL2	0.05	1	50	50	0.033	12/11/2017	11:26
3	CAL3	0.10	1	50	50	0.065	12/11/2017	11:26
4	CAL4	0.30	1	50	50	0.186	12/11/2017	11:27
5	CAL5	0.50	1	50	50	0.330	12/11/2017	11:27
6	CAL6	1.00	1	50	50	0.649	12/11/2017	11:28



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB92067

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.495	12/11/2017	11:29
2	ICB		1	50	50	0.000	0.001	12/11/2017	11:29
3	CCV1	0.50	1	50	50	0.327	0.504	12/11/2017	11:30
4	CCB1		1	50	50	0.000	0.001	12/11/2017	11:30
5	PB104906BL		1	50	50	0.000	0.001	12/11/2017	11:31
6	PB104906BS	0.50	1	50	50	0.320	0.493	12/11/2017	11:31
7	I6766-01		1	50	50	0.092	0.143	12/11/2017	11:32
8	I6766-01DUP		1	50	50	0.090	0.140	12/11/2017	11:32
9	I6766-01MS	0.50	1	50	50	0.390	0.601	12/11/2017	11:33
10	I6766-01MSD	0.50	1	50	50	0.387	0.596	12/11/2017	11:33
11	I6766-02		1	50	50	0.044	0.069	12/11/2017	11:34
12	I6766-03		1	50	50	0.208	0.321	12/11/2017	11:34
13	I6766-04		1	50	50	0.097	0.151	12/11/2017	11:35
14	I6766-05		1	50	50	0.227	0.350	12/11/2017	11:35
15	CCV2	0.50	1	50	50	0.328	0.506	12/11/2017	11:36
16	CCB2		1	50	50	0.000	0.001	12/11/2017	11:36
17	I6766-06		1	50	50	0.093	0.144	12/11/2017	11:37
18	I6766-07		1	50	50	0.087	0.135	12/11/2017	11:37
19	I6766-08		1	50	50	0.038	0.060	12/11/2017	11:38
20	I6766-09		1	50	50	0.091	0.141	12/11/2017	11:38
21	I6766-10		1	50	50	0.092	0.143	12/11/2017	11:39
22	I6766-11		1	50	50	0.168	0.260	12/11/2017	11:39
23	I6766-12		1	50	50	0.041	0.064	12/11/2017	11:40
24	CCV3	0.50	1	50	50	0.326	0.503	12/11/2017	11:40
25	CCB3		1	50	50	0.000	0.001	12/11/2017	11:41