

245.1


LB92071

INSTRUMENT ID: CNL

Linear

 μ Abs.:

32142



A= 0.0000e+000

B= 3.0883e-004

C= -1.9779e-002

Rho= 0.9996823

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.004	-0.004	51	0.000	51				
0.2	0.200	0.191	-0.009	682	0.0 %	682				
2.5	2.500	2.535	0.035	8274	0.0 %	8274				
5.0	5.000	4.895	-0.105	15915	0.0 %	15915				
7.5	7.500	7.676	0.176	24918	0.0 %	24918				
10.0	10.000	9.907	-0.093	32142	0.0 %	32142				

LB92071 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Dec 2017 12:03:05

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	30	51	-	0.0000	245.1	PPB	11 Dec 2017 12:15:23	S	Std
0.2 - 1	30.2	682	-	0.2000	245.1	PPB	11 Dec 2017 12:17:31	S	Std
2.5 - 1	32.5	8274	-	2.5000	245.1	PPB	11 Dec 2017 12:19:40	S	Std
5.0 - 1	35	15915	-	5.0000	245.1	PPB	11 Dec 2017 12:21:48	S	Std
7.5 - 1	37.5	24918	-	7.5000	245.1	PPB	11 Dec 2017 12:23:56	S	Std
10.0 - 1	37.5	32142	-	10.0000	245.1	PPB	11 Dec 2017 12:26:04	S	Std
ICV02 - 1	ICV02	13033	4.0052	-245.1	PPB	PPB	11 Dec 2017 12:28:45	U	SMPL
ICB02 - 1	ICB02	-85	-0.0460	-245.1	PPB	PPB	11 Dec 2017 12:30:51	U	SMPL
CCV11 - 1	CCV11	15922	4.8974	-245.1	PPB	PPB	11 Dec 2017 12:32:58	U	SMPL
CCB11 - 1	CCB11	-85	-0.0460	-245.1	PPB	PPB	11 Dec 2017 12:35:05	U	SMPL
CRA - 1	CRA	689	0.1930	-245.1	PPB	PPB	11 Dec 2017 12:37:12	U	SMPL
HighStd - 1	HighStd	32665	10.0682	-245.1	PPB	PPB	11 Dec 2017 12:39:19	U	SMPL
ChkStd - 1	ChkStd	23363	7.1955	-245.1	PPB	PPB	11 Dec 2017 12:41:26	U	SMPL
PB104876BL - 1	PBW	-50	-0.0352	-245.1	PPB	PPB	11 Dec 2017 12:43:35	U	SMPL
PB104876BS - 1	LCSW	12565	3.8607	-245.1	PPB	PPB	11 Dec 2017 12:45:44	U	SMPL
I6750-01 - 1	COMP	241	0.0546	-245.1	PPB	PPB	11 Dec 2017 12:54:54	U	SMPL
I6750-01DUP - 1	COMPDU	269	0.0633	-245.1	PPB	PPB	11 Dec 2017 12:57:01	U	SMPL
I6750-01MS - 1	COMPMS	10522	3.2298	-245.1	PPB	PPB	11 Dec 2017 12:59:09	U	SMPL
I6750-01MSD - 1	COMPMSD	10603	3.2548	-245.1	PPB	PPB	11 Dec 2017 13:01:18	U	SMPL
I6762-01 - 1	120417-FOPI-F-D-B	39	-0.0077	-245.1	PPB	PPB	11 Dec 2017 13:03:26	U	SMPL
CCV12 - 1	CCV12	16136	4.9635	-245.1	PPB	PPB	11 Dec 2017 13:05:34	U	SMPL
CCB12 - 1	CCB12	-80	-0.0445	-245.1	PPB	PPB	11 Dec 2017 13:07:41	U	SMPL
I6762-02 - 1	120417-FOPI-F-D-S	151	0.0269	-245.1	PPB	PPB	11 Dec 2017 13:09:48	U	SMPL
I6762-03 - 1	120417-FOPI-F-U-B	156	0.0284	-245.1	PPB	PPB	11 Dec 2017 13:11:55	U	SMPL
I6762-04 - 1	120417-FOPI-F-U-S	164	0.0309	-245.1	PPB	PPB	11 Dec 2017 13:14:02	U	SMPL
I6762-05 - 1	120417-TIDO-F-D-B	175	0.0343	-245.1	PPB	PPB	11 Dec 2017 13:16:09	U	SMPL
I6762-06 - 1	120417-TIDO-F-D-S	167	0.0318	-245.1	PPB	PPB	11 Dec 2017 13:18:16	U	SMPL
I6762-07 - 1	120417-TIDO-F-U-B	153	0.0275	-245.1	PPB	PPB	11 Dec 2017 13:20:24	U	SMPL
I6762-08 - 1	120417-TIDO-F-U-S	140	0.0235	-245.1	PPB	PPB	11 Dec 2017 13:22:32	U	SMPL
I6762-09 - 1	120417-JETT-E-D-B	161	0.0299	-245.1	PPB	PPB	11 Dec 2017 13:24:40	U	SMPL
I6762-10 - 1	120417-JETT-E-D-S	164	0.0309	-245.1	PPB	PPB	11 Dec 2017 13:26:49	U	SMPL
I6762-11 - 1	120417-JETT-E-U-B	197	0.0411	-245.1	PPB	PPB	11 Dec 2017 13:28:56	U	SMPL
CCV13 - 1	CCV13	16087	4.9484	-245.1	PPB	PPB	11 Dec 2017 13:31:02	U	SMPL
CCB13 - 1	CCB13	-79	-0.0442	-245.1	PPB	PPB	11 Dec 2017 13:33:09	U	SMPL
I6762-12 - 1	120417-JETT-E-U-S	163	0.0306	-245.1	PPB	PPB	11 Dec 2017 13:35:17	U	SMPL
PB104877BL - 1	PBW	157	0.0287	-245.1	PPB	PPB	11 Dec 2017 13:37:24	U	SMPL
PB104877BS - 1	LCSW	12670	3.8931	-245.1	PPB	PPB	11 Dec 2017 13:39:31	U	SMPL
I6763-01 - 1	120517-JETT-F-D-B	124	0.0185	-245.1	PPB	PPB	11 Dec 2017 13:41:38	U	SMPL
I6763-02 - 1	120517-JETT-F-D-M	211	0.0454	-245.1	PPB	PPB	11 Dec 2017 13:43:45	U	SMPL
I6763-03 - 1	120517-JETT-F-D-S	237	0.0534	-245.1	PPB	PPB	11 Dec 2017 13:45:52	U	SMPL
I6763-04 - 1	120517-JETT-F-U-B	176	0.0346	-245.1	PPB	PPB	11 Dec 2017 13:48:01	U	SMPL
I6763-05 - 1	120517-JETT-F-U-M	231	0.0516	-245.1	PPB	PPB	11 Dec 2017 13:50:09	U	SMPL
I6763-06 - 1	120517-JETT-F-U-S	214	0.0463	-245.1	PPB	PPB	11 Dec 2017 13:52:17	U	SMPL
I6763-07 - 1	120517-TIDO-F-D-B	402	0.1044	-245.1	PPB	PPB	11 Dec 2017 13:54:24	U	SMPL
CCV14 - 1	CCV14	15798	4.8592	-245.1	PPB	PPB	11 Dec 2017 13:56:32	U	SMPL
CCB14 - 1	CCB14	-76	-0.0432	-245.1	PPB	PPB	11 Dec 2017 13:58:39	U	SMPL
I6763-08 - 1	120517-TIDO-F-D-S	368	0.0939	-245.1	PPB	PPB	11 Dec 2017 14:00:47	U	SMPL
I6763-09 - 1	120517-TIDO-F-U-B	333	0.0831	-245.1	PPB	PPB	11 Dec 2017 14:02:54	U	SMPL
I6763-10 - 1	120517-TIDO-F-U-S	211	0.0454	-245.1	PPB	PPB	11 Dec 2017 14:05:01	U	SMPL
I6763-11 - 1	120517-FOPI-F-D-B	116	0.0160	-245.1	PPB	PPB	11 Dec 2017 14:07:08	U	SMPL
I6763-12 - 1	120517-FOPI-F-D-S	54	-0.0031	-245.1	PPB	PPB	11 Dec 2017 14:09:16	U	SMPL
I6763-13 - 1	120517-FOPI-F-U-B	216	0.0469	-245.1	PPB	PPB	11 Dec 2017 14:11:23	U	SMPL
I6763-14 - 1	120517-FOPI-F-U-S	214	0.0463	-245.1	PPB	PPB	11 Dec 2017 14:13:32	U	SMPL
I6763-14DUP - 1	120517-FOPI-F-U-SDUP	194	0.0401	-245.1	PPB	PPB	11 Dec 2017 14:15:40	U	SMPL
I6763-14MS - 1	120517-FOPI-F-U-SMS	14890	4.5787	-245.1	PPB	PPB	11 Dec 2017 14:17:48	U	SMPL
I6763-14MSD - 1	120517-FOPI-F-U-SMSD	14872	4.5732	-245.1	PPB	PPB	11 Dec 2017 14:19:56	U	SMPL
CCV15 - 1	CCV15	15660	4.8165	-245.1	PPB	PPB	11 Dec 2017 14:22:03	U	SMPL
CCB15 - 1	CCB15	-123	-0.0578	-245.1	PPB	PPB	11 Dec 2017 14:24:10	U	SMPL
I750-01LX5 - 1		46	-0.0056	-245.1	PPB	PPB	11 Dec 2017 14:26:19	U	SMPL
I750-01A - 1		11050	3.3928	-245.1	PPB	PPB	11 Dec 2017 14:28:26	U	SMPL
I763-14LX5 - 1		-63	-0.0392	-245.1	PPB	PPB	11 Dec 2017 14:30:33	U	SMPL
I763-14A - 1		15909	4.8934	-245.1	PPB	PPB	11 Dec 2017 14:32:42	U	SMPL
CCV16 - 1	CCV16	15696	4.8276	-245.1	PPB	PPB	11 Dec 2017 14:34:50	U	SMPL
CCB16 - 1	CCB16	-101	-0.0510	-245.1	PPB	PPB	11 Dec 2017 14:36:57	U	SMPL