


7471A

LB 99080
INSTRUMENT 115: CV1

Linear

 μ Abs.:

42642



A= 0.0000e+000

B= 2.3687e-004

C= 8.3868e-003

Rho= 0.9995673

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.008	0.008	-3	0.000	-3				
0.2	0.200	0.214	0.014	870	0.0 %	870				
2.5	2.500	2.485	-0.015	10455	0.0 %	10455				
5.0	5.000	5.102	0.102	21505	0.0 %	21505				
7.5	7.500	7.281	-0.219	30704	0.0 %	30704	0			
10.0	10.000	10.109	0.109	42642	0.0 %	42642				

LB92080 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 11 Dec 2017 15:32:46

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	-3	-	0.00007471A	PPB		11 Dec 2017 15:34:28	S	Std
0.2 - 1	50.2	870	-	0.20007471A	PPB		11 Dec 2017 15:36:45	S	Std
2.5 - 1	52.5	10455	-	2.50007471A	PPB		11 Dec 2017 15:39:03	S	Std
5.0 - 1	55	21505	-	5.00007471A	PPB		11 Dec 2017 15:41:21	S	Std
7.5 - 1	57.5	30704	-	7.50007471A	PPB		11 Dec 2017 15:43:38	S	Std
10.0 - 1	59	42642	-	10.00007471A	PPB		11 Dec 2017 15:48:30	S	Std
ICV03 - 1	ICV03	15972	3.7917	-7471A	PPB		11 Dec 2017 15:51:21	U	SMPL
ICB03 - 1	ICB03	-124	-0.0210	-7471A	PPB		11 Dec 2017 15:53:36	U	SMPL
CCV18 - 1	CCV18	21695	5.1474	-7471A	PPB		11 Dec 2017 15:55:53	U	SMPL
CCB18 - 1	CCB18	-149	-0.0269	-7471A	PPB		11 Dec 2017 15:58:08	U	SMPL
CRA - 1	CRA	847	0.2090	-7471A	PPB		11 Dec 2017 16:00:25	U	SMPL
HighStd - 1	HighStd	42977	10.1885	-7471A	PPB		11 Dec 2017 16:02:41	U	SMPL
ChkStd - 1	ChkStd	28429	6.7425	-7471A	PPB		11 Dec 2017 16:04:57	U	SMPL
PB104878BL - 1	PBS	-41	-0.0013	-7471A	PPB		11 Dec 2017 16:07:15	U	SMPL
PB104878BS - 1	LCSS	15348	3.6439	-7471A	PPB		11 Dec 2017 16:09:34	U	SMPL
I6677-01 - 1	OR-03-120717-A	1744	0.4215	-7471A	PPB		11 Dec 2017 16:11:50	U	SMPL
I6677-01DUP - 1	OR-03-120717-ADUP	1436	0.3485	-7471A	PPB		11 Dec 2017 16:14:07	U	SMPL
I6677-01MS - 1	OR-03-120717-AMS	16891	4.0094	-7471A	PPB		11 Dec 2017 16:16:25	U	SMPL
I6677-01MSD - 1	OR-03-120717-AMSD	17012	4.0381	-7471A	PPB		11 Dec 2017 16:18:42	U	SMPL
I6677-03 - 1	OR-03-120717-B	990	0.2429	-7471A	PPB		11 Dec 2017 16:20:59	U	SMPL
CCV19 - 1	CCV19	21786	5.1689	-7471A	PPB		11 Dec 2017 16:23:16	U	SMPL
CCB19 - 1	CCB19	-166	-0.0309	-7471A	PPB		11 Dec 2017 16:25:31	U	SMPL
I6764-01 - 1	MAP-8-E(0-1)	10222	2.4297	-7471A	PPB		11 Dec 2017 16:27:48	U	SMPL
I6764-02 - 1	MAP-8-E(4-5)	24506	5.8132	-7471A	PPB		11 Dec 2017 16:30:04	U	SMPL
I6764-03 - 1	MAP-8-E(7-8)	101	0.0323	-7471A	PPB		11 Dec 2017 16:32:20	U	SMPL
I6764-04 - 1	MAP-1-E(0-1)	5365	1.2792	-7471A	PPB		11 Dec 2017 16:34:37	U	SMPL
I6764-05 - 1	MAP-1-E(4-5)	5282	1.2596	-7471A	PPB		11 Dec 2017 16:36:54	U	SMPL
I6764-06 - 1	MAP-1-E(7-8)	1715	0.4146	-7471A	PPB		11 Dec 2017 16:39:11	U	SMPL
I6764-07 - 1	MAP-2-E(0-1)	5621	1.3399	-7471A	PPB		11 Dec 2017 16:41:28	U	SMPL
I6764-08 - 1	MAP-2-E(7-8)	1024	0.2509	-7471A	PPB		11 Dec 2017 16:43:45	U	SMPL
I6764-09 - 1	MAP-3-E(0-1)	26957	6.3938	-7471A	PPB		11 Dec 2017 16:46:03	U	SMPL
I6764-10 - 1	MAP-3-E(4-5)	41764	9.9012	-7471A	PPB		11 Dec 2017 16:48:19	U	SMPL
CCV20 - 1	CCV20	21390	5.0751	-7471A	PPB		11 Dec 2017 16:50:37	U	SMPL
CCB20 - 1	CCB20	-178	-0.0338	-7471A	PPB		11 Dec 2017 16:52:56	U	SMPL
I6764-11 - 1	MAP-3-E(7-8)	3034	0.7271	-7471A	PPB		11 Dec 2017 16:55:13	U	SMPL
I6677-01LX5 - 1		243	0.0659	-7471A	PPB		11 Dec 2017 16:57:30	U	SMPL
I6677-01A - 1		18568	4.4067	-7471A	PPB		11 Dec 2017 16:59:46	U	SMPL
CCV21 - 1	CCV21	21444	5.0879	-7471A	PPB		11 Dec 2017 17:02:02	U	SMPL
CCB21 - 1	CCB21	-163	-0.0302	-7471A	PPB		11 Dec 2017 17:04:19	U	SMPL