



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92092

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 98.2

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/11/2017	17:20
2	CAL2	1	Water			7.00	12/11/2017	17:21
3	CAL3	1	Water			10.04	12/11/2017	17:22
4	ICV	1	Water			6.99	12/11/2017	17:25
5	CCV1	1	Water			2.01	12/11/2017	17:26
6	I6675-04	1	Solid	20.02	20	8.46	12/11/2017	17:29
7	I6675-04DUP	1	Solid	20.03	20	8.47	12/11/2017	17:30
8	I6675-06	1	Solid	20.04	20	7.94	12/11/2017	17:31
9	I6868-03	1	Solid	20.02	20	8.65	12/11/2017	17:32
10	I6872-03	1	Solid	20.01	20	10.40	12/11/2017	17:34
11	CCV2	1	Water			12.01	12/11/2017	17:35

LB 92092

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6868 WorkList ID : 106399 Date : 12/11/2017 4:51:32 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/11/2017	Solid	I6675-04	Corrosivity	Cool 4 deg C	PSEG05		SU-03-121117-A	12/11/2017	9045C
12/11/2017	Solid	I6675-06	Corrosivity	Cool 4 deg C	PSEG05		SU-03-121117-B	12/11/2017	9045C
12/11/2017	Solid	I6868-03	Corrosivity	Cool 4 deg C	PSEG01	J62	RBR-251683-RBR-400031-24	12/11/2017	9045C
12/11/2017	Solid	I6872-03	Corrosivity	Cool 4 deg C	PSEG01	J62	293-311	12/11/2017	9045C

Date/Time 12-11-17 5:00 PM
 Received by: JL
 Relinquished by: CP

Date/Time 12-11-17 6:30 PM
 Received by: CP
 Relinquished by: Jo