



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB92103

Reagent/Standard	Lot/Log #
COD ICV-LCS std, 50ppm	WP60075
COD calibration std. 100 ppm	WP60072
COD calibration std. 10 ppm	WP60071
COD calibration std. 150 ppm	WP60073
COD calibration std. 50 ppm	WP60074
COD calibration std. 0 ppm	WP60070
COD Digestion Vials Low Level 0-150Mg/L	W2291
COD ICV-LCS std, 50ppm	WP60278
COD CCV std, 50ppm	WP60279

Temp In (C): <u>150</u>	Date In: <u>12/12/2017</u>	Time In: <u>09:40</u>
Temp Out (C): <u>150</u>	Date Out: <u>12/12/2017</u>	Time Out: <u>11:43</u>

Intercept: -1.029

Slope: 1.0231

Regression: 0.9999

Seq	Lab ID	TrueValue (mg/l)	DF	MATRIX	Reading	Anal Date	Anal Time
1	CAL1	0	1	Water	0.000	12/05/2017	12:27
2	CAL2	10	1	Water	9.000	12/05/2017	12:27
3	CAL3	50	1	Water	49.000	12/05/2017	12:28
4	CAL4	100	1	Water	101.000	12/05/2017	12:28
5	CAL5	150	1	Water	153.000	12/05/2017	12:29



Analytical Summary Report

Analysis Method: SM5220 D

ANALYST: Indiana

Parameter: COD

SUPERVISOR REVIEW BY: apatel

Run Number: LB92103

Seq	Lab ID	True Value (mg/l)	Initial Weight (g)	Final Vol (ml)	DF	MATRIX	Reading	Result	AnalDate	AnalTime
1	ICV	50			1	Water	50.000	49.877	12/05/2017	12:29
2	ICB				1	Water	0.000	1.006	12/05/2017	12:30
3	CCV1	50			1	Water	51.000	50.854	12/12/2017	12:40
4	CCB1				1	Water	0.000	1.006	12/12/2017	12:40
5	LB92103BL				1	Water	0.000	1.006	12/12/2017	12:41
6	LB92103BS	50			1	Water	50.000	49.877	12/12/2017	12:41
7	I6813-03				10	Water	14.000	14.690	12/12/2017	12:42
8	I6813-03DUP				10	Water	14.000	14.690	12/12/2017	12:42
9	I6813-03MS	50			10	Water	20.000	20.554	12/12/2017	12:43
10	I6813-03MSD	50			10	Water	20.000	20.554	12/12/2017	12:43
11	CCV2	50			1	Water	51.000	50.854	12/12/2017	12:44
12	CCB2				1	Water	0.000	1.006	12/12/2017	12:44

2892103

WORKLIST(Hardcopy Internal Chain)

WorkList Name : COD121117

WorkList ID : 106394

Date : 12/11/2017 3:41:41 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/15/2017	Water	I6813-03	COD	Conc H2SO4 4c	PSEG01	M52	Drill-Cutting-Spoils	12/08/2017	SM5220 D

Date/Time 12/12/17 9:30
Received by: IB
Relinquished by: AS

Date/Time 12/12/17 11:00
Received by: IB
Relinquished by: AS