

245.1


LB 92104

INSTRUMENT ID: CV1

Linear

 μ Abs.:

34234



A= 0.0000e+000

B= 2.8824e-004

C= 9.9325e-003

Rho= 0.9996996

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.011	0.011	3	0.000	3				
0.2	0.200	0.196	-0.004	647	0.0 %	647				
2.5	2.500	2.407	-0.093	8315	0.0 %	8315				
5.0	5.000	5.062	0.062	17527	0.0 %	17527				
7.5	7.500	7.646	0.146	26493	0.0 %	26493				
10.0	10.000	9.878	-0.122	34234	0.0 %	34234				

LB92104 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 12 Dec 2017 11:53:42

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	245.1	PPB	12 Dec 2017 11:56:48	S	Std
0.2 - 1	502	647	-	0.2000	245.1	PPB	12 Dec 2017 11:58:57	S	Std
2.5 - 1	52.5	8315	-	2.5000	245.1	PPB	12 Dec 2017 12:01:05	S	Std
5.0 - 1	55	17527	-	5.0000	245.1	PPB	12 Dec 2017 12:03:13	S	Std
7.5 - 1	57.5	26493	-	7.5000	245.1	PPB	12 Dec 2017 12:05:21	S	Std
10.0 - 1	510	34234	-	10.0000	245.1	PPB	12 Dec 2017 12:07:29	S	Std
ICV05 - 1	ICV05	13507	3.9033	-245.1	PPB	PPB	12 Dec 2017 12:12:24	U	SMPL
ICB05 - 1	ICB05	-96	-0.0177	-245.1	PPB	PPB	12 Dec 2017 12:14:30	U	SMPL
CCV28 - 1	CCV28	17334	5.0064	-245.1	PPB	PPB	12 Dec 2017 12:16:37	U	SMPL
CCB28 - 1	CCB28	-105	-0.0203	-245.1	PPB	PPB	12 Dec 2017 12:18:44	U	SMPL
CRA - 1	CRA	662	0.2008	-245.1	PPB	PPB	12 Dec 2017 12:20:51	U	SMPL
HighStd - 1	HighStd	34227	9.8757	-245.1	PPB	PPB	12 Dec 2017 12:22:57	U	SMPL
ChkStd - 1	ChkStd	24005	6.9292	-245.1	PPB	PPB	12 Dec 2017 12:25:04	U	SMPL
PB104918BL - 1	PBW	-88	-0.0154	-245.1	PPB	PPB	12 Dec 2017 12:27:13	U	SMPL
PB104918BS - 1	LCSW	13086	3.7819	-245.1	PPB	PPB	12 Dec 2017 12:29:22	U	SMPL
I6763-03 - 1	120517-JETT-F-D-S	-53	-0.0053	-245.1	PPB	PPB	12 Dec 2017 12:31:30	U	SMPL
I6763-05 - 1	120517-JETT-F-U-M	45	0.0229	-245.1	PPB	PPB	12 Dec 2017 12:33:37	U	SMPL
I6763-07 - 1	120517-TIDO-F-D-B	82	0.0336	-245.1	PPB	PPB	12 Dec 2017 12:35:46	U	SMPL
I6763-08 - 1	120517-TIDO-F-D-S	41	0.0218	-245.1	PPB	PPB	12 Dec 2017 12:37:54	U	SMPL
I6763-09 - 1	120517-TIDO-F-U-B	72	0.0307	-245.1	PPB	PPB	12 Dec 2017 12:40:01	U	SMPL
CCV29 - 1	CCV29	17206	4.9695	-245.1	PPB	PPB	12 Dec 2017 12:42:07	U	SMPL
CCB29 - 1	CCB29	-92	-0.0166	-245.1	PPB	PPB	12 Dec 2017 12:44:14	U	SMPL
I6816-01 - 1	120717-TIFR-F-D-B	16	0.0145	-245.1	PPB	PPB	12 Dec 2017 12:46:21	U	SMPL
I6816-02 - 1	120717-TIFR-F-D-M	70	0.0301	-245.1	PPB	PPB	12 Dec 2017 12:48:28	U	SMPL
I6816-03 - 1	120717-TIFR-F-D-S	58	0.0267	-245.1	PPB	PPB	12 Dec 2017 12:50:35	U	SMPL
I6816-04 - 1	120717-TIFR-F-U-B	24	0.0169	-245.1	PPB	PPB	12 Dec 2017 12:52:42	U	SMPL
I6816-05 - 1	120717-TIFR-F-U-M	47	0.0235	-245.1	PPB	PPB	12 Dec 2017 12:54:49	U	SMPL
I6816-06 - 1	120717-TIFR-F-U-S	37	0.0206	-245.1	PPB	PPB	12 Dec 2017 12:56:57	U	SMPL
I6816-07 - 1	120717-TIDO-F-D-B	58	0.0267	-245.1	PPB	PPB	12 Dec 2017 12:59:04	U	SMPL
I6816-08 - 1	120717-TIDO-F-D-S	57	0.0264	-245.1	PPB	PPB	12 Dec 2017 13:01:13	U	SMPL
I6816-09 - 1	120717-TIDO-F-U-B	48	0.0238	-245.1	PPB	PPB	12 Dec 2017 13:03:21	U	SMPL
I6816-10 - 1	120717-TIDO-F-U-S	48	0.0238	-245.1	PPB	PPB	12 Dec 2017 13:05:28	U	SMPL
CCV30 - 1	CCV30	17366	5.0156	-245.1	PPB	PPB	12 Dec 2017 13:07:35	U	SMPL
CCB30 - 1	CCB30	-91	-0.0163	-245.1	PPB	PPB	12 Dec 2017 13:09:42	U	SMPL
I6816-11 - 1	120717-FOPI-F-D-B	20	0.0157	-245.1	PPB	PPB	12 Dec 2017 13:11:50	U	SMPL
I6816-12 - 1	120717-FOPI-F-D-S	22	0.0163	-245.1	PPB	PPB	12 Dec 2017 13:13:57	U	SMPL
I6816-12DUP - 1	120717-FOPI-F-D-SDUP	22	0.0163	-245.1	PPB	PPB	12 Dec 2017 13:16:04	U	SMPL
I6816-12MS - 1	120717-FOPI-F-D-SMS	15191	4.3887	-245.1	PPB	PPB	12 Dec 2017 13:18:11	U	SMPL
I6816-12MSD - 1	120717-FOPI-F-D-SMSD	14943	4.3172	-245.1	PPB	PPB	12 Dec 2017 13:20:19	U	SMPL
PB104919BL - 1	PBW	-50	-0.0045	-245.1	PPB	PPB	12 Dec 2017 13:22:26	U	SMPL
PB104919BS - 1	LCSW	13229	3.8231	-245.1	PPB	PPB	12 Dec 2017 13:24:34	U	SMPL
I6817-01 - 1	120617-TIDO-F-D-B	-38	-0.0010	-245.1	PPB	PPB	12 Dec 2017 13:26:42	U	SMPL
I6817-02 - 1	120617-TIDO-F-D-S	25	0.0171	-245.1	PPB	PPB	12 Dec 2017 13:28:51	U	SMPL
I6817-03 - 1	120617-TIDO-F-U-B	68	0.0295	-245.1	PPB	PPB	12 Dec 2017 13:30:58	U	SMPL
CCV31 - 1	CCV31	17137	4.9496	-245.1	PPB	PPB	12 Dec 2017 13:33:05	U	SMPL
CCB31 - 1	CCB31	-86	-0.0149	-245.1	PPB	PPB	12 Dec 2017 13:35:13	U	SMPL
I6817-04 - 1	120617-TIDO-F-U-S	38	0.0209	-245.1	PPB	PPB	12 Dec 2017 13:37:21	U	SMPL
I6817-05 - 1	120617-JETT-F-D-B	56	0.0261	-245.1	PPB	PPB	12 Dec 2017 13:39:28	U	SMPL
I6817-06 - 1	120617-JETT-F-D-M	48	0.0238	-245.1	PPB	PPB	12 Dec 2017 13:41:35	U	SMPL
I6817-07 - 1	120617-JETT-F-D-S	39	0.0212	-245.1	PPB	PPB	12 Dec 2017 13:43:43	U	SMPL
I6817-08 - 1	120617-JETT-F-U-B	44	0.0226	-245.1	PPB	PPB	12 Dec 2017 13:45:50	U	SMPL
I6817-09 - 1	120617-JETT-F-U-M	40	0.0215	-245.1	PPB	PPB	12 Dec 2017 13:47:58	U	SMPL
I6817-10 - 1	120617-JETT-F-U-S	35	0.0200	-245.1	PPB	PPB	12 Dec 2017 13:50:06	U	SMPL
I6817-11 - 1	120617-FOPI-F-D-B	41	0.0218	-245.1	PPB	PPB	12 Dec 2017 13:52:14	U	SMPL
I6817-12 - 1	120617-FOPI-F-D-S	46	0.0232	-245.1	PPB	PPB	12 Dec 2017 13:54:23	U	SMPL
I6817-13 - 1	120617-FOPI-F-U-B	48	0.0238	-245.1	PPB	PPB	12 Dec 2017 13:56:30	U	SMPL
CCV32 - 1	CCV32	17139	4.9502	-245.1	PPB	PPB	12 Dec 2017 13:58:37	U	SMPL
CCB32 - 1	CCB32	-56	-0.0062	-245.1	PPB	PPB	12 Dec 2017 14:00:44	U	SMPL
I6817-14 - 1	120617-FOPI-F-U-S	60	0.0272	-245.1	PPB	PPB	12 Dec 2017 14:02:52	U	SMPL
I6817-14DUP - 1	120617-FOPI-F-U-SDUP	39	0.0212	-245.1	PPB	PPB	12 Dec 2017 14:05:00	U	SMPL
I6817-14MS - 1	120617-FOPI-F-U-SMS	14860	4.2932	-245.1	PPB	PPB	12 Dec 2017 14:07:07	U	SMPL
I6817-14MSD - 1	120617-FOPI-F-U-SMSD	14855	4.2918	-245.1	PPB	PPB	12 Dec 2017 14:09:14	U	SMPL
6824-06 - 1	TW-1	-80	-0.0131	-245.1	PPB	PPB	12 Dec 2017 14:11:21	U	SMPL
6824-07 - 1	TW-4	54	0.0255	-245.1	PPB	PPB	12 Dec 2017 14:13:29	U	SMPL
6824-08 - 1	TW-3	33	0.0194	-245.1	PPB	PPB	12 Dec 2017 14:15:37	U	SMPL
6824-09 - 1	TW-2	23	0.0166	-245.1	PPB	PPB	12 Dec 2017 14:17:45	U	SMPL
CCV33 - 1	CCV33	17007	4.9121	-245.1	PPB	PPB	12 Dec 2017 14:19:54	U	SMPL
CCB33 - 1	CCB33	-81	-0.0134	-245.1	PPB	PPB	12 Dec 2017 14:22:00	U	SMPL
6816-12LX5 - 1		10	0.0128	-245.1	PPB	PPB	12 Dec 2017 14:24:08	U	SMPL
6816-12A - 1		15714	4.5394	-245.1	PPB	PPB	12 Dec 2017 14:26:16	U	SMPL
6817-14LX5 - 1		-75	-0.0117	-245.1	PPB	PPB	12 Dec 2017 14:28:23	U	SMPL
6817-14A - 1		15585	4.5022	-245.1	PPB	PPB	12 Dec 2017 14:30:30	U	SMPL
CCV34 - 1	CCV34	17282	4.9914	-245.1	PPB	PPB	12 Dec 2017 14:32:38	U	SMPL
CCB34 - 1	CCB34	-93	-0.0169	-245.1	PPB	PPB	12 Dec 2017 14:34:46	U	SMPL