



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB92111

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	28.0	82.4	12/12/2017	09:15
2	I6813-04		1	>100.0	>212.0	12/12/2017	09:50
3	I6813-04DUP		1	>100.0	>212.0	12/12/2017	10:15

Result = (Celsius * 1.8) + 32

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP121217

WorkList ID : 106413

Date : 12/12/2017 8:22:20 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/15/2017	Water	I6813-04 C	Flash Point	Cool 4 deg C	PSEG01	M52	Drill-Cutting-Spoils	12/08/2017	1010A

Date/Time 12/12/17 8:30 AM
 Received by: KS
 Relinquished by: CP

Date/Time 12/12/17 10:20 AM
 Received by: CP
 Relinquished by: KS