


ISM02.4

LB92113
INSTRUMENT ID: C42

Linear

 μ Abs.:

44368



A= 0.0000e+000

B= 2.2545e-004

C= -1.1366e-002

Rho= 0.9999773

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	400
0.00	0.000	-0.009	-0.009	12	0.000	12					-3
0.2	0.200	0.194	-0.006	910	0.0 %	910					0
2.5	2.500	2.491	-0.009	11101	0.0 %	11101					1
5.0	5.000	5.054	0.054	22469	0.0 %	22469					0
7.5	7.500	7.478	-0.022	33218	0.0 %	33218					0
10.0	10.000	9.991	-0.009	44368	0.0 %	44368					0

12/12/15

LB92113
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	12	-		0 ISM02.4	PPB	12/12/2017 15:33	Std
	0.2 S0.2	910	-		0.2 ISM02.4	PPB	12/12/2017 15:36	Std
	2.5 S2.5	11101	-		2.5 ISM02.4	PPB	12/12/2017 15:38	Std
	5 S5	22469	-		5 ISM02.4	PPB	12/12/2017 15:40	Std
	7.5 S7.5	33218	-		7.5 ISM02.4	PPB	12/12/2017 15:43	Std
	10 S10	44368	-		10 ISM02.4	PPB	12/12/2017 15:45	Std
ICV	ICV	17454	3.9237 -		ISM02.4	PPB	12/12/2017 15:48	SMPL
ICB	ICB	-122	-0.0389 -		ISM02.4	PPB	12/12/2017 15:50	SMPL
CCV035	CCV035	22449	5.0498 -		ISM02.4	PPB	12/12/2017 15:52	SMPL
CCB035	CCB035	-131	-0.0409 -		ISM02.4	PPB	12/12/2017 15:54	SMPL
PB104931BL	PBS004	70	0.0044 -		ISM02.4	PPB	12/12/2017 15:57	SMPL
I6756-01	MA4Z59	2020	0.444 -		ISM02.4	PPB	12/12/2017 15:59	SMPL
I6756-02	MA4Z60	9095	2.0391 -		ISM02.4	PPB	12/12/2017 16:01	SMPL
I6756-03	MA4Z61	5632	1.2584 -		ISM02.4	PPB	12/12/2017 16:03	SMPL
I6756-04	MA4Z62	7972	1.7859 -		ISM02.4	PPB	12/12/2017 16:06	SMPL
I6756-05	MA4Z63	35177	7.9194 -		ISM02.4	PPB	12/12/2017 16:08	SMPL
I6756-06	MA4Z64	31823	7.1632 -		ISM02.4	PPB	12/12/2017 16:10	SMPL
I6756-07	MA4Z65	33000	7.4285 -		ISM02.4	PPB	12/12/2017 16:13	SMPL
I6756-08	MA4Z66	306	0.0576 -		ISM02.4	PPB	12/12/2017 16:15	SMPL
I6756-09	MA4Z67	5233	1.1684 -		ISM02.4	PPB	12/12/2017 16:17	SMPL
I6756-10	MA4Z68	1808	0.3963 -		ISM02.4	PPB	12/12/2017 16:20	SMPL
I6756-11	MA4Z69	2234	0.4923 -		ISM02.4	PPB	12/12/2017 16:22	SMPL
I6756-12	MA4Z70	31640	7.1219 -		ISM02.4	PPB	12/12/2017 16:24	SMPL
I6756-13	MA4Z71	2296	0.5063 -		ISM02.4	PPB	12/12/2017 16:26	SMPL
I6756-14	MA4Z72	5701	1.2739 -		ISM02.4	PPB	12/12/2017 16:29	SMPL
I6756-15	MA4Z73	4006	0.8918 -		ISM02.4	PPB	12/12/2017 16:31	SMPL
I6756-16	MA4Z74	4362	0.9721 -		ISM02.4	PPB	12/12/2017 16:33	SMPL
I6756-17	MA4Z74D	4927	1.0994 -		ISM02.4	PPB	12/12/2017 16:35	SMPL
I6756-18	MA4Z74S	15627	3.5118 -		ISM02.4	PPB	12/12/2017 16:38	SMPL
CCV036	CCV036	22242	5.0031 -		ISM02.4	PPB	12/12/2017 16:46	SMPL
CCB036	CCB036	-116	-0.0375 -		ISM02.4	PPB	12/12/2017 16:48	SMPL
I6756-19	MA4Z78	14143	3.1772 -		ISM02.4	PPB	12/12/2017 16:50	SMPL
I6756-20	MA4Z79	34	-0.0037 -		ISM02.4	PPB	12/12/2017 16:59	SMPL
CCV037	CCV037	22025	4.9542 -		ISM02.4	PPB	12/12/2017 17:01	SMPL
CCB037	CCB037	-124	-0.0393 -		ISM02.4	PPB	12/12/2017 17:04	SMPL