


7471B

LB92116
 INSTRUMENT ID: CV2

Linear

μ Abs.:

43159



A= 0.0000e+000

B= 2.3402e-004

C= 1.8438e-002

Rho= 0.9997542

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	-24	0.000	-24				
0.2	0.200	0.223	0.023	873	0.0 %	873	0			
2.5	2.500	2.530	0.030	10732	0.0 %	10732				
5.0	5.000	4.970	-0.030	21158	0.0 %	21158				
7.5	7.500	7.346	-0.154	31313	0.0 %	31313				
10.0	10.000	10.118	0.118	43159	0.0 %	43159				

LB92116 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 12 Dec 2017 18:12:24

Sample ID	Extended ID	µ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-24	-	0.00007471B	-	PPB	12 Dec 2017 18:13:48	S	Std
0.2 - 1	873	-	-	0.20007471B	-	PPB	12 Dec 2017 18:16:07	S	Std
2.5 - 1	10732	-	-	2.50007471B	-	PPB	12 Dec 2017 18:18:24	S	Std
5.0 - 1	21158	-	-	5.00007471B	-	PPB	12 Dec 2017 18:20:42	S	Std
7.5 - 1	31313	-	-	7.50007471B	-	PPB	12 Dec 2017 18:22:59	S	Std
10.0 - 1	43159	-	-	10.00007471B	-	PPB	12 Dec 2017 18:25:16	S	Std
ICV06 - 1	ICV06	16333	3.8407	-7471B	-	PPB	12 Dec 2017 18:28:12	U	SMPL
ICB06 - 1	ICB06	170	0.0582	-7471B	-	PPB	12 Dec 2017 18:30:28	U	SMPL
CCV38 - 1	CCV38	22122	5.1954	-7471B	-	PPB	12 Dec 2017 18:32:43	U	SMPL
CCB38 - 1	CCB38	123	0.0472	-7471B	-	PPB	12 Dec 2017 18:34:59	U	SMPL
CRA - 1	CRA	874	0.2230	-7471B	-	PPB	12 Dec 2017 18:37:16	U	SMPL
HighStd - 1	HighStd	43793	10.2668	-7471B	-	PPB	12 Dec 2017 18:39:32	U	SMPL
ChkStd - 1	ChkStd	29086	6.8251	-7471B	-	PPB	12 Dec 2017 18:41:48	U	SMPL
PB104929BL - 1	PBS	-146	-0.0157	-7471B	-	PPB	12 Dec 2017 18:44:07	U	SMPL
PB104929BS - 1	LCSS	15144	3.5624	-7471B	-	PPB	12 Dec 2017 18:46:26	U	SMPL
I6802-01 - 1	CONCRETE-17273	2052	0.4986	-7471B	-	PPB	12 Dec 2017 18:48:43	U	SMPL
I6813-01 - 1	Soil-Cuttings	2735	0.6585	-7471B	-	PPB	12 Dec 2017 18:55:53	U	SMPL
I6813-01DUP - 1	Soil-CuttingsDUP	2633	0.6346	-7471B	-	PPB	12 Dec 2017 18:58:11	U	SMPL
I6813-01MS - 1	Soil-CuttingsMS	19252	4.5238	-7471B	-	PPB	12 Dec 2017 19:00:28	U	SMPL
I6813-01MSD - 1	Soil-CuttingsMSD	18219	4.2820	-7471B	-	PPB	12 Dec 2017 19:02:44	U	SMPL
CCV39 - 1	CCV39	21410	5.0288	-7471B	-	PPB	12 Dec 2017 19:05:00	U	SMPL
CCB39 - 1	CCB39	148	0.0531	-7471B	-	PPB	12 Dec 2017 19:07:17	U	SMPL
I6814-01 - 1	PHMHC-WC1	38351	8.9933	-7471B	-	PPB	12 Dec 2017 19:09:33	U	SMPL
I6828-01 - 1	DBWW-2-0-4	11015	2.5962	-7471B	-	PPB	12 Dec 2017 19:11:49	U	SMPL
I6828-02 - 1	DBWW-2-4-6	56859	13.3245	-7471B	-	PPB	12 Dec 2017 19:14:07	U	SMPL
I6828-03 - 1	DBWW-2-6-10	1493	0.3678	-7471B	-	PPB	12 Dec 2017 19:16:23	U	SMPL
I6828-04 - 1	DBEW-2-0-4	13646	3.2119	-7471B	-	PPB	12 Dec 2017 19:18:42	U	SMPL
I6828-05 - 1	DBEW-2-4-10	1525	0.3753	-7471B	-	PPB	12 Dec 2017 19:20:59	U	SMPL
I6841-01 - 1	WC-01-01	47608	11.1596	-7471B	-	PPB	12 Dec 2017 19:23:16	U	SMPL
I6842-01 - 1	WC-04-01	36502	8.5606	-7471B	-	PPB	12 Dec 2017 19:25:34	U	SMPL
I6843-01 - 1	WC-04-02	163702	38.3278	-7471B	-	PPB	12 Dec 2017 19:27:53	U	SMPL
I6844-01 - 1	WC-9974103-01	27773	6.5178	-7471B	-	PPB	12 Dec 2017 19:30:11	U	SMPL
CCV40 - 1	CCV40	21192	4.9778	-7471B	-	PPB	12 Dec 2017 19:32:37	U	SMPL
CCB40 - 1	CCB40	128	0.0484	-7471B	-	PPB	12 Dec 2017 19:34:55	U	SMPL
I6845-01 - 1	WC-10257-01	87946	20.5994	-7471B	-	PPB	12 Dec 2017 19:37:13	U	SMPL
I6846-01 - 1	WC-34972326-01	28793	6.7565	-7471B	-	PPB	12 Dec 2017 19:39:29	U	SMPL
I6847-01 - 1	WC-646507-01	103294	24.1912	-7471B	-	PPB	12 Dec 2017 19:41:49	U	SMPL
I6848-01 - 1	WC-9871-01	32593	7.6458	-7471B	-	PPB	12 Dec 2017 19:44:07	U	SMPL
I6868-01 - 1	RBR-251683-RBR-400031-2428	17121	4.0251	-7471B	-	PPB	12 Dec 2017 19:46:26	U	SMPL
I6872-01 - 1	293-311	1834	0.4476	-7471B	-	PPB	12 Dec 2017 19:48:45	U	SMPL
PB104930BL - 1	PBS	12	0.0212	-7471B	-	PPB	12 Dec 2017 19:51:03	U	SMPL
PB104930BS - 1	LCSS	17922	4.2125	-7471B	-	PPB	12 Dec 2017 19:53:21	U	SMPL
I6822-01 - 1	SB-1(0-2)	1348	0.3339	-7471B	-	PPB	12 Dec 2017 19:55:39	U	SMPL
I6822-01DUP - 1	SB-1(0-2)DUP	1391	0.3440	-7471B	-	PPB	12 Dec 2017 19:57:55	U	SMPL
CCV41 - 1	CCV41	21475	5.0440	-7471B	-	PPB	12 Dec 2017 20:00:11	U	SMPL
CCB41 - 1	CCB41	-136	-0.0134	-7471B	-	PPB	12 Dec 2017 20:02:28	U	SMPL
I6822-01MS - 1	SB-1(0-2)MS	16413	3.8594	-7471B	-	PPB	12 Dec 2017 20:04:45	U	SMPL
I6822-01MSD - 1	SB-1(0-2)MSD	16857	3.9633	-7471B	-	PPB	12 Dec 2017 20:07:01	U	SMPL
I6822-02 - 1	SB-1(13-15)	28916	6.7853	-7471B	-	PPB	12 Dec 2017 20:09:18	U	SMPL
I6822-03 - 1	SB-4(0-2)	27470	6.4469	-7471B	-	PPB	12 Dec 2017 20:11:34	U	SMPL
I6822-04 - 1	SB-4(10-12)	5550	1.3172	-7471B	-	PPB	12 Dec 2017 20:13:53	U	SMPL
I6822-05 - 1	SB-3(8-10)	59329	13.9025	-7471B	-	PPB	12 Dec 2017 20:16:11	U	SMPL
I6822-06 - 1	SB-3(13-15)	7042	1.6664	-7471B	-	PPB	12 Dec 2017 20:18:29	U	SMPL
I6822-07 - 1	SB-2(0-2)	11118	2.6203	-7471B	-	PPB	12 Dec 2017 20:20:48	U	SMPL
I6822-08 - 1	SB-2(13-15)	28438	6.6735	-7471B	-	PPB	12 Dec 2017 20:23:06	U	SMPL
I6822-09 - 1	SB-9(0-2)	189612	44.3912	-7471B	-	PPB	12 Dec 2017 20:25:23	U	SMPL
CCV42 - 1	CCV42	20725	4.8685	-7471B	-	PPB	12 Dec 2017 20:27:40	U	SMPL
CCB42 - 1	CCB42	-243	-0.0384	-7471B	-	PPB	12 Dec 2017 20:30:10	U	SMPL
I6822-10 - 1	SB-9(13-15)	44467	10.4245	-7471B	-	PPB	12 Dec 2017 20:32:28	U	SMPL
I6822-11 - 1	SB-8(0-2)	70061	16.4140	-7471B	-	PPB	12 Dec 2017 20:34:44	U	SMPL
I6822-12 - 1	SB-8(13-15)	38420	9.0094	-7471B	-	PPB	12 Dec 2017 20:37:02	U	SMPL
I6822-13 - 1	SB-6(0-2)	12438	2.9292	-7471B	-	PPB	12 Dec 2017 20:39:22	U	SMPL
I6822-14 - 1	SB-6(13-15)	3265	0.7825	-7471B	-	PPB	12 Dec 2017 20:41:40	U	SMPL
I6835-01 - 1	WC-4539-02	55009	12.8916	-7471B	-	PPB	12 Dec 2017 20:43:57	U	SMPL
I6836-01 - 1	WC-70-01	30559	7.1698	-7471B	-	PPB	12 Dec 2017 20:46:15	U	SMPL
I6837-01 - 1	WC-69-01	39418	9.2430	-7471B	-	PPB	12 Dec 2017 20:48:34	U	SMPL
I6838-01 - 1	WC-7317-01	27168	6.3763	-7471B	-	PPB	12 Dec 2017 20:50:53	U	SMPL
I6839-01 - 1	WC-41446854-01	45315	10.6230	-7471B	-	PPB	12 Dec 2017 20:53:12	U	SMPL
CCV43 - 1	CCV43	21211	4.9822	-7471B	-	PPB	12 Dec 2017 20:55:31	U	SMPL
CCB43 - 1	CCB43	123	0.0472	-7471B	-	PPB	12 Dec 2017 20:57:50	U	SMPL
I6840-01 - 1	WC-41446854-02	58287	13.6587	-7471B	-	PPB	12 Dec 2017 21:00:08	U	SMPL
I6813-01LX5 - 1		380	0.1074	-7471B	-	PPB	12 Dec 2017 21:02:25	U	SMPL
I6813-01A - 1		19092	4.4863	-7471B	-	PPB	12 Dec 2017 21:04:44	U	SMPL
I6822-01LX5 - 1		347	0.0996	-7471B	-	PPB	12 Dec 2017 21:07:00	U	SMPL
I6822-01A - 1		17839	4.1931	-7471B	-	PPB	12 Dec 2017 21:09:17	U	SMPL
I6841-01DLX2 - 1		23911	5.6141	-7471B	-	PPB	12 Dec 2017 21:15:04	U	SMPL
I6843-01DLX10 - 1		16629	3.9099	-7471B	-	PPB	12 Dec 2017 21:17:21	U	SMPL
I6845-01DLX5 - 1		17066	4.0122	-7471B	-	PPB	12 Dec 2017 21:19:40	U	SMPL
I6847-01DLX5 - 1		20684	4.8589	-7471B	-	PPB	12 Dec 2017 21:21:57	U	SMPL
CV44 - 1	CCV44	21296	5.0021	-7471B	-	PPB	12 Dec 2017 21:24:14	U	SMPL
CB44 - 1	CCB44	-87	-0.0019	-7471B	-	PPB	12 Dec 2017 21:26:32	U	SMPL
I6828-02DLX5 - 1		11255	2.6523	-7471B	-	PPB	12 Dec 2017 21:28:50	U	SMPL
I6822-05DLX5 - 1		12058	2.8402	-7471B	-	PPB	12 Dec 2017 21:31:07	U	SMPL
I6822-09DLX10 - 1		18926	4.4475	-7471B	-	PPB	12 Dec 2017 21:33:24	U	SMPL
I6822-10DLX2 - 1		21826	5.1261	-7471B	-	PPB	12 Dec 2017 21:35:41	U	SMPL
I6822-11DLX5 - 1		13613	3.2041	-7471B	-	PPB	12 Dec 2017 21:37:57	U	SMPL
I6835-01DLX2 - 1		26706	6.2681	-7471B	-	PPB	12 Dec 2017 21:40:15	U	SMPL
I6839-01DLX2 - 1		22954	5.3901	-7471B	-	PPB	12 Dec 2017 21:42:32	U	SMPL
I6840-01DLX5 - 1		11279	2.6579	-7471B	-	PPB	12 Dec 2017 21:44:51	U	SMPL
CV45 - 1	CCV45	21497	5.0491	-7471B	-	PPB	12 Dec 2017 21:47:10	U	SMPL
CB45 - 1	CCB45	59	0.0322	-7471B	-	PPB	12 Dec 2017 21:49:27	U	SMPL