



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92117

Temp.: 22.3 °C

BalanceID: WC SC-4

SLOPE.: 98.1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/12/2017	17:10
2	CAL2	1	Water			7.01	12/12/2017	17:11
3	CAL3	1	Water			10.02	12/12/2017	17:13
4	ICV	1	Water			6.99	12/12/2017	17:15
5	CCV1	1	Water			2.01	12/12/2017	17:16
6	I6676-06	1	Solid	20.02	20	9.34	12/12/2017	17:18
7	I6684-04	1	Solid	20.03	20	6.67	12/12/2017	17:19
8	I6686-04	1	Solid	20.04	20	7.78	12/12/2017	17:21
9	I6885-03	1	Solid	20.03	20	7.45	12/12/2017	17:22
10	I6885-03DUP	1	Solid	20.05	20	7.46	12/12/2017	17:24
11	CCV2	1	Water			12.01	12/12/2017	17:26

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6881

WorkList ID : 106464

Date : 12/12/2017 3:45:58 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/12/2017	Solid	I6676-06	Corrosivity	Cool 4 deg C	PSEG05		OK-03-121217	12/12/2017	9045C
12/12/2017	Solid	I6684-04	Corrosivity	Cool 4 deg C	PSEG05	D52	BU-02-121217	12/12/2017	9045C
12/11/2017	Solid	I6686-04	Corrosivity	Cool 4 deg C	PSEG05		TR-04-121117	12/11/2017	9045C
12/08/2017	Solid	I6885-03	Corrosivity	Cool 4 deg C	PREF02	D11	WE-003	12/08/2017	9045C

Date/Time 12.12.17 3:40 PM
 Received by: CD
 Relinquished by: JD

Date/Time 12.12.17 6:38 PM
 Received by: CD
 Relinquished by: JD