

7470A


LB 92127

INSTRUMENT ID: CV 4

Linear

 μ Abs.:

33154



A= 0.0000e+000

B= 3.0097e-004

C= -8.5791e-003

Rho= 0.9995178

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.002	-0.002	23	0.000	23				
0.2	0.200	0.180	-0.020	627	0.0 %	627				
2.5	2.500	2.411	-0.089	8038	0.0 %	8038				
5.0	5.000	5.243	0.243	17449	0.0 %	17449				
7.5	7.500	7.398	-0.102	24609	0.0 %	24609				
10.0	10.000	9.970	-0.030	33154	0.0 %	33154				

LB92127 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 13 Dec 2017 10:55:45

Sample ID	Extended ID	μ Abs	Conc	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	23	-	0.0000	7470A	PPB	13 Dec 2017 11:41:08	S	Std
0.2 - 1	50.2	627	-	0.2000	7470A	PPB	13 Dec 2017 11:43:17	S	Std
2.5 - 1	52.5	8038	-	2.5000	7470A	PPB	13 Dec 2017 11:45:25	S	Std
5.0 - 1	55	17449	-	5.0000	7470A	PPB	13 Dec 2017 11:47:33	S	Std
7.5 - 1	57.5	24609	-	7.5000	7470A	PPB	13 Dec 2017 11:49:42	S	Std
10.0 - 1	60	33154	-	10.0000	7470A	PPB	13 Dec 2017 11:54:17	S	Std
ICV07 - 1	ICV07	13327	4.0025	-	7470A	PPB	13 Dec 2017 11:57:19	U	SMPL
ICB07 - 1	ICB07	152	0.0372	-	7470A	PPB	13 Dec 2017 11:59:25	U	SMPL
CCV46 - 1	CCV46	17386	5.2241	-	7470A	PPB	13 Dec 2017 12:01:31	U	SMPL
CCB46 - 1	CCB46	150	0.0366	-	7470A	PPB	13 Dec 2017 12:03:38	U	SMPL
CRA - 1	CRA	634	0.1822	-	7470A	PPB	13 Dec 2017 12:05:46	U	SMPL
HighStd - 1	HighStd	32927	9.9015	-	7470A	PPB	13 Dec 2017 12:07:52	U	SMPL
ChkStd - 1	ChkStd	22670	6.8144	-	7470A	PPB	13 Dec 2017 12:09:59	U	SMPL
PB105000BL - 1	PBW	64	0.0107	-	7470A	PPB	13 Dec 2017 12:12:08	U	SMPL
PB105000BS - 1	LCSW	13200	3.9642	-	7470A	PPB	13 Dec 2017 12:16:38	U	SMPL
PB105000TB - 1	PB104920TB	172	0.0432	-	7470A	PPB	13 Dec 2017 12:18:46	U	SMPL
I6675-04 - 1	SU-03-121117-A	251	0.0670	-	7470A	PPB	13 Dec 2017 12:20:54	U	SMPL
I6675-06 - 1	SU-03-121117-B	259	0.0694	-	7470A	PPB	13 Dec 2017 12:23:02	U	SMPL
I6675-06DUP - 1	SU-03-121117-BDUP	249	0.0664	-	7470A	PPB	13 Dec 2017 12:25:11	U	SMPL
I6675-06MS - 1	SU-03-121117-BMS	13238	3.9757	-	7470A	PPB	13 Dec 2017 12:27:17	U	SMPL
CCV47 - 1	CCV47	17348	5.2127	-	7470A	PPB	13 Dec 2017 12:29:24	U	SMPL
CCB47 - 1	CCB47	164	0.0408	-	7470A	PPB	13 Dec 2017 12:31:31	U	SMPL
I6675-06MSD - 1	SU-03-121117-BMSD	13634	4.0949	-	7470A	PPB	13 Dec 2017 12:33:38	U	SMPL
I6868-03 - 1	RBR-251683-RBR-400031-2428	177	0.0447	-	7470A	PPB	13 Dec 2017 12:35:45	U	SMPL
I6872-03 - 1	293-311	246	0.0655	-	7470A	PPB	13 Dec 2017 12:37:52	U	SMPL
I6675-06LX5 - 1		41	0.0038	-	7470A	PPB	13 Dec 2017 12:39:59	U	SMPL
I6675-06A - 1		13979	4.1987	-	7470A	PPB	13 Dec 2017 12:42:07	U	SMPL
CCV48 - 1	CCV48	17193	5.1660	-	7470A	PPB	13 Dec 2017 12:44:14	U	SMPL
CCB48 - 1	CCB48	153	0.0375	-	7470A	PPB	13 Dec 2017 12:46:22	U	SMPL