


ISM02.4

LB92133
INSTRUMENT ID: CV2

Linear

 μ Abs.:

41223



A= 0.0000e+000

B= 2.4376e-004

C= -1.0849e-002

Rho= 0.9999521

Accept=Accepted

slope
intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.003	0.003	56	0.000	56					-4
0.2	0.200	0.193	-0.007	836	0.0 %	836					1
2.5	2.500	2.513	0.013	10353	0.0 %	10353					1
5.0	5.000	5.027	0.027	20669	0.0 %	20669					-1
7.5	7.500	7.426	-0.074	30510	0.0 %	30510					0
10.0	10.000	10.038	0.038	41223	0.0 %	41223					

no
12/13/00

LB92133
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Stnd Conc	Method	Units	Date	Type
	0 S0.0	56	-		0 ISM02.4	PPB	12/13/2017 14:10	Std
	0.2 S0.2	836	-		0.2 ISM02.4	PPB	12/13/2017 14:12	Std
	2.5 S2.5	10353	-		2.5 ISM02.4	PPB	12/13/2017 14:14	Std
	5 S5	20669	-		5 ISM02.4	PPB	12/13/2017 14:17	Std
	7.5 S7.5	30510	-		7.5 ISM02.4	PPB	12/13/2017 14:19	Std
	10 S10	41223	-		10 ISM02.4	PPB	12/13/2017 14:24	Std
ICV	ICV	15786	3.8372	-	ISM02.4	PPB	12/13/2017 14:27	SMPL
ICB	ICB	-31	-0.0184	-	ISM02.4	PPB	12/13/2017 14:29	SMPL
CCV049	CCV049	20715	5.0387	-	ISM02.4	PPB	12/13/2017 14:31	SMPL
CCB049	CCB049	-6	-0.0123	-	ISM02.4	PPB	12/13/2017 14:33	SMPL
PB104985BL	PBW003	40	-0.0011	-	ISM02.4	PPB	12/13/2017 14:36	SMPL
I6754-01	MA4Z81	119	0.0182	-	ISM02.4	PPB	12/13/2017 14:38	SMPL
I6754-02	MA4Z82	167	0.0299	-	ISM02.4	PPB	12/13/2017 14:40	SMPL
I6754-03	MA4Z85	56	0.0028	-	ISM02.4	PPB	12/13/2017 14:42	SMPL
I6754-04	MA4Z86	105	0.0147	-	ISM02.4	PPB	12/13/2017 14:45	SMPL
I6754-05	MA4Z87	138	0.0228	-	ISM02.4	PPB	12/13/2017 14:47	SMPL
I6754-06	MA4Z88	109	0.0157	-	ISM02.4	PPB	12/13/2017 14:49	SMPL
I6754-07	MA4Z89	77	0.0079	-	ISM02.4	PPB	12/13/2017 14:52	SMPL
I6754-08	MA4Z90	122	0.0189	-	ISM02.4	PPB	12/13/2017 14:54	SMPL
I6754-09	MA4Z91	157	0.0274	-	ISM02.4	PPB	12/13/2017 14:56	SMPL
I6754-10	MA4ZB7	205	0.0391	-	ISM02.4	PPB	12/13/2017 14:58	SMPL
I6754-11	MA4ZB8	233	0.0459	-	ISM02.4	PPB	12/13/2017 15:01	SMPL
I6754-12	MA4ZB8D	181	0.0333	-	ISM02.4	PPB	12/13/2017 15:03	SMPL
I6754-13	MA4ZB8S	4554	1.0992	-	ISM02.4	PPB	12/13/2017 15:05	SMPL
I6754-14	MA4ZB9	82	0.0091	-	ISM02.4	PPB	12/13/2017 15:07	SMPL
I6754-15	MA4ZC0	145	0.0245	-	ISM02.4	PPB	12/13/2017 15:10	SMPL
PB104986BL	PBW003	48	0.0009	-	ISM02.4	PPB	12/13/2017 15:12	SMPL
I6755-01	MA4Z83	211	0.0406	-	ISM02.4	PPB	12/13/2017 15:14	SMPL
I6755-02	MA4Z84	169	0.0303	-	ISM02.4	PPB	12/13/2017 15:17	SMPL
I6755-03	MA4Z92	103	0.0143	-	ISM02.4	PPB	12/13/2017 15:19	SMPL
CCV050	CCV050	20567	5.0026	-	ISM02.4	PPB	12/13/2017 15:21	SMPL
CCB050	CCB050	-20	-0.0157	-	ISM02.4	PPB	12/13/2017 15:23	SMPL
I6755-04	MA4Z93	101	0.0138	-	ISM02.4	PPB	12/13/2017 15:26	SMPL
I6755-05	MA4Z94	180	0.033	-	ISM02.4	PPB	12/13/2017 15:28	SMPL
I6755-06	MA4Z95	138	0.0228	-	ISM02.4	PPB	12/13/2017 15:30	SMPL
I6755-07	MA4Z96	177	0.0323	-	ISM02.4	PPB	12/13/2017 15:33	SMPL
I6755-08	MA4Z97	103	0.0143	-	ISM02.4	PPB	12/13/2017 15:35	SMPL
I6755-09	MA4Z98	118	0.0179	-	ISM02.4	PPB	12/13/2017 15:37	SMPL
I6755-10	MA4ZC1	189	0.0352	-	ISM02.4	PPB	12/13/2017 15:39	SMPL
I6755-11	MA4ZC2	191	0.0357	-	ISM02.4	PPB	12/13/2017 15:42	SMPL
I6755-12	MA4ZC2D	168	0.0301	-	ISM02.4	PPB	12/13/2017 15:44	SMPL
I6755-13	MA4ZC2S	4375	1.0556	-	ISM02.4	PPB	12/13/2017 15:46	SMPL
I6755-14	MA4ZC3	112	0.0165	-	ISM02.4	PPB	12/13/2017 15:49	SMPL
I6755-15	MA4ZC4	169	0.0303	-	ISM02.4	PPB	12/13/2017 15:51	SMPL
PB104987BL	PBW003	27	-0.0043	-	ISM02.4	PPB	12/13/2017 15:53	SMPL
I6804-01	MA4ZF3	13223	3.2124	-	ISM02.4	PPB	12/13/2017 15:55	SMPL

INSTRUMENT ID : CV1

I6804-04	MA4ZF6	52	0.0018 -	ISM02.4	PPB	12/13/2017 15:58 SMPL
I6804-05	MA4ZF7	184	0.034 -	ISM02.4	PPB	12/13/2017 16:00 SMPL
I6804-06	MA4ZF8	217	0.042 -	ISM02.4	PPB	12/13/2017 16:02 SMPL
I6804-07	MA4ZF9	304	0.0633 -	ISM02.4	PPB	12/13/2017 16:04 SMPL
I6804-08	MA4ZF9D	334	0.0706 -	ISM02.4	PPB	12/13/2017 16:07 SMPL
I6804-09	MA4ZF9S	4520	1.091 -	ISM02.4	PPB	12/13/2017 16:09 SMPL
CCV051	CCV051	20244	4.9238 -	ISM02.4	PPB	12/13/2017 16:11 SMPL
CCB051	CCB051	-30	-0.0182 -	ISM02.4	PPB	12/13/2017 16:14 SMPL
PB104988BL	PBW003	63	0.0045 -	ISM02.4	PPB	12/13/2017 16:16 SMPL
I6692-22	MA4YX4	3	-0.0101 -	ISM02.4	PPB	12/13/2017 16:18 SMPL
I6774-16	MA4ZC5	11	-0.0082 -	ISM02.4	PPB	12/13/2017 16:20 SMPL
I6805-01	MA4ZG0	120	0.0184 -	ISM02.4	PPB	12/13/2017 16:23 SMPL
I6805-02	MA4ZG1	91	0.0113 -	ISM02.4	PPB	12/13/2017 16:25 SMPL
I6805-03	MA4ZG2	144	0.0243 -	ISM02.4	PPB	12/13/2017 16:27 SMPL
I6805-04	MA4ZG3	137	0.0225 -	ISM02.4	PPB	12/13/2017 16:30 SMPL
I6805-05	MA4ZG3D	154	0.0267 -	ISM02.4	PPB	12/13/2017 16:32 SMPL
I6805-06	MA4ZG3S	4367	1.0537 -	ISM02.4	PPB	12/13/2017 16:34 SMPL
I6807-17	MA4ZE9	27	-0.0043 -	ISM02.4	PPB	12/13/2017 16:36 SMPL
CCV052	CCV052	20355	4.9509 -	ISM02.4	PPB	12/13/2017 16:46 SMPL
CCB052	CCB052	-36	-0.0196 -	ISM02.4	PPB	12/13/2017 16:48 SMPL