

7471B


LB 92152

INSTRUMENT ID: CV2

Linear

 μ Abs.:

42326



A= 0.0000e+000

B= 2.3580e-004

C= -5.5904e-003

Rho= 0.9999476

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.015	0.015	88	0.000	88				
0.2	0.200	0.201	0.001	875	0.0 %	875	0			
2.5	2.500	2.491	-0.009	10589	0.0 %	10589				
5.0	5.000	4.948	-0.052	21008	0.0 %	21008				
7.5	7.500	7.570	0.070	32128	0.0 %	32128				
10.0	10.000	9.975	-0.025	42326	0.0 %	42326				

LB92152 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 13 Dec 2017 17:44:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	88	-	0.0000	7471B	PPB	13 Dec 2017 17:45:39	S	Std
0.2 - 1	50.2	875	-	0.2000	7471B	PPB	13 Dec 2017 17:47:56	S	Std
2.5 - 1	52.5	10589	-	2.5000	7471B	PPB	13 Dec 2017 17:50:14	S	Std
5.0 - 1	55	21008	-	5.0000	7471B	PPB	13 Dec 2017 17:52:31	S	Std
7.5 - 1	57.5	32128	-	7.5000	7471B	PPB	13 Dec 2017 17:54:48	S	Std
10.0 - 1	511	42326	-	10.0000	7471B	PPB	13 Dec 2017 17:59:37	S	Std
ICV08 - 1	ICV08	15636	3.6813	-	7471B	PPB	13 Dec 2017 18:02:21	U	SMPL
ICB08 - 1	ICB08	-47	-0.0167	-	7471B	PPB	13 Dec 2017 18:04:36	U	SMPL
CCV53 - 1	CCV53	21227	4.9997	-	7471B	PPB	13 Dec 2017 18:06:53	U	SMPL
CCB53 - 1	CCB53	-62	-0.0202	-	7471B	PPB	13 Dec 2017 18:09:09	U	SMPL
CRA - 1	CRA	828	0.1896	-	7471B	PPB	13 Dec 2017 18:11:26	U	SMPL
HighStd - 1	HighStd	43152	10.1695	-	7471B	PPB	13 Dec 2017 18:13:42	U	SMPL
ChkStd - 1	ChkStd	30978	7.2989	-	7471B	PPB	13 Dec 2017 18:15:58	U	SMPL
PB104978BL - 1	PBS	-67	-0.0214	-	7471B	PPB	13 Dec 2017 18:18:16	U	SMPL
PB104978BS - 1	LCSS	15659	3.6867	-	7471B	PPB	13 Dec 2017 18:20:34	U	SMPL
I6675-03 - 1	SU-03-121117-A	1195	0.2762	-	7471B	PPB	13 Dec 2017 18:22:51	U	SMPL
I6675-05 - 1	SU-03-121117-B	786	0.1797	-	7471B	PPB	13 Dec 2017 18:25:09	U	SMPL
I6686-03 - 1	TR-04-121117	1072	0.2472	-	7471B	PPB	13 Dec 2017 18:27:26	U	SMPL
I6827-01 - 1	WC-4539-01	40905	9.6397	-	7471B	PPB	13 Dec 2017 18:34:35	U	SMPL
I6827-01DUP - 1	WC-4539-01DUP	40634	9.5758	-	7471B	PPB	13 Dec 2017 18:39:41	U	SMPL
CCV54 - 1	CCV54	20901	4.9228	-	7471B	PPB	13 Dec 2017 18:41:57	U	SMPL
CCB54 - 1	CCB54	-72	-0.0226	-	7471B	PPB	13 Dec 2017 18:44:15	U	SMPL
I6827-01MS - 1	WC-4539-01MS	57485	13.5492	-	7471B	PPB	13 Dec 2017 18:49:12	U	SMPL
I6827-01MSD - 1	WC-4539-01MSD	56848	13.3990	-	7471B	PPB	13 Dec 2017 18:51:28	U	SMPL
I6830-01 - 1	MAP-3-E(12-13)	327	0.0715	-	7471B	PPB	13 Dec 2017 18:53:46	U	SMPL
I6830-02 - 1	MAP-3-E(17-18)	256	0.0548	-	7471B	PPB	13 Dec 2017 18:56:05	U	SMPL
I6830-03 - 1	MAP-3-E(13-14)	468	0.1048	-	7471B	PPB	13 Dec 2017 18:58:22	U	SMPL
I6830-04 - 1	MAP-5-E(12-13)	355	0.0781	-	7471B	PPB	13 Dec 2017 19:00:38	U	SMPL
I6830-05 - 1	MAP-5-E(15-16)	214	0.0449	-	7471B	PPB	13 Dec 2017 19:02:56	U	SMPL
I6875-01 - 1	TP-02A-2	288708	68.0708	-	7471B	PPB	13 Dec 2017 19:05:13	U	SMPL
I6875-02 - 1	TP-03A-2	24933	5.8735	-	7471B	PPB	13 Dec 2017 19:07:31	U	SMPL
I6875-03 - 1	TP-04A0-2	98893	23.3130	-	7471B	PPB	13 Dec 2017 19:10:11	U	SMPL
CCV55 - 1	CCV55	20620	4.8565	-	7471B	PPB	13 Dec 2017 19:12:29	U	SMPL
CCB55 - 1	CCB55	-130	-0.0362	-	7471B	PPB	13 Dec 2017 19:14:48	U	SMPL
I6877-01 - 1	SB-5(0-2)	25119	5.9174	-	7471B	PPB	13 Dec 2017 19:17:06	U	SMPL
I6877-02 - 1	SB-5(13-15)	7419	1.7438	-	7471B	PPB	13 Dec 2017 19:19:22	U	SMPL
I6877-03 - 1	SB-7(0-2)	32646	7.6922	-	7471B	PPB	13 Dec 2017 19:21:39	U	SMPL
I6877-04 - 1	SB-7(13-15)	4980	1.1687	-	7471B	PPB	13 Dec 2017 19:23:56	U	SMPL
I6885-02 - 1	WC-003	19605	4.6172	-	7471B	PPB	13 Dec 2017 19:26:14	U	SMPL
I6886-05 - 1	SEEP-DIRT-5	418	0.0930	-	7471B	PPB	13 Dec 2017 19:28:31	U	SMPL
I6827-01LX5 - 1		8136	1.9129	-	7471B	PPB	13 Dec 2017 19:30:49	U	SMPL
I6827-01A - 1		56760	13.3782	-	7471B	PPB	13 Dec 2017 19:33:06	U	SMPL
CCV56 - 1	CCV56	21247	5.0044	-	7471B	PPB	13 Dec 2017 19:37:28	U	SMPL
CCB56 - 1	CCB56	-60	-0.0197	-	7471B	PPB	13 Dec 2017 19:39:44	U	SMPL
I6827-01DLX2 - 1		22683	5.3430	-	7471B	PPB	13 Dec 2017 19:42:01	U	SMPL
I6827-01DUPDLX2 - 1		21343	5.0270	-	7471B	PPB	13 Dec 2017 19:44:18	U	SMPL
I6827-01MSDLX2 - 1		30727	7.2397	-	7471B	PPB	13 Dec 2017 19:46:34	U	SMPL
I6827-01MSDDLX2 - 1		30465	7.1780	-	7471B	PPB	13 Dec 2017 19:48:52	U	SMPL
I6827-01LDLX10 - 1		4397	1.0312	-	7471B	PPB	13 Dec 2017 19:51:09	U	SMPL
I6827-01ADLX2 - 1		29638	6.9830	-	7471B	PPB	13 Dec 2017 19:53:27	U	SMPL
I6875-01DLX10 - 1		30381	7.1581	-	7471B	PPB	13 Dec 2017 19:55:44	U	SMPL
I6875-03DLX5 - 1		19303	4.5460	-	7471B	PPB	13 Dec 2017 19:58:02	U	SMPL
CCV57 - 1	CCV57	21103	4.9704	-	7471B	PPB	13 Dec 2017 20:00:21	U	SMPL
CCB57 - 1	CCB57	-75	-0.0233	-	7471B	PPB	13 Dec 2017 20:02:39	U	SMPL