



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92153

Temp.: 22.4 °C

BalanceID: WC SC-4

SLOPE.: 97.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.02 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/13/2017	17:15
2	CAL2	1	Water			7.00	12/13/2017	17:16
3	CAL3	1	Water			10.04	12/13/2017	17:19
4	ICV	1	Water			6.99	12/13/2017	17:20
5	CCV1	1	Water			2.01	12/13/2017	17:21
6	I6678-02	1	Solid	20.02	20	8.51	12/13/2017	17:22
7	I6679-04	1	Solid	20.03	20	8.57	12/13/2017	17:24
8	I6679-06	1	Solid	20.01	20	8.60	12/13/2017	17:25
9	I6679-08	1	Solid	20.03	20	8.40	12/13/2017	17:26
10	I6889-03	1	Solid	20.05	20	6.30	12/13/2017	17:28
11	I6889-05	1	Solid	20.03	20	6.90	12/13/2017	17:29
12	I6889-05DUP	1	Solid	20.02	20	6.91	12/13/2017	17:30
13	CCV2	1	Water			12.02	12/13/2017	17:31

WORKLIST(Hardcopy Internal Chain)

1292153

WorkList Name : CORROSIONITY I6891

WorkList ID : 106520

Date : 12/13/2017 4:54:57 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2017	Solid	I6678-02	Corrosivity	Cool 4 deg C	PSEG05	D52	JC-01-121317	12/13/2017	9045C
12/13/2017	Solid	I6679-04	Corrosivity	Cool 4 deg C	PSEG05		HR-05-121317-A	12/13/2017	9045C
12/13/2017	Solid	I6679-06	Corrosivity	Cool 4 deg C	PSEG05		HR-05-121317-B	12/13/2017	9045C
12/13/2017	Solid	I6679-08	Corrosivity	Cool 4 deg C	PSEG05		HR-05-121317-C	12/13/2017	9045C
12/13/2017	Solid	I6889-03	Corrosivity	Cool 4 deg C	PSEG01		GIS-1(0-3)	12/13/2017	9045C
12/13/2017	Solid	I6889-05	Corrosivity	Cool 4 deg C	PSEG01		T-1(0-3)	12/13/2017	9045C

Date/Time 12-13-17 4:54 PM
 Received by: JH
 Relinquished by: CA

Date/Time 12-13-17 6:40 PM
 Received by: CA
 Relinquished by: JH