

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB92160

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP56414
calibration std. phosphate 1 ppm	WP60256
calibration std. phosphate 0.5 ppm	WP60255
calibration std. phosphate 0.3 ppm	WP60254
calibration std. 0 ppm	WP60249
phosphate ICV std.	WP60258
phosphate CCV std.	WP60257
calibration std. phosphate 0.1 ppm	WP60253
calibration std. phosphate 0.05 ppm	WP60252
Combined reagent	WP60358
11 N sulfuric acid	WP57045
Phenolphthalein indicator	WP57038
Sodium hydroxide, 1N	WP57036

Intercept: -0.0007Slope: 0.6498Regression: 0.999842

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	12/14/2017	12:00
2	CAL2	0.05	1	50	50	0.033	12/14/2017	12:01
3	CAL3	0.10	1	50	50	0.066	12/14/2017	12:01
4	CAL4	0.30	1	50	50	0.186	12/14/2017	12:02
5	CAL5	0.50	1	50	50	0.329	12/14/2017	12:02
6	CAL6	1.00	1	50	50	0.649	12/14/2017	12:03



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB92160

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.326	0.503	12/14/2017	12:03
2	ICB		1	50	50	0.000	0.001	12/14/2017	12:04
3	CCV1	0.50	1	50	50	0.327	0.504	12/14/2017	12:04
4	CCB1		1	50	50	0.000	0.001	12/14/2017	12:05
5	PB105043BL		1	50	50	0.000	0.001	12/14/2017	12:06
6	PB105043BS	0.50	1	50	50	0.318	0.490	12/14/2017	12:06
7	I6895-01		1	50	50	1.639	2.523	12/14/2017	12:07
8	I6895-01DUP		1	50	50	1.766	2.719	12/14/2017	12:07
9	I6895-01MS	0.50	1	50	50	1.910	2.940	12/14/2017	12:08
10	I6895-01MSD	0.50	1	50	50	1.709	2.631	12/14/2017	12:08
11	I6895-01		10	50	50	0.408	0.629	12/14/2017	12:20
12	I6895-01DUP		10	50	50	0.409	0.631	12/14/2017	12:20
13	I6895-01MS	0.50	10	50	50	0.438	0.675	12/14/2017	12:21
14	I6895-01MSD	0.50	10	50	50	0.436	0.672	12/14/2017	12:21
15	CCV2	0.50	1	50	50	0.329	0.507	12/14/2017	12:22
16	CCB2		1	50	50	0.000	0.001	12/14/2017	12:22