



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92170

Temp.: 23.2 °C

BalanceID: WC SC-4

SLOPE.: 98.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	12/14/2017	17:30
2	CAL2	1	Water			7.01	12/14/2017	17:31
3	CAL3	1	Water			10.01	12/14/2017	17:32
4	ICV	1	Water			7.01	12/14/2017	17:33
5	CCV1	1	Water			2.01	12/14/2017	17:35
6	I6680-12	1	Solid	20.02	20	9.50	12/14/2017	17:36
7	I6680-12DUP	1	Solid	20.03	20	9.51	12/14/2017	17:37
8	I6680-14	1	Solid	20.04	20	7.60	12/14/2017	17:38
9	I6680-16	1	Solid	20.02	20	10.36	12/14/2017	17:40
10	I6680-18	1	Solid	20.03	20	9.64	12/14/2017	17:41
11	I6744-12	1	Solid	20.04	20	8.53	12/14/2017	17:42
12	I6911-02	1	Solid	20.02	20	11.21	12/14/2017	17:44
13	I6919-02	1	Solid	20.04	20	8.43	12/14/2017	17:45
14	CCV2	1	Water			12.01	12/14/2017	17:46

2342170

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY I6919 WorkList ID : 106572 Date : 12/14/2017 5:20:25 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/14/2017	Solid	I6880-12	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-121417-A	12/14/2017	9045C
12/14/2017	Solid	I6880-14	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-121417-B	12/14/2017	9045C
12/14/2017	Solid	I6880-16	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-121417-C	12/14/2017	9045C
12/14/2017	Solid	I6880-18	Corrosivity	Cool 4 deg C	PSEG05	D52	CL-02-121417-D	12/14/2017	9045C
12/04/2017	Solid	I6744-12	Corrosivity	Cool 4 deg C	WOOD06		PROS- 3-E(11-12)	12/04/2017	9045C
12/14/2017	Solid	I6911-02	Corrosivity	Cool 4 deg C	PSEG01	D52	SL-TRANSFORMERS	12/14/2017	9045C
12/14/2017	Solid	I6919-02	Corrosivity	Cool 4 deg C	PSEG01	D52	339	12/14/2017	9045C

Date/Time 12.16.17 5:20 PM
 Received by: JH
 Relinquished by: JH

Date/Time 12.14.17 6:46 PM
 Received by: JH
 Relinquished by: CP