

LB 92176


ISM02.4

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40578



A= 0.0000e+000

B= 2.4818e-004

C= -5.1745e-003

Rho= 0.9998888

Accept=Accepted

slope
4.157e-004

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.003	0.003	32	0.000	32					-
0.2	0.200	0.192	-0.008	793	0.0 %	793					-4
2.5	2.500	2.538	0.038	10248	0.0 %	10248					2
5.0	5.000	5.012	0.012	20218	0.0 %	20218					0
7.5	7.500	7.390	-0.110	29796	0.0 %	29796					-1
10.0	10.000	10.065	0.065	40578	0.0 %	40578					1

NB
12/15/10

LB92176
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	32	-		0 ISM02.4	PPB	12/15/2017 9:49	Std
	0.2 S0.2	793	-		0.2 ISM02.4	PPB	12/15/2017 9:52	Std
	2.5 S2.5	10248	-		2.5 ISM02.4	PPB	12/15/2017 9:54	Std
	5 S5	20218	-		5 ISM02.4	PPB	12/15/2017 9:56	Std
	7.5 S7.5	29796	-		7.5 ISM02.4	PPB	12/15/2017 9:59	Std
	10 S10	40578	-		10 ISM02.4	PPB	12/15/2017 10:01	Std
ICV	ICV	15272	3.785 -		ISM02.4	PPB	12/15/2017 10:04	SMPL
ICB	ICB	-16	-0.0091 -		ISM02.4	PPB	12/15/2017 10:06	SMPL
CCV063	CCV063	20427	5.0644 -		ISM02.4	PPB	12/15/2017 10:08	SMPL
CCB063	CCB063	-49	-0.0173 -		ISM02.4	PPB	12/15/2017 10:10	SMPL
PB105040BL	PBW003	89	0.0169 -		ISM02.4	PPB	12/15/2017 10:13	SMPL
I6800-01	MC0291	135	0.0283 -		ISM02.4	PPB	12/15/2017 10:15	SMPL
I6800-02	MC0293	92	0.0177 -		ISM02.4	PPB	12/15/2017 10:17	SMPL
I6800-05	MC0292	119	0.0244 -		ISM02.4	PPB	12/15/2017 10:19	SMPL
I6800-06	MC0296	177	0.0388 -		ISM02.4	PPB	12/15/2017 10:22	SMPL
I6800-07	MC0296D	165	0.0358 -		ISM02.4	PPB	12/15/2017 10:24	SMPL
I6800-08	MC0296S	4212	1.0402 -		ISM02.4	PPB	12/15/2017 10:26	SMPL
I6800-13	MC0297	116	0.0236 -		ISM02.4	PPB	12/15/2017 10:29	SMPL
I6800-15	MC02D7	135	0.0283 -		ISM02.4	PPB	12/15/2017 10:31	SMPL
PB105041BL	PBW003	178	0.039 -		ISM02.4	PPB	12/15/2017 10:33	SMPL
I6803-01	MBDYW5	163	0.0353 -		ISM02.4	PPB	12/15/2017 10:35	SMPL
I6803-02	MBDYW6	135	0.0283 -		ISM02.4	PPB	12/15/2017 10:38	SMPL
I6803-03	MBDYZ4	139	0.0293 -		ISM02.4	PPB	12/15/2017 10:40	SMPL
I6803-04	MBDYW1	152	0.0325 -		ISM02.4	PPB	12/15/2017 10:42	SMPL
I6803-05	MBDYW2	145	0.0308 -		ISM02.4	PPB	12/15/2017 10:45	SMPL
I6803-06	MBDYW9	95	0.0184 -		ISM02.4	PPB	12/15/2017 10:47	SMPL
I6803-07	MBDYZ5	85	0.0159 -		ISM02.4	PPB	12/15/2017 10:49	SMPL
I6803-08	MBDYT4	136	0.0286 -		ISM02.4	PPB	12/15/2017 10:51	SMPL
I6803-09	MBDYT6	177	0.0388 -		ISM02.4	PPB	12/15/2017 10:54	SMPL
I6803-10	MBDYT8	171	0.0373 -		ISM02.4	PPB	12/15/2017 10:56	SMPL
CCV064	CCV064	20244	5.0189 -		ISM02.4	PPB	12/15/2017 10:58	SMPL
CCB064	CCB064	-52	-0.0181 -		ISM02.4	PPB	12/15/2017 11:00	SMPL
I6803-11	MBDYT8D	24	0.0008 -		ISM02.4	PPB	12/15/2017 11:03	SMPL
I6803-12	MBDYT8S	4296	1.061 -		ISM02.4	PPB	12/15/2017 11:05	SMPL
I6803-13	MBDYT9	76	0.0137 -		ISM02.4	PPB	12/15/2017 11:07	SMPL
I6803-14	MBDYW0	55	0.0085 -		ISM02.4	PPB	12/15/2017 11:10	SMPL
I6803-15	MBDYW3	36	0.0038 -		ISM02.4	PPB	12/15/2017 11:12	SMPL
I6803-16	MBDYZ0	101	0.0199 -		ISM02.4	PPB	12/15/2017 11:14	SMPL
I6803-17	MBDYZ6	44	0.0057 -		ISM02.4	PPB	12/15/2017 11:16	SMPL
I6803-18	MBDZ07	63	0.0105 -		ISM02.4	PPB	12/15/2017 11:19	SMPL
CCV065	CCV065	20208	5.01 -		ISM02.4	PPB	12/15/2017 11:21	SMPL
CCB065	CCB065	-22	-0.0106 -		ISM02.4	PPB	12/15/2017 11:23	SMPL