



TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatelANALYST: AparnaDate: 12/15/2017DATE & TIME IN: 12/15/2017 11:20OVEN TEMP: 103 °CRun Number: LB92180DATE & TIME OUT: 12/15/2017 15:20OVEN TEMP: 103 °CBalanceID: WC SC-3OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB92180BL	LB92180BL	1.4091	1.4091	100	1.4091	1.4091	1.4091	0	0
2	LB92180BS	LB92180BS	1.4107	1.4107	100	1.4603	1.4603	1.4603	0.0496	496
3	I6926-01	SPDES-1	1.4138	1.4138	800	1.4152	1.4152	1.4152	0.0014	1.8

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS121517-2

WorkList ID : 106604

Date : 12/15/2017 11:38:19 AM

1392180

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/15/2017	Water	I6926-01	TSS	Cool 4 deg C	SKAN06	F11	SPDES-1	12/14/2017	SM2540 D

Date/Time 12/15/17, 11:05
 Received by: AS
 Relinquished by: CP

Date/Time 12/15/17, 12:00
 Received by: CP
 Relinquished by: AS