


7470A

LB92182
INSTRUMENT ID: C42

Linear

 μ Abs.:

35187



A= 0.0000e+000

B= 2.8349e-004

C= -9.2738e-003

Rho= 0.9998538

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.003	0.003	44	0.000	44				
0.2	0.200	0.190	-0.010	702	0.0 %	702				
2.5	2.500	2.432	-0.068	8611	0.0 %	8611				
5.0	5.000	5.132	0.132	18136	0.0 %	18136				
7.5	7.500	7.477	-0.023	26409	0.0 %	26409				
10.0	10.000	9.966	-0.034	35187	0.0 %	35187				

LB92182 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 15 Dec 2017 12:16:24

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc/Method	Units	Date	Type	Type
0.0 - 1	S0	44	-	0.00007470A	PPB	15 Dec 2017 12:18:02	S	Std
0.2 - 1	S0.2	702	-	0.20007470A	PPB	15 Dec 2017 12:20:10	S	Std
2.5 - 1	S2.5	8611	-	2.50007470A	PPB	15 Dec 2017 12:22:19	S	Std
5.0 - 1	S5	18136	-	5.00007470A	PPB	15 Dec 2017 12:24:27	S	Std
7.5 - 1	S7.5	26409	-	7.50007470A	PPB	15 Dec 2017 12:26:35	S	Std
10.0 - 1	S10	35187	-	10.00007470A	PPB	15 Dec 2017 12:28:43	S	Std
ICV10 - 1	ICV10	13896	3.9301	-7470A	PPB	15 Dec 2017 12:31:54	U	SMPL
ICB10 - 1	ICB10	-74	-0.0303	-7470A	PPB	15 Dec 2017 12:34:00	U	SMPL
CCV66 - 1	CCV66	18201	5.1505	-7470A	PPB	15 Dec 2017 12:36:07	U	SMPL
CCB66 - 1	CCB66	-106	-0.0393	-7470A	PPB	15 Dec 2017 12:38:13	U	SMPL
CRA - 1	CRA	692	0.1869	-7470A	PPB	15 Dec 2017 12:40:21	U	SMPL
HighStd - 1	HighStd	35119	9.9465	-7470A	PPB	15 Dec 2017 12:42:27	U	SMPL
ChkStd - 1	ChkStd	24163	6.8406	-7470A	PPB	15 Dec 2017 12:44:34	U	SMPL
PB105051BL - 1	PBW	-142	-0.0495	-7470A	PPB	15 Dec 2017 12:46:44	U	SMPL
PB105051BS - 1	LCSW	14292	4.0423	-7470A	PPB	15 Dec 2017 12:48:53	U	SMPL
PB105051TB - 1	PB105011TB	-66	-0.0280	-7470A	PPB	15 Dec 2017 12:51:00	U	SMPL
I6676-06 - 1	OK-03-121217	6	-0.0076	-7470A	PPB	15 Dec 2017 12:53:08	U	SMPL
I6678-02 - 1	JC-01-121317	62	0.0083	-7470A	PPB	15 Dec 2017 12:55:17	U	SMPL
I6678-02DUP - 1	JC-01-121317DUP	62	0.0083	-7470A	PPB	15 Dec 2017 12:57:25	U	SMPL
I6678-02MS - 1	JC-01-121317MS	13162	3.7220	-7470A	PPB	15 Dec 2017 12:59:32	U	SMPL
CCV67 - 1	CCV67	18084	5.1173	-7470A	PPB	15 Dec 2017 13:01:38	U	SMPL
CCB67 - 1	CCB67	-100	-0.0376	-7470A	PPB	15 Dec 2017 13:03:45	U	SMPL
I6678-02MSD - 1	JC-01-121317MSD	13342	3.7730	-7470A	PPB	15 Dec 2017 13:05:52	U	SMPL
I6679-04 - 1	HR-05-121317-A	-23	-0.0158	-7470A	PPB	15 Dec 2017 13:07:59	U	SMPL
I6679-06 - 1	HR-05-121317-B	25	-0.0022	-7470A	PPB	15 Dec 2017 13:10:06	U	SMPL
I6679-08 - 1	HR-05-121317-C	39	0.0018	-7470A	PPB	15 Dec 2017 13:12:14	U	SMPL
I6684-04 - 1	BU-02-121217	24	-0.0025	-7470A	PPB	15 Dec 2017 13:14:22	U	SMPL
I6889-03 - 1	GIS-1(0-3)	23	-0.0028	-7470A	PPB	15 Dec 2017 13:16:30	U	SMPL
I6889-05 - 1	T-1(0-3)	12	-0.0059	-7470A	PPB	15 Dec 2017 13:18:38	U	SMPL
I6889-06 - 1	FSA(0-3)	41	0.0023	-7470A	PPB	15 Dec 2017 13:20:47	U	SMPL
I6678-02LX5 - 1		33	0.0001	-7470A	PPB	15 Dec 2017 13:22:55	U	SMPL
I6678-02A - 1		13955	3.9468	-7470A	PPB	15 Dec 2017 13:25:02	U	SMPL
CCV68 - 1	CCV68	18159	5.1386	-7470A	PPB	15 Dec 2017 13:27:09	U	SMPL
CCB68 - 1	CCB68	-91	-0.0351	-7470A	PPB	15 Dec 2017 13:29:16	U	SMPL