

TOTAL SUSPENDED SOLIDS - SM2540D

SUPERVISOR: apatel

ANALYST: Aparna

Date: 12/15/2017

DATE & TIME IN: 12/15/2017 11:45

OVEN TEMP: 103 °C

Run Number: LB92184

DATE & TIME OUT: 12/15/2017 15:45

OVEN TEMP: 103 °C

BalanceID: WC SC-3

OvenID: WC Oven-2

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Volume (ml)	1st Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	2nd Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Final Empty Dish+Sample weight after 1.5hr drying @103-@105°C (g)	Weight (g)	Result mg/L
1	LB92184BL	LB92184BL	1.4125	1.4125	100	1.4125	1.4125	1.4125	0	0
2	LB92184BS	LB92184BS	1.4925	1.4925	100	1.5455	1.5455	1.5455	0.053	530
3	I6937-01	001M-WILLETS-PT-BLVD	1.3986	1.3986	400	1.4107	1.4107	1.4107	0.0121	30.3
4	I6937-01DUP	001M-WILLETS-PT-BLVD DUP	1.4082	1.4082	400	1.4203	1.4203	1.4203	0.0121	30.3
5	I6937-02	002M-35TH-AVE	1.4111	1.4111	300	1.4215	1.4215	1.4215	0.0104	34.7

A = Sample Volume (ml)

B = Empty Dish Weight (g)

C = Final Empty Dish + Sample weight after 1.5 hr drying @105°C(g)

D = Weight (g)

Weight (g) = C - B

Result mg/L = $\frac{D}{A} \times 1000 \times 1000$

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TSS121517-3

WorkList ID : 106607

Date : 12/15/2017 12:13:59 PM

1892184

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/15/2017	Water	I6937-01	TSS	Cool 4 deg C	TULL01	E52	001M-WILLETS-PT-BLVD	12/14/2017	SM2540 D
12/15/2017	Water	I6937-02	TSS	Cool 4 deg C	TULL01	E52	002M-35TH-AVE	12/14/2017	SM2540 D

Date/Time 12/15/17, 11:20
 Received by: AS
 Relinquished by: CP

Date/Time 12/15/17, 12:00
 Received by: CP
 Relinquished by: AS