


7471A

LB 92187
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42291



A= 0.0000e+000

B= 2.3633e-004

C= -1.3255e-002

Rho= 0.9997376

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.005	-0.005	35	0.000	35				
0.2	0.200	0.188	-0.012	851	0.0 %	851				
2.5	2.500	2.602	0.102	11066	0.0 %	11066				
5.0	5.000	4.845	-0.155	20557	0.0 %	20557				
7.5	7.500	7.589	0.089	32168	0.0 %	32168	0			
10.0	10.000	9.981	-0.019	42291	0.0 %	42291				

LB92187 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 Dec 2017 13:45:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	35	-	0.00007471A	PPB		15 Dec 2017 13:45:47	S	Std
0.2 - 1	50.2	851	-	0.20007471A	PPB		15 Dec 2017 13:48:05	S	Std
2.5 - 1	52.5	11066	-	2.50007471A	PPB		15 Dec 2017 13:50:22	S	Std
5.0 - 1	55	20557	-	5.00007471A	PPB		15 Dec 2017 13:52:40	S	Std
7.5 - 1	57.5	32168	-	7.50007471A	PPB		15 Dec 2017 13:54:57	S	Std
10.0 - 1	510	42291	-	10.00007471A	PPB		15 Dec 2017 13:59:54	S	Std
ICV11 - 1	ICV11	15679	3.6921	-7471A	PPB		15 Dec 2017 14:02:35	U	SMPL
ICB11 - 1	ICB11	-113	-0.0400	-7471A	PPB		15 Dec 2017 14:04:51	U	SMPL
CCV69 - 1	CCV69	20435	4.8161	-7471A	PPB		15 Dec 2017 14:07:07	U	SMPL
CCB69 - 1	CCB69	-91	-0.0348	-7471A	PPB		15 Dec 2017 14:09:23	U	SMPL
CRA - 1	CRA	862	0.1905	-7471A	PPB		15 Dec 2017 14:11:40	U	SMPL
HighStd - 1	HighStd	43434	10.2514	-7471A	PPB		15 Dec 2017 14:13:55	U	SMPL
ChkStd - 1	ChkStd	28725	6.7753	-7471A	PPB		15 Dec 2017 14:16:11	U	SMPL
PB105054BL - 1	PBS	-57	-0.0267	-7471A	PPB		15 Dec 2017 14:18:29	U	SMPL
PB105054BS - 1	LCSS	14663	3.4520	-7471A	PPB		15 Dec 2017 14:20:47	U	SMPL
I6676-05 - 1	OK-03-121217	796	0.1749	-7471A	PPB		15 Dec 2017 14:23:05	U	SMPL
I6679-03 - 1	HR-05-121317-A	1597	0.3642	-7471A	PPB		15 Dec 2017 14:27:39	U	SMPL
I6679-05 - 1	HR-05-121317-B	2474	0.5714	-7471A	PPB		15 Dec 2017 14:29:57	U	SMPL
I6679-07 - 1	HR-05-121317-C	3278	0.7614	-7471A	PPB		15 Dec 2017 14:32:13	U	SMPL
CCV70 - 1	CCV70	20834	4.9104	-7471A	PPB		15 Dec 2017 14:34:29	U	SMPL
CCB70 - 1	CCB70	-100	-0.0369	-7471A	PPB		15 Dec 2017 14:36:44	U	SMPL
I6680-11 - 1	CL-02-121417-A	1358	0.3077	-7471A	PPB		15 Dec 2017 14:39:01	U	SMPL
I6680-11DUP - 1	CL-02-121417-ADUP	1201	0.2706	-7471A	PPB		15 Dec 2017 14:41:17	U	SMPL
I6680-11MS - 1	CL-02-121417-AMS	16428	3.8691	-7471A	PPB		15 Dec 2017 14:43:33	U	SMPL
I6680-11MSD - 1	CL-02-121417-AMSD	16441	3.8722	-7471A	PPB		15 Dec 2017 14:45:49	U	SMPL
I6680-13 - 1	CL-02-121417-B	2514	0.5809	-7471A	PPB		15 Dec 2017 14:48:06	U	SMPL
I6680-15 - 1	CL-02-121417-C	1043	0.2332	-7471A	PPB		15 Dec 2017 14:50:23	U	SMPL
I6680-17 - 1	CL-02-121417-D	1718	0.3928	-7471A	PPB		15 Dec 2017 14:52:41	U	SMPL
I6684-03 - 1	BU-02-121217	729	0.1590	-7471A	PPB		15 Dec 2017 14:54:58	U	SMPL
I6744-05 - 1	PROS-3-E(13-14)	625	0.1344	-7471A	PPB		15 Dec 2017 14:57:16	U	SMPL
I6889-02 - 1	GIS-1(0-3)	1137	0.2554	-7471A	PPB		15 Dec 2017 14:59:32	U	SMPL
CCV71 - 1	CCV71	20864	4.9175	-7471A	PPB		15 Dec 2017 15:01:48	U	SMPL
CCB71 - 1	CCB71	-94	-0.0355	-7471A	PPB		15 Dec 2017 15:04:05	U	SMPL
I6889-04 - 1	T-1(0-3)	1701	0.3887	-7471A	PPB		15 Dec 2017 15:06:22	U	SMPL
I6911-01 - 1	SL-TRANSFORMERS	250	0.0458	-7471A	PPB		15 Dec 2017 15:08:38	U	SMPL
I6919-01 - 1	339	2892	0.6702	-7471A	PPB		15 Dec 2017 15:10:55	U	SMPL
I6680-11LX5 - 1		282	0.0534	-7471A	PPB		15 Dec 2017 15:13:11	U	SMPL
I6680-11A - 1		17271	4.0684	-7471A	PPB		15 Dec 2017 15:15:28	U	SMPL
I6678-01 - 1	JC-01-121317	2437	0.5627	-7471A	PPB		15 Dec 2017 15:17:45	U	SMPL
CCV72 - 1	CCV72	20555	4.8445	-7471A	PPB		15 Dec 2017 15:20:02	U	SMPL
CCB72 - 1	CCB72	-109	-0.0390	-7471A	PPB		15 Dec 2017 15:22:19	U	SMPL