

7471A


LB92196

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43005



A= 0.0000e+000

B= 2.3239e-004

C= 1.5289e-002

Rho= 0.9996662

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.009	0.009	-28	0.000	-28				
0.2	0.200	0.210	0.010	840	0.0 %	840				
2.5	2.500	2.587	0.087	11065	0.0 %	11065				
5.0	5.000	4.801	-0.199	20592	0.0 %	20592				
7.5	7.500	7.584	0.084	32568	0.0 %	32568	0			
10.0	10.000	10.009	0.009	43005	0.0 %	43005				

LB92196 INSTRUMENT ID : CV1

Method: 7471A

Operator: Admin

Date of Analysis: 15 Dec 2017 15:48:15

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-28	-	0.0000	7471A	PPB	15 Dec 2017 15:49:48	S	Std
0.2 - 1	50.2	840	-	0.2000	7471A	PPB	15 Dec 2017 15:52:05	S	Std
2.5 - 1	52.5	11065	-	2.5000	7471A	PPB	15 Dec 2017 15:54:23	S	Std
5.0 - 1	55	20592	-	5.0000	7471A	PPB	15 Dec 2017 15:56:41	S	Std
7.5 - 1	57.5	32568	-	7.5000	7471A	PPB	15 Dec 2017 15:58:58	S	Std
10.0 - 1	510	43005	-	10.0000	7471A	PPB	15 Dec 2017 16:01:15	S	Std
ICV12 - 1	ICV12	15735	3.6720	-7471A	PPB	PPB	15 Dec 2017 16:03:57	U	SMPL
ICB12 - 1	ICB12	-132	-0.0154	-7471A	PPB	PPB	15 Dec 2017 16:06:13	U	SMPL
CCV73 - 1	CCV73	20654	4.8151	-7471A	PPB	PPB	15 Dec 2017 16:08:28	U	SMPL
CCB73 - 1	CCB73	-102	-0.0084	-7471A	PPB	PPB	15 Dec 2017 16:10:44	U	SMPL
CRA - 1	CRA	848	0.2124	-7471A	PPB	PPB	15 Dec 2017 16:13:01	U	SMPL
HighStd - 1	HighStd	43088	10.0287	-7471A	PPB	PPB	15 Dec 2017 16:15:17	U	SMPL
ChkStd - 1	ChkStd	28788	6.7054	-7471A	PPB	PPB	15 Dec 2017 16:17:33	U	SMPL
PB105015BL - 1	PBS	-59	0.0016	-7471A	PPB	PPB	15 Dec 2017 16:19:51	U	SMPL
PB105015BS - 1	LCSS	14735	3.4396	-7471A	PPB	PPB	15 Dec 2017 16:22:09	U	SMPL
I6883-01 - 1	EP-01	296	0.0841	-7471A	PPB	PPB	15 Dec 2017 16:24:27	U	SMPL
I6888-01 - 1	EPE-122-2(0-5)	14841	3.4642	-7471A	PPB	PPB	15 Dec 2017 16:26:44	U	SMPL
I6888-01DUP - 1	EPE-122-2(0-5)DUP	15272	3.5644	-7471A	PPB	PPB	15 Dec 2017 16:29:01	U	SMPL
I6888-01MS - 1	EPE-122-2(0-5)MS	30721	7.1547	-7471A	PPB	PPB	15 Dec 2017 16:35:47	U	SMPL
I6888-01MSD - 1	EPE-122-2(0-5)MSD	30352	7.0689	-7471A	PPB	PPB	15 Dec 2017 16:38:03	U	SMPL
CCV74 - 1	CCV74	20320	4.7375	-7471A	PPB	PPB	15 Dec 2017 16:40:21	U	SMPL
CCB74 - 1	CCB74	-132	-0.0154	-7471A	PPB	PPB	15 Dec 2017 16:42:38	U	SMPL
I6888-02 - 1	EPE-122-2(5-10)	258	0.0752	-7471A	PPB	PPB	15 Dec 2017 16:44:55	U	SMPL
I6888-03 - 1	EPE-122-2(25-27)	185	0.0583	-7471A	PPB	PPB	15 Dec 2017 16:47:11	U	SMPL
I6888-04 - 1	EPE-122-2(30-32)	110	0.0409	-7471A	PPB	PPB	15 Dec 2017 16:49:27	U	SMPL
I6888-05 - 1	EPE-122-2(40-42)	631	0.1619	-7471A	PPB	PPB	15 Dec 2017 16:51:43	U	SMPL
I6888-06 - 1	EPE-122-2(45-47)	106	0.0399	-7471A	PPB	PPB	15 Dec 2017 16:54:00	U	SMPL
I6888-07 - 1	EPE-122-2(50-52)	465	0.1234	-7471A	PPB	PPB	15 Dec 2017 16:56:16	U	SMPL
I6888-08 - 1	EPE-122-2(55-57)	781	0.1968	-7471A	PPB	PPB	15 Dec 2017 16:58:34	U	SMPL
I6888-09 - 1	EPE-122-2(62-64)	824	0.2068	-7471A	PPB	PPB	15 Dec 2017 17:00:52	U	SMPL
I6888-10 - 1	EPE-122-2(65-67)	500	0.1315	-7471A	PPB	PPB	15 Dec 2017 17:03:09	U	SMPL
I6888-11 - 1	EPE-122-2(70-72)	245	0.0722	-7471A	PPB	PPB	15 Dec 2017 17:05:26	U	SMPL
CCV75 - 1	CCV75	20556	4.7924	-7471A	PPB	PPB	15 Dec 2017 17:07:42	U	SMPL
CCB75 - 1	CCB75	-107	-0.0096	-7471A	PPB	PPB	15 Dec 2017 17:09:58	U	SMPL
I6888-12 - 1	EPE-122-2(75-77)	356	0.0980	-7471A	PPB	PPB	15 Dec 2017 17:12:15	U	SMPL
I6908-01 - 1	1	315	0.0885	-7471A	PPB	PPB	15 Dec 2017 17:14:32	U	SMPL
I6908-03 - 1	2	1846	0.4443	-7471A	PPB	PPB	15 Dec 2017 17:16:48	U	SMPL
I6908-05 - 1	3	844	0.2114	-7471A	PPB	PPB	15 Dec 2017 17:19:05	U	SMPL
I6908-07 - 1	4	2839	0.6751	-7471A	PPB	PPB	15 Dec 2017 17:21:21	U	SMPL
I6908-09 - 1	5	2027	0.4864	-7471A	PPB	PPB	15 Dec 2017 17:23:38	U	SMPL
I6908-11 - 1	6	621	0.1596	-7471A	PPB	PPB	15 Dec 2017 17:25:55	U	SMPL
I6908-13 - 1	7	3287	0.7792	-7471A	PPB	PPB	15 Dec 2017 17:28:12	U	SMPL
I6888-01LX5 - 1		2805	0.6672	-7471A	PPB	PPB	15 Dec 2017 17:30:30	U	SMPL
I6888-01A - 1		31287	7.2862	-7471A	PPB	PPB	15 Dec 2017 17:32:46	U	SMPL
CCV76 - 1	CCV76	20463	4.7708	-7471A	PPB	PPB	15 Dec 2017 17:35:03	U	SMPL
CCB76 - 1	CCB76	-104	-0.0089	-7471A	PPB	PPB	15 Dec 2017 17:37:21	U	SMPL