



Analytical Summary Report

Analysis Method: SM 4500-PH B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92201

Temp.: 18.1 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/15/2017	15:00
2	CAL2	1	Water			6.99	12/15/2017	15:01
3	CAL3	1	Water			10.02	12/15/2017	15:02
4	ICV	1	Water			7.00	12/15/2017	15:03
5	CCV1	1	Water			2.01	12/15/2017	15:05
6	I6945-01	1	Water			4.65	12/15/2017	15:06
7	I6945-01DUP	1	Water			4.66	12/15/2017	15:07
8	I6947-01	1	Water			7.39	12/15/2017	15:09
9	CCV2	1	Water			12.01	12/15/2017	15:10

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6945

WorkList ID : 106621

Date : 12/15/2017 2:41:07 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/13/2017	Water	I6945-01-17	pH	Cool 4 deg C	STER14	E52	MINI-CELL	12/13/2017	SM 4500-PH B
12/13/2017	Water	I6947-01-17	pH	Cool 4 deg C	STER14	E52	JS-SYSTEM	12/13/2017	SM 4500-PH B

Date/Time

Received by:

Relinquished by:

12-15-17 3:00 PM

AP

CD

Date/Time

Received by:

Relinquished by:

12-15-17 4:20 PM

CD

AP