

ISM02.4


LB 92206

INSTRUMENT ID: C42

Linear

 μ Abs.:

43644



A= 0.0000e+000

B= 2.2881e-004

C= -7.1985e-003

Rho= 0.9999791

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0
0.00	0.000	-0.004	-0.004	15	0.000	15					1
0.2	0.200	0.206	0.006	932	0.0 %	932					3
2.5	2.500	2.501	0.001	10963	0.0 %	10963					0
5.0	5.000	4.972	-0.028	21759	0.0 %	21759					-1
7.5	7.500	7.546	0.046	33010	0.0 %	33010					1
10.0	10.000	9.979	-0.021	43644	0.0 %	43644					0

LB
12/15/12

LB92206
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	15	-		0 ISM02.4	PPB	12/15/2017 17:55	Std
	0.2 S0.2	932	-		0.2 ISM02.4	PPB	12/15/2017 17:58	Std
	2.5 S2.5	10963	-		2.5 ISM02.4	PPB	12/15/2017 18:00	Std
	5 S5	21759	-		5 ISM02.4	PPB	12/15/2017 18:02	Std
	7.5 S7.5	33010	-		7.5 ISM02.4	PPB	12/15/2017 18:04	Std
	10 S10	43644	-		10 ISM02.4	PPB	12/15/2017 18:09	Std
ICV	ICV	16976	3.8771 -		ISM02.4	PPB	12/15/2017 18:15	SMPL
ICB	ICB	-78	-0.025 -		ISM02.4	PPB	12/15/2017 18:17	SMPL
CCV077	CCV077	21968	5.0193 -		ISM02.4	PPB	12/15/2017 18:19	SMPL
CCB077	CCB077	-120	-0.0347 -		ISM02.4	PPB	12/15/2017 18:22	SMPL
PB105068BL	PBS004	131	0.0228 -		ISM02.4	PPB	12/15/2017 18:24	SMPL
I6735-01	MA4Z49	997	0.2209 -		ISM02.4	PPB	12/15/2017 18:26	SMPL
I6735-02	MA4Z50	50339	11.511 -		ISM02.4	PPB	12/15/2017 18:28	SMPL
I6735-03	MA4Z51	5909	1.3448 -		ISM02.4	PPB	12/15/2017 18:31	SMPL
I6735-04	MA4Z52	8498	1.9372 -		ISM02.4	PPB	12/15/2017 18:33	SMPL
I6735-05	MA4Z53	3984	0.9044 -		ISM02.4	PPB	12/15/2017 18:35	SMPL
I6735-06	MA4Z54	5721	1.3018 -		ISM02.4	PPB	12/15/2017 18:38	SMPL
I6735-07	MA4Z54S	16701	3.8142 -		ISM02.4	PPB	12/15/2017 18:40	SMPL
I6735-08	MA4Z54D	5647	1.2849 -		ISM02.4	PPB	12/15/2017 18:42	SMPL
PB105069BL	PBS004	92	0.0139 -		ISM02.4	PPB	12/15/2017 18:44	SMPL
I6774-01	MA4Z99	1941	0.4369 -		ISM02.4	PPB	12/15/2017 18:47	SMPL
I6774-02	MA4ZA0	2942	0.666 -		ISM02.4	PPB	12/15/2017 18:49	SMPL
I6774-03	MA4ZA1	19925	4.5519 -		ISM02.4	PPB	12/15/2017 18:51	SMPL
I6774-04	MA4ZA2	21937	5.0122 -		ISM02.4	PPB	12/15/2017 18:53	SMPL
I6774-05	MA4ZA3	3121	0.7069 -		ISM02.4	PPB	12/15/2017 18:56	SMPL
I6774-06	MA4ZA4	5985	1.3622 -		ISM02.4	PPB	12/15/2017 18:58	SMPL
I6774-07	MA4ZA5	3391	0.7687 -		ISM02.4	PPB	12/15/2017 19:00	SMPL
I6774-08	MA4ZA6	3203	0.7257 -		ISM02.4	PPB	12/15/2017 19:03	SMPL
I6774-09	MA4ZA7	6983	1.5906 -		ISM02.4	PPB	12/15/2017 19:05	SMPL
I6774-10	MA4ZA8	5053	1.149 -		ISM02.4	PPB	12/15/2017 19:07	SMPL
CCV078	CCV078	21706	4.9594 -		ISM02.4	PPB	12/15/2017 19:09	SMPL
CCB078	CCB078	-102	-0.0305 -		ISM02.4	PPB	12/15/2017 19:12	SMPL
I6774-11	MA4ZA9	2325	0.5248 -		ISM02.4	PPB	12/15/2017 19:14	SMPL
I6774-12	MA4ZB0	5647	1.2849 -		ISM02.4	PPB	12/15/2017 19:16	SMPL
I6774-13	MA4ZB1	2452	0.5538 -		ISM02.4	PPB	12/15/2017 19:19	SMPL
I6774-14	MA4ZB5	13705	3.1287 -		ISM02.4	PPB	12/15/2017 19:21	SMPL
I6774-15	MA4ZB6	-29	-0.0138 -		ISM02.4	PPB	12/15/2017 19:23	SMPL
I6774-17	MA4ZC6	3009	0.6813 -		ISM02.4	PPB	12/15/2017 19:25	SMPL
I6774-18	MA4ZC7	2552	0.5767 -		ISM02.4	PPB	12/15/2017 19:28	SMPL
I6774-19	MA4ZC8	8800	2.0063 -		ISM02.4	PPB	12/15/2017 19:30	SMPL
I6774-20	MA4ZC8D	8405	1.916 -		ISM02.4	PPB	12/15/2017 19:32	SMPL
I6774-21	MA4ZC8S	18916	4.321 -		ISM02.4	PPB	12/15/2017 19:38	SMPL
I6774-22	MA4ZC9	25690	5.871 -		ISM02.4	PPB	12/15/2017 19:40	SMPL
I6774-23	MA4ZD0	14248	3.2529 -		ISM02.4	PPB	12/15/2017 19:42	SMPL
I6774-24	MA4ZD1	16200	3.6996 -		ISM02.4	PPB	12/15/2017 19:44	SMPL
PB105071BL	PBS004	149	0.0269 -		ISM02.4	PPB	12/15/2017 19:47	SMPL

LB92206
INSTRUMENT ID : CV1

I6800-03	MC0298	454	0.0967 -	ISM02.4	PPB	12/15/2017 19:49 SMPL
I6800-04	MC02A1	2842	0.6431 -	ISM02.4	PPB	12/15/2017 19:51 SMPL
I6800-09	MC0299	976	0.2161 -	ISM02.4	PPB	12/15/2017 19:54 SMPL
I6800-10	MC02B3	9527	2.1727 -	ISM02.4	PPB	12/15/2017 20:01 SMPL
I6800-11	MC02B3D	9572	2.183 -	ISM02.4	PPB	12/15/2017 20:03 SMPL
I6800-12	MC02B3S	19495	4.4535 -	ISM02.4	PPB	12/15/2017 20:06 SMPL
CCV079	CCV079	21989	5.0241 -	ISM02.4	PPB	12/15/2017 20:08 SMPL
CCB079	CCB079	-54	-0.0196 -	ISM02.4	PPB	12/15/2017 20:13 SMPL
I6800-14	MC02B7	1383	0.3092 -	ISM02.4	PPB	12/15/2017 20:15 SMPL
I6800-16	MC02E1	1062	0.2358 -	ISM02.4	PPB	12/15/2017 20:17 SMPL
PB105072BL	PBS004	225	0.0443 -	ISM02.4	PPB	12/15/2017 20:20 SMPL
I6807-01	MA4ZD2	46762	10.6925 -	ISM02.4	PPB	12/15/2017 20:22 SMPL
I6807-02	MA4ZD3	18377	4.1977 -	ISM02.4	PPB	12/15/2017 20:24 SMPL
I6807-03	MA4ZD4	12198	2.7838 -	ISM02.4	PPB	12/15/2017 20:27 SMPL
I6807-04	MA4ZD5	11911	2.7182 -	ISM02.4	PPB	12/15/2017 20:29 SMPL
I6807-05	MA4ZD6	5074	1.1538 -	ISM02.4	PPB	12/15/2017 20:31 SMPL
I6807-06	MA4ZD7	391	0.0823 -	ISM02.4	PPB	12/15/2017 20:33 SMPL
I6807-07	MA4ZD8	9798	2.2347 -	ISM02.4	PPB	12/15/2017 20:36 SMPL
I6807-08	MA4ZE2	-28	-0.0136 -	ISM02.4	PPB	12/15/2017 20:38 SMPL
I6807-09	MA4ZE3	14071	3.2124 -	ISM02.4	PPB	12/15/2017 20:40 SMPL
I6807-10	MA4ZE4	4923	1.1192 -	ISM02.4	PPB	12/15/2017 20:43 SMPL
I6807-11	MA4ZE5	9453	2.1558 -	ISM02.4	PPB	12/15/2017 20:45 SMPL
I6807-12	MA4ZE6	765	0.1678 -	ISM02.4	PPB	12/15/2017 20:47 SMPL
I6807-13	MA4ZE7	545	0.1175 -	ISM02.4	PPB	12/15/2017 20:49 SMPL
I6807-14	MA4ZE8	10870	2.48 -	ISM02.4	PPB	12/15/2017 20:57 SMPL
I6807-15	MA4ZE8D	10919	2.4912 -	ISM02.4	PPB	12/15/2017 20:59 SMPL
I6807-16	MA4ZE8S	21743	4.9679 -	ISM02.4	PPB	12/15/2017 21:01 SMPL
CCV080	CCV080	21736	4.9663 -	ISM02.4	PPB	12/15/2017 21:03 SMPL
CCB080	CCB080	-102	-0.0305 -	ISM02.4	PPB	12/15/2017 21:06 SMPL
I6735-02DLX2	MA4Z50	25284	5.7781 -	ISM02.4	PPB	12/15/2017 21:08 SMPL
I6807-01DLX2	MA4ZD2	23225	5.307 -	ISM02.4	PPB	12/15/2017 21:10 SMPL
CCV081	CCV081	21524	4.9177 -	ISM02.4	PPB	12/15/2017 21:31 SMPL
CCB081	CCB081	-106	-0.0315 -	ISM02.4	PPB	12/15/2017 21:33 SMPL