



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92207

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/15/2017	18:20
2	CAL2	1	Water			7.01	12/15/2017	18:21
3	CAL3	1	Water			10.04	12/15/2017	18:22
4	ICV	1	Water			6.99	12/15/2017	18:24
5	CCV1	1	Water			2.01	12/15/2017	18:25
6	I6675-08	1	Solid	20.02	20	7.78	12/15/2017	18:26
7	I6675-08DUP	1	Solid	20.03	20	7.79	12/15/2017	18:29
8	I6677-06	1	Solid	20.02	20	7.93	12/15/2017	18:32
9	I6677-08	1	Solid	20.04	20	7.36	12/15/2017	18:34
10	I6680-20	1	Solid	20.01	20	8.52	12/15/2017	18:35
11	I6927-03	1	Solid	20.02	20	8.18	12/15/2017	18:36
12	I6940-03	1	Solid	20.01	20	11.00	12/15/2017	18:37
13	I6957-05	1	Solid	20.02	20	12.01	12/15/2017	18:39
14	I6957-06	1	Solid	20.04	20	9.40	12/15/2017	18:42
15	I6957-07	1	Solid	20.01	20	10.16	12/15/2017	18:44
16	CCV2	1	Water			2.02	12/15/2017	18:45
17	I6957-08	1	Solid	20.05	20	9.16	12/15/2017	18:46
18	I6958-02	1	Solid	20.03	20	6.26	12/15/2017	18:47
19	CCV3	1	Water			12.02	12/15/2017	18:49

WORKLIST(Hardcopy Internal Chain)

2392207

WorkList Name : CORROSIONITY I6940

WorkList ID : 106624

Date : 12/15/2017 2:43:24 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/15/2017	Solid	I6675-08	Corrosivity	Cool 4 deg C	PSEG05		SU-02-121517	12/15/2017	9045C
12/15/2017	Solid	I6677-06	Corrosivity	Cool 4 deg C	PSEG05		OR-01-121517-A	12/15/2017	9045C
12/15/2017	Solid	I6677-08	Corrosivity	Cool 4 deg C	PSEG05		OR-01-121517-B	12/15/2017	9045C
12/15/2017	Solid	I6680-20	Corrosivity	Cool 4 deg C	PSEG05		PATERSON-02-121517	12/15/2017	9045C
12/15/2017	Solid	I6927-03	Corrosivity	Cool 4 deg C	PSEG01	D52	R-4297-39456	12/15/2017	9045C
12/15/2017	Solid	I6940-03	Corrosivity	Cool 4 deg C	PSEG01	D52	VB16206-VB16211	12/15/2017	9045C
12/14/2017	Solid	I6957-05	Corrosivity	Cool 4 deg C	ENTA05	E41	HALSTED-NW	12/14/2017	9045C
12/14/2017	Solid	I6957-06	Corrosivity	Cool 4 deg C	ENTA05	E41	HALSTED-SW	12/14/2017	9045C
12/14/2017	Solid	I6957-07	Corrosivity	Cool 4 deg C	ENTA05	E41	HALSTED-EW	12/14/2017	9045C
12/14/2017	Solid	I6957-08	Corrosivity	Cool 4 deg C	ENTA05	E41	HALSTED-WW	12/14/2017	9045C
12/15/2017	Solid	I6958-02	Corrosivity	Cool 4 deg C	PSEG01	D52	DRUM-1-4	12/15/2017	9045C

Date/Time 12-15-17 2:40 PM
 Received by: JD
 Relinquished by: UN

Date/Time 12-15-17 7:00 PM
 Received by: JP
 Relinquished by: CR