



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92208

Temp.: 23.5 °C

BalanceID: WC SC-4

SLOPE.: 98.4

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/15/2017	18:20
2	CAL2	1	Water			7.01	12/15/2017	18:21
3	CAL3	1	Water			10.04	12/15/2017	18:22
4	ICV	1	Water			6.99	12/15/2017	18:24
5	CCV1	1	Water			2.01	12/15/2017	18:25
6	I6957-01	1	Solid	20.02	20	12.01	12/15/2017	18:39
7	I6957-01DUP	1	Solid	20.03	20	12.02	12/15/2017	18:40
8	I6957-02	1	Solid	20.04	20	9.40	12/15/2017	18:41
9	I6957-03	1	Solid	20.01	20	10.16	12/15/2017	18:42
10	I6957-04	1	Solid	20.05	20	9.16	12/15/2017	18:45
11	CCV2	1	Water			12.02	12/15/2017	18:49

1592208

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph i6957

WorkList ID : 106643

Date : 12/15/2017 5:46:42 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/14/2017	Solid	I6957-01	pH	Cool 4 deg C	ENTA05	E41	HALSTED-NW	12/14/2017	9045C
12/14/2017	Solid	I6957-02	pH	Cool 4 deg C	ENTA05	E41	HALSTED-SW	12/14/2017	9045C
12/14/2017	Solid	I6957-03	pH	Cool 4 deg C	ENTA05	E41	HALSTED-EW	12/14/2017	9045C
12/14/2017	Solid	I6957-04	pH	Cool 4 deg C	ENTA05	E41	HALSTED-VWV	12/14/2017	9045C

Date/Time 12.15.17 2:14PM
Received by: JH
Relinquished by: CS

Date/Time 12.15.17 7:10PM
Received by: CP
Relinquished by: JP