


7470A

LB 92219
INSTRUMENT ID: CV1

Linear

 μ Abs.:

36850



A= 0.0000e+000

B= 2.7077e-004

C= 2.8095e-002

Rho= 0.9998698

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.033	0.033	18	0.000	18				
0.2	0.200	0.216	0.016	694	0.0 %	694				
2.5	2.500	2.495	-0.005	9111	0.0 %	9111				
5.0	5.000	4.879	-0.121	17914	0.0 %	17914				
7.5	7.500	7.572	0.072	27860	0.0 %	27860				
10.0	10.000	10.006	0.006	36850	0.0 %	36850				

LB92219 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Dec 2017 09:46:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	18	-	0.0000	7470A	PPB	18 Dec 2017 09:46:27	S	Std
0.2 - 1	50.2	694	-	0.2000	7470A	PPB	18 Dec 2017 09:48:35	S	Std
2.5 - 1	5225	9111	-	2.5000	7470A	PPB	18 Dec 2017 09:50:43	S	Std
5.0 - 1	55	17914	-	5.0000	7470A	PPB	18 Dec 2017 09:52:52	S	Std
7.5 - 1	57.5	27860	-	7.5000	7470A	PPB	18 Dec 2017 09:55:00	S	Std
10.0 - 1	510	36850	-	10.0000	7470A	PPB	18 Dec 2017 09:59:32	S	Std
ICV13 - 1	ICV13	13752	3.7517	-	7470A	PPB	18 Dec 2017 10:04:29	U	SMPL
ICB13 - 1	ICB13	-76	0.0075	-	7470A	PPB	18 Dec 2017 10:06:35	U	SMPL
CCV66 - 1	CCV66	18404	5.0113	-	7470A	PPB	18 Dec 2017 10:08:42	U	SMPL
CCB66 - 1	CCB66	-71	0.0089	-	7470A	PPB	18 Dec 2017 10:10:48	U	SMPL
CRA - 1	CRA	707	0.2195	-	7470A	PPB	18 Dec 2017 10:12:55	U	SMPL
HighStd - 1	HighStd	36902	10.0199	-	7470A	PPB	18 Dec 2017 10:15:02	U	SMPL
ChkStd - 1	ChkStd	24919	6.7753	-	7470A	PPB	18 Dec 2017 10:17:09	U	SMPL
PB105100BL - 1	PBW	-145	-0.0112	-	7470A	PPB	18 Dec 2017 10:19:19	U	SMPL
PB105100BS - 1	LCSW PB105049 RB	13475	3.6767	-	7470A	PPB	18 Dec 2017 10:21:28	U	SMPL
PB105100TB - 1	PB105100TB	-52	0.0140	-	7470A	PPB	18 Dec 2017 10:23:35	U	SMPL
I6680-12 - 1	CL-02-121417-A	-6	0.0265	-	7470A	PPB	18 Dec 2017 10:25:43	U	SMPL
I6680-14 - 1	CL-02-121417-B	24	0.0346	-	7470A	PPB	18 Dec 2017 10:27:51	U	SMPL
I6680-16 - 1	CL-02-121417-C	45	0.0403	-	7470A	PPB	18 Dec 2017 10:30:00	U	SMPL
I6680-18 - 1	CL-02-121417-D	37	0.0381	-	7470A	PPB	18 Dec 2017 10:32:06	U	SMPL
CCV67 - 1	CCV67	18359	4.9991	-	7470A	PPB	18 Dec 2017 10:34:13	U	SMPL
CCB67 - 1	CCB67	-89	0.0040	-	7470A	PPB	18 Dec 2017 10:36:20	U	SMPL
I6744-12 - 1	PROS- 3-E(11-12)	0	0.0281	-	7470A	PPB	18 Dec 2017 10:38:27	U	SMPL
I6911-02 - 1	SL-TRANSFORMERS	7	0.0300	-	7470A	PPB	18 Dec 2017 10:40:34	U	SMPL
I6916-01 - 1	OILY-STONES	52	0.0422	-	7470A	PPB	18 Dec 2017 10:42:41	U	SMPL
I6916-02 - 1	OILY-DEBRIS	33	0.0370	-	7470A	PPB	18 Dec 2017 10:44:48	U	SMPL
I6916-03 - 1	OIL-DRUM	22	0.0341	-	7470A	PPB	18 Dec 2017 10:46:55	U	SMPL
I6919-02 - 1	339	13	0.0316	-	7470A	PPB	18 Dec 2017 10:49:03	U	SMPL
I6919-02DUP - 1	339DUP	28	0.0357	-	7470A	PPB	18 Dec 2017 10:51:11	U	SMPL
I6919-02MS - 1	339MS	12877	3.5147	-	7470A	PPB	18 Dec 2017 10:53:20	U	SMPL
I6919-02MSD - 1	339MSD	13535	3.6929	-	7470A	PPB	18 Dec 2017 10:55:28	U	SMPL
CCV68 - 1	CCV68	18322	4.9891	-	7470A	PPB	18 Dec 2017 10:57:35	U	SMPL
CCB68 - 1	CCB68	-90	0.0037	-	7470A	PPB	18 Dec 2017 10:59:42	U	SMPL
I6919-02LX5 - 1		21	0.0338	-	7470A	PPB	18 Dec 2017 11:01:50	U	SMPL
I6919-02A - 1		14288	3.8968	-	7470A	PPB	18 Dec 2017 11:03:57	U	SMPL
CCV69 - 1	CCV69	18486	5.0335	-	7470A	PPB	18 Dec 2017 11:06:04	U	SMPL
CCB69 - 1	CCB69	-81	0.0062	-	7470A	PPB	18 Dec 2017 11:08:11	U	SMPL