


245.1

LB92220
INSTRUMENT ID: CY2

Linear

 μ Abs.:

33993



A= 0.0000e+000

B= 2.9489e-004

C= 2.9096e-003

Rho= 0.9999223

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.004	0.004	3	0.000	3				
0.05	0.050	0.062	0.012	202	0.0 %	202				
0.2	0.200	0.215	0.015	719	0.0 %	719				
2.5	2.500	2.526	0.026	8557	0.0 %	8557				
5.0	5.000	4.888	-0.112	16567	0.0 %	16567				
7.5	7.500	7.527	0.027	25514	0.0 %	25514				
10.0	10.000	10.027	0.027	33993	0.0 %	33993				

LB92220 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 18 Dec 2017 13:04:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	3	-	0.0000	245.1	PPB	18 Dec 2017 13:05:15	S	Std
0.05 - 1	50.05	202	-	0.0500	245.1	PPB	18 Dec 2017 13:07:24	S	Std
0.2 - 1	50.2	719	-	0.2000	245.1	PPB	18 Dec 2017 13:09:32	S	Std
2.5 - 1	52.5	8557	-	2.5000	245.1	PPB	18 Dec 2017 13:11:41	S	Std
5.0 - 1	55	16567	-	5.0000	245.1	PPB	18 Dec 2017 13:13:48	S	Std
7.5 - 1	57.5	25514	-	7.5000	245.1	PPB	18 Dec 2017 13:15:56	S	Std
10.0 - 1	510	33993	-	10.0000	245.1	PPB	18 Dec 2017 13:30:12	S	Std
ICV14 - 1	ICV14	13221	3.9017	-245.1	PPB	PPB	18 Dec 2017 13:32:57	U	SMPL
ICB14 - 1	ICB14	-83	-0.0216	-245.1	PPB	PPB	18 Dec 2017 13:35:03	U	SMPL
CCV70 - 1	CCV70	16712	4.9312	-245.1	PPB	PPB	18 Dec 2017 13:37:10	U	SMPL
CCB70 - 1	CCB70	-86	-0.0225	-245.1	PPB	PPB	18 Dec 2017 13:39:16	U	SMPL
CRA - 1	CRA	711	0.2126	-245.1	PPB	PPB	18 Dec 2017 13:41:24	U	SMPL
HighStd - 1	HighStd	33864	9.9892	-245.1	PPB	PPB	18 Dec 2017 13:43:30	U	SMPL
ChkStd - 1	ChkStd	24432	7.2078	-245.1	PPB	PPB	18 Dec 2017 13:45:37	U	SMPL
PB105052BL - 1	PBW	-156	-0.0431	-245.1	PPB	PPB	18 Dec 2017 13:47:46	U	SMPL
PB105052BS - 1	LCSW	13566	4.0034	-245.1	PPB	PPB	18 Dec 2017 13:49:54	U	SMPL
I6880-01 - 1	120817-TIDO-F-D-B	-60	-0.0148	-245.1	PPB	PPB	18 Dec 2017 13:52:02	U	SMPL
I6880-02 - 1	120817-TIDO-F-D-S	33	0.0126	-245.1	PPB	PPB	18 Dec 2017 13:54:10	U	SMPL
I6880-03 - 1	120817-TIDO-F-U-B	31	0.0121	-245.1	PPB	PPB	18 Dec 2017 13:56:18	U	SMPL
I6880-04 - 1	120817-TIDO-F-U-S	39	0.0144	-245.1	PPB	PPB	18 Dec 2017 13:58:27	U	SMPL
I6880-05 - 1	120817-FOPI-F-D-B	34	0.0129	-245.1	PPB	PPB	18 Dec 2017 14:00:33	U	SMPL
CCV71 - 1	CCV71	16657	4.9150	-245.1	PPB	PPB	18 Dec 2017 14:02:40	U	SMPL
CCB71 - 1	CCB71	-81	-0.0210	-245.1	PPB	PPB	18 Dec 2017 14:04:47	U	SMPL
I6880-06 - 1	120817-FOPI-F-D-S	35	0.0132	-245.1	PPB	PPB	18 Dec 2017 14:06:54	U	SMPL
I6880-07 - 1	120817-FOPI-F-U-B	28	0.0112	-245.1	PPB	PPB	18 Dec 2017 14:09:01	U	SMPL
I6880-08 - 1	120817-FOPI-F-U-S	46	0.0165	-245.1	PPB	PPB	18 Dec 2017 14:11:07	U	SMPL
I6880-09 - 1	120817-JETT-F-D-B	44	0.0159	-245.1	PPB	PPB	18 Dec 2017 14:13:14	U	SMPL
I6880-10 - 1	120817-JETT-F-D-M	60	0.0206	-245.1	PPB	PPB	18 Dec 2017 14:15:22	U	SMPL
I6880-11 - 1	120817-JETT-F-D-S	43	0.0156	-245.1	PPB	PPB	18 Dec 2017 14:17:30	U	SMPL
I6880-12 - 1	120817-JETT-F-U-B	57	0.0197	-245.1	PPB	PPB	18 Dec 2017 14:19:37	U	SMPL
I6880-13 - 1	120817-JETT-F-U-M	75	0.0250	-245.1	PPB	PPB	18 Dec 2017 14:21:46	U	SMPL
I6880-14 - 1	120817-JETT-F-U-S	48	0.0171	-245.1	PPB	PPB	18 Dec 2017 14:23:54	U	SMPL
CCV72 - 1	CCV72	16450	4.8539	-245.1	PPB	PPB	18 Dec 2017 14:28:08	U	SMPL
CCB72 - 1	CCB72	-81	-0.0210	-245.1	PPB	PPB	18 Dec 2017 14:30:15	U	SMPL
I6886-01DUP - 1	SEEP-DRAINS-1DUP	101	0.0327	-245.1	PPB	PPB	18 Dec 2017 14:32:23	U	SMPL
I6886-01MS - 1	SEEP-DRAINS-1MS	12983	3.8315	-245.1	PPB	PPB	18 Dec 2017 14:34:30	U	SMPL
I6886-01MSD - 1	SEEP-DRAINS-1MSD	12740	3.7599	-245.1	PPB	PPB	18 Dec 2017 14:36:37	U	SMPL
I6886-02 - 1	SEEP-DRAINS-2	-59	-0.0145	-245.1	PPB	PPB	18 Dec 2017 14:38:44	U	SMPL
I6886-03 - 1	SEEP-DRAINS-3	28	0.0112	-245.1	PPB	PPB	18 Dec 2017 14:40:52	U	SMPL
I6886-04 - 1	SEEP-DRAINS-4	26	0.0106	-245.1	PPB	PPB	18 Dec 2017 14:42:59	U	SMPL
PB105053BL - 1	PBW	24	0.0100	-245.1	PPB	PPB	18 Dec 2017 14:45:08	U	SMPL
PB105053BS - 1	LCSW	13186	3.8914	-245.1	PPB	PPB	18 Dec 2017 14:47:16	U	SMPL
6882-01 - 1	121117-JETT-E-D-B	-18	-0.0024	-245.1	PPB	PPB	18 Dec 2017 14:49:25	U	SMPL
6882-02 - 1	121117-JETT-E-D-S	54	0.0188	-245.1	PPB	PPB	18 Dec 2017 14:51:32	U	SMPL
CCV73 - 1	CCV73	16582	4.8928	-245.1	PPB	PPB	18 Dec 2017 14:53:39	U	SMPL
CCB73 - 1	CCB73	-90	-0.0236	-245.1	PPB	PPB	18 Dec 2017 14:55:46	U	SMPL
6882-03 - 1	121117-JETT-E-U-B	50	0.0177	-245.1	PPB	PPB	18 Dec 2017 14:57:54	U	SMPL
6882-04 - 1	121117-JETT-E-U-S	53	0.0185	-245.1	PPB	PPB	18 Dec 2017 15:00:01	U	SMPL
6882-05 - 1	121117-TIDO-E-D-M	53	0.0185	-245.1	PPB	PPB	18 Dec 2017 15:02:08	U	SMPL
6882-06 - 1	121117-TIDO-E-U-M	72	0.0241	-245.1	PPB	PPB	18 Dec 2017 15:04:15	U	SMPL
6882-07 - 1	121117-FOPI-E-D-M	54	0.0188	-245.1	PPB	PPB	18 Dec 2017 15:06:22	U	SMPL
6882-08 - 1	121117-FOPI-E-U-M	51	0.0179	-245.1	PPB	PPB	18 Dec 2017 15:08:30	U	SMPL
6882-08DUP - 1	121117-FOPI-E-U-MDUP	38	0.0141	-245.1	PPB	PPB	18 Dec 2017 15:10:38	U	SMPL
6882-08MS - 1	121117-FOPI-E-U-MMS	15613	4.6071	-245.1	PPB	PPB	18 Dec 2017 15:12:46	U	SMPL
6882-08MSD - 1	121117-FOPI-E-U-MMSD	15301	4.5151	-245.1	PPB	PPB	18 Dec 2017 15:14:55	U	SMPL
PB105088BL - 1	PBW	-69	-0.0174	-245.1	PPB	PPB	18 Dec 2017 15:17:02	U	SMPL
CCV74 - 1	CCV74	16036	4.7318	-245.1	PPB	PPB	18 Dec 2017 15:19:09	U	SMPL
CCB74 - 1	CCB74	-69	-0.0174	-245.1	PPB	PPB	18 Dec 2017 15:21:16	U	SMPL
PB105088BS - 1	LCSW	13578	4.0070	-245.1	PPB	PPB	18 Dec 2017 15:23:24	U	SMPL
6909-01 - 1	121217-DBRI-E-D-B	-8	0.0006	-245.1	PPB	PPB	18 Dec 2017 15:25:31	U	SMPL
6909-02 - 1	121217-DBRI-E-D-S	59	0.0203	-245.1	PPB	PPB	18 Dec 2017 15:27:38	U	SMPL
6909-03 - 1	121217-DBRI-E-U-B	62	0.0212	-245.1	PPB	PPB	18 Dec 2017 15:29:46	U	SMPL
6909-04 - 1	121217-DBRI-E-U-S	59	0.0203	-245.1	PPB	PPB	18 Dec 2017 15:31:53	U	SMPL
6909-05 - 1	121217-TIDO-E-D-B	68	0.0230	-245.1	PPB	PPB	18 Dec 2017 15:34:00	U	SMPL
6909-06 - 1	121217-TIDO-E-D-S	66	0.0224	-245.1	PPB	PPB	18 Dec 2017 15:36:08	U	SMPL
6909-07 - 1	121217-TIDO-E-U-B	53	0.0185	-245.1	PPB	PPB	18 Dec 2017 15:38:16	U	SMPL
6909-08 - 1	121217-TIDO-E-U-S	68	0.0230	-245.1	PPB	PPB	18 Dec 2017 15:40:25	U	SMPL
6909-09 - 1	121217-FOPI-E-D-M	115	0.0368	-245.1	PPB	PPB	18 Dec 2017 15:42:32	U	SMPL
CCV75 - 1	CCV75	16504	4.8698	-245.1	PPB	PPB	18 Dec 2017 15:44:40	U	SMPL
CCB75 - 1	CCB75	-55	-0.0133	-245.1	PPB	PPB	18 Dec 2017 15:46:47	U	SMPL
6909-10 - 1	121217-FOPI-E-U-M	87	0.0286	-245.1	PPB	PPB	18 Dec 2017 15:48:55	U	SMPL
6910-01 - 1	121317-DBRI-E-D-M	73	0.0244	-245.1	PPB	PPB	18 Dec 2017 15:51:03	U	SMPL
6910-02 - 1	121317-DBRI-E-U-M	54	0.0188	-245.1	PPB	PPB	18 Dec 2017 15:53:10	U	SMPL
6910-03 - 1	121317-TIDO-E-D-M	57	0.0197	-245.1	PPB	PPB	18 Dec 2017 15:55:17	U	SMPL
6910-04 - 1	121317-TIDO-E-U-M	67	0.0227	-245.1	PPB	PPB	18 Dec 2017 15:57:25	U	SMPL
6910-04DUP - 1	121317-TIDO-E-U-MDUP	88	0.0289	-245.1	PPB	PPB	18 Dec 2017 15:59:33	U	SMPL
6910-04MS - 1	121317-TIDO-E-U-MMS	15761	4.6507	-245.1	PPB	PPB	18 Dec 2017 16:01:42	U	SMPL
6910-04MSD - 1	121317-TIDO-E-U-MMSD	15732	4.6422	-245.1	PPB	PPB	18 Dec 2017 16:03:50	U	SMPL
6886-01 - 1	SEEP-DRAINS-1	95	0.0309	-245.1	PPB	PPB	18 Dec 2017 16:05:59	U	SMPL
CV76 - 1	CCV76	15911	4.6950	-245.1	PPB	PPB	18 Dec 2017 16:08:06	U	SMPL
CB76 - 1	CCB76	-67	-0.0168	-245.1	PPB	PPB	18 Dec 2017 16:10:14	U	SMPL

LB92220 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 18 Dec 2017 13:04:22

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
I6886-01LX5 - 1		25	0.0103		-245.1	PPB	18 Dec 2017 16:12:23	U	SMPL
I6886-01A - 1		13417	3.9595		-245.1	PPB	18 Dec 2017 16:14:30	U	SMPL
I6882-08LX5 - 1		-70	-0.0177		-245.1	PPB	18 Dec 2017 16:16:38	U	SMPL
I6882-08A - 1		15935	4.7020		-245.1	PPB	18 Dec 2017 16:18:45	U	SMPL
I6910-04LX5 - 1		-25	-0.0045		-245.1	PPB	18 Dec 2017 16:20:53	U	SMPL
I6910-04A - 1		13541	3.9961		-245.1	PPB	18 Dec 2017 16:23:01	U	SMPL
CCV77 - 1	CCV77	16092	4.7483		-245.1	PPB	18 Dec 2017 16:45:22	U	SMPL
CCB77 - 1	CCB77	-80	-0.0207		-245.1	PPB	18 Dec 2017 16:47:30	U	SMPL