



Analytical Summary Report

Analysis Method: 9045C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB92235

Temp.: 15.6 °C

SLOPE.: 97.5

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			3.99	12/18/2017	19:05
2	CAL2	1	Water			7.01	12/18/2017	19:06
3	CAL3	1	Water			10.01	12/18/2017	19:07
4	ICV	1	Water			7.00	12/18/2017	19:09
5	CCV1	1	Water			2.01	12/18/2017	19:10
6	I6983-01	1	Water			7.45	12/18/2017	19:11
7	I6983-01DUP	1	Water			7.46	12/18/2017	19:13
8	CCV2	1	Water			12.01	12/18/2017	19:15

LB 92235

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH I6983 WorkList ID : 106700 Date : 12/18/2017 7:07:36 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2017	Water	I6983-01-11	pH	Cool 4 deg C	DEWB01	F42	PHMHB-WCW	12/18/2017	9045C

Date/Time 12-18-17 7:07 PM
Received by: JD
Relinquished by: CD

Date/Time 12-18-17 8:00 PM
Received by: CD
Relinquished by: JD