



Analytical Summary Report

Analysis Method: 9045C

Analyst By : Jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB92236

Temp.: 24.2 °C

BalanceID: WC SC-4

SLOPE.: 98.8

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2270
buffer solution pH 7 yellow	W2159
PH 10.01 BUFFER,COLOR CD 475ML	W2154
buffer solution pH 7 yellow	W2219
pH 2.00 Buffer	W2289
Buffer Solution, PH12 (500ml)	W2267

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.01 Control Limits[+/- 0.1].

True Value of CCV2 = 2.01 Control Limits[+/- 0.1].

True Value of CCV3 = 12.01 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water			4.01	12/18/2017	19:20
2	CAL2	1	Water			7.01	12/18/2017	19:21
3	CAL3	1	Water			10.03	12/18/2017	19:22
4	ICV	1	Water			6.99	12/19/2017	19:24
5	CCV1	1	Water			2.01	12/19/2017	19:25
6	I6676-09	1	Solid	20.02	20	7.70	12/18/2017	19:26
7	I6676-09DUP	1	Solid	20.04	20	7.71	12/18/2017	19:27
8	I6682-02	1	Solid	20.01	20	8.50	12/18/2017	19:30
9	I6682-04	1	Solid	20.03	20	7.40	12/18/2017	19:31
10	I6682-06	1	Solid	20.05	20	6.96	12/18/2017	19:32
11	I6973-03	1	Solid	20.02	20	7.80	12/18/2017	19:34
12	I6975-02	1	Solid	20.01	20	7.39	12/18/2017	19:35
13	I6975-04	1	Solid	20.03	20	5.55	12/18/2017	19:36
14	I6981-13	1	Solid	20.04	20	9.62	12/18/2017	19:37
15	I6981-14	1	Solid	20.06	20	9.15	12/18/2017	19:38
16	CCV2	1	Water			2.01	12/18/2017	19:40
17	I6981-15	1	Solid	20.03	20	8.41	12/18/2017	19:41
18	I6981-16	1	Solid	20.03	20	8.61	12/18/2017	19:42
19	CCV3	1	Water			12.02	12/18/2017	19:45

1892236

WORKLIST(Hardcopy Internal Chain)

WorkList Name: CORROSIONITY I6975 WorkList ID: 106697 Date: 12/18/2017 4:09:15 PM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/18/2017	Solid	I6676-09	Corrosivity	Cool 4 deg C	PSEG05		OK-01-121817	12/18/2017	9045C
12/18/2017	Solid	I6682-02	Corrosivity	Cool 4 deg C	PSEG05	F52	NB-01-121817-A	12/18/2017	9045C
12/18/2017	Solid	I6682-04	Corrosivity	Cool 4 deg C	PSEG05	F52	NB-01-121817-B	12/18/2017	9045C
12/18/2017	Solid	I6682-06	Corrosivity	Cool 4 deg C	PSEG05	F52	NB-01-121817-C	12/18/2017	9045C
12/18/2017	Solid	I6973-03	Corrosivity	Cool 4 deg C	WSP001		B11(0-10)-COMP	12/18/2017	9045C
12/18/2017	Solid	I6975-02	Corrosivity	Cool 4 deg C	PSEG01	F52	72-12000	12/18/2017	9045C
12/18/2017	Solid	I6975-04	Corrosivity	Cool 4 deg C	PSEG01	F52	72-11989	12/18/2017	9045C
12/18/2017	Solid	I6981-13	Corrosivity	Cool 4 deg C	DEWB01	F51	PHMHA-WCS	12/18/2017	9045C
12/18/2017	Solid	I6981-14	Corrosivity	Cool 4 deg C	DEWB01	F51	PHMHA-WCD	12/18/2017	9045C
12/18/2017	Solid	I6981-15	Corrosivity	Cool 4 deg C	DEWB01	F51	PHMHB-WCS	12/18/2017	9045C
12/18/2017	Solid	I6981-16	Corrosivity	Cool 4 deg C	DEWB01	F51	PHMHB-WCD	12/18/2017	9045C

Date/Time 12-18-17 4:00 PM
Received by: JP
Relinquished by: JP

Date/Time 12-18-17 7:35 PM
Received by: JP
Relinquished by: JP