


7470A

LB 92241
INSTRUMENT ID: CY1

Linear

 μ Abs.:

34407



A= 0.0000e+000

B= 2.9132e-004

C= 3.5400e-002

Rho= 0.9997426

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.045	0.045	34	0.000	34				
0.2	0.200	0.203	0.003	575	0.0 %	575				
2.5	2.500	2.411	-0.089	8153	0.0 %	8153				
5.0	5.000	5.110	0.110	17420	0.0 %	17420				
7.5	7.500	7.372	-0.128	25184	0.0 %	25184				
10.0	10.000	10.059	0.059	34407	0.0 %	34407				

LB92241 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 19 Dec 2017 09:38:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	34	-	0.0000	7470A	PPB	19 Dec 2017 09:45:19	S	Std
0.2 - 1	50.2	575	-	0.2000	7470A	PPB	19 Dec 2017 09:47:27	S	Std
2.5 - 1	52.5	8153	-	2.5000	7470A	PPB	19 Dec 2017 09:49:36	S	Std
5.0 - 1	55	17420	-	5.0000	7470A	PPB	19 Dec 2017 09:51:44	S	Std
7.5 - 1	57.5	25184	-	7.5000	7470A	PPB	19 Dec 2017 09:53:52	S	Std
10.0 - 1	510	34407	-	10.0000	7470A	PPB	19 Dec 2017 09:56:00	S	Std
ICV15 - 1	ICV15	12998	3.8220	-	7470A	PPB	19 Dec 2017 09:58:49	U	SMPL
ICB15 - 1	ICB15	-40	0.0237	-	7470A	PPB	19 Dec 2017 10:00:55	U	SMPL
CCV78 - 1	CCV78	17421	5.1105	-	7470A	PPB	19 Dec 2017 10:03:02	U	SMPL
CCB78 - 1	CCB78	-102	0.0057	-	7470A	PPB	19 Dec 2017 10:05:08	U	SMPL
CRA - 1	CRA	579	0.2041	-	7470A	PPB	19 Dec 2017 10:07:16	U	SMPL
HighStd - 1	HighStd	34463	10.0752	-	7470A	PPB	19 Dec 2017 10:09:23	U	SMPL
ChkStd - 1	ChkStd	24790	7.2573	-	7470A	PPB	19 Dec 2017 10:11:30	U	SMPL
PB105141BL - 1	PBW	-151	-0.0086	-	7470A	PPB	19 Dec 2017 10:13:39	U	SMPL
PB105141BS - 1	LCSW	12900	3.7935	-	7470A	PPB	19 Dec 2017 10:15:48	U	SMPL
PB105141TB - 1	PB105085TB	-92	0.0086	-	7470A	PPB	19 Dec 2017 10:17:55	U	SMPL
I6675-08 - 1	SU-04-121517	5	0.0369	-	7470A	PPB	19 Dec 2017 10:20:04	U	SMPL
I6677-06 - 1	OR-01-121517-A	35	0.0456	-	7470A	PPB	19 Dec 2017 10:22:12	U	SMPL
I6677-08 - 1	OR-01-121517-B	28	0.0436	-	7470A	PPB	19 Dec 2017 10:24:20	U	SMPL
I6677-08DUP - 1	OR-01-121517-BDUP	24	0.0424	-	7470A	PPB	19 Dec 2017 10:26:27	U	SMPL
CCV79 - 1	CCV79	17451	5.1193	-	7470A	PPB	19 Dec 2017 10:28:33	U	SMPL
CCB79 - 1	CCB79	-88	0.0098	-	7470A	PPB	19 Dec 2017 10:30:40	U	SMPL
I6677-08MS - 1	OR-01-121517-BMS	12986	3.8185	-	7470A	PPB	19 Dec 2017 10:32:48	U	SMPL
I6677-08MSD - 1	OR-01-121517-BMSD	12989	3.8194	-	7470A	PPB	19 Dec 2017 10:34:54	U	SMPL
I6680-20 - 1	PATERSON-02-121517	-53	0.0200	-	7470A	PPB	19 Dec 2017 10:37:01	U	SMPL
I6927-03 - 1	R-4297-39456	29	0.0438	-	7470A	PPB	19 Dec 2017 10:39:08	U	SMPL
I6939-01 - 1	RO-202973	20	0.0412	-	7470A	PPB	19 Dec 2017 10:41:15	U	SMPL
I6939-02 - 1	RO-202646	23	0.0421	-	7470A	PPB	19 Dec 2017 10:43:23	U	SMPL
I6940-03 - 1	VB16206-VB16211	18	0.0406	-	7470A	PPB	19 Dec 2017 10:45:32	U	SMPL
I6957-05 - 1	HALSTED-NW	69	0.0555	-	7470A	PPB	19 Dec 2017 10:47:40	U	SMPL
I6957-06 - 1	HALSTED-SW	21	0.0415	-	7470A	PPB	19 Dec 2017 10:49:48	U	SMPL
I6957-07 - 1	HALSTED-EW	65	0.0543	-	7470A	PPB	19 Dec 2017 10:51:55	U	SMPL
CCV80 - 1	CCV80	17317	5.0802	-	7470A	PPB	19 Dec 2017 10:54:02	U	SMPL
CCB80 - 1	CCB80	-67	0.0159	-	7470A	PPB	19 Dec 2017 10:56:09	U	SMPL
I6957-08 - 1	HALSTED-WW	47	0.0491	-	7470A	PPB	19 Dec 2017 10:58:17	U	SMPL
I6958-02 - 1	DRUM-1-4	10	0.0383	-	7470A	PPB	19 Dec 2017 11:00:24	U	SMPL
I6677-08LX5 - 1		14	0.0395	-	7470A	PPB	19 Dec 2017 11:02:31	U	SMPL
I6677-08A - 1		13532	3.9776	-	7470A	PPB	19 Dec 2017 11:04:38	U	SMPL
CCV81 - 1	CCV81	17479	5.1274	-	7470A	PPB	19 Dec 2017 11:06:45	U	SMPL
CCB81 - 1	CCB81	-82	0.0115	-	7470A	PPB	19 Dec 2017 11:08:53	U	SMPL