


ISM02.4

LB 92244
INSTRUMENT ID: CV1

Linear

 μ Abs.:

41362



A= 0.0000e+000

B= 2.4314e-004

C= 9.3301e-004

Rho= 0.9999194

Accept=Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/b
0.00	0.000	0.003	0.003	9	0.000	9					1
0.2	0.200	0.204	0.004	835	0.0 %	835					2
2.5	2.500	2.510	0.010	10321	0.0 %	10321					0
5.0	5.000	5.021	0.021	20648	0.0 %	20648					0
7.5	7.500	7.404	-0.096	30447	0.0 %	30447					-1
10.0	10.000	10.058	0.058	41362	0.0 %	41362					1

NB
12/19/18

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	9	-		0 ISM02.4	PPB	12/19/2017 11:44	Std
	0.2 S0.2	835	-		0.2 ISM02.4	PPB	12/19/2017 11:46	Std
	2.5 S2.5	10321	-		2.5 ISM02.4	PPB	12/19/2017 11:49	Std
	5 S5	20648	-		5 ISM02.4	PPB	12/19/2017 11:51	Std
	7.5 S7.5	30447	-		7.5 ISM02.4	PPB	12/19/2017 11:53	Std
	10 S10	41362	-		10 ISM02.4	PPB	12/19/2017 11:55	Std
ICV	ICV	15407	3.747 -		ISM02.4	PPB	12/19/2017 11:58	SMPL
ICB	ICB	-82	-0.019 -		ISM02.4	PPB	12/19/2017 12:00	SMPL
CCV082	CCV082	20489	4.9826 -		ISM02.4	PPB	12/19/2017 12:03	SMPL
CCB082	CCB082	-95	-0.0222 -		ISM02.4	PPB	12/19/2017 12:05	SMPL
PB105130BL	PBW003	0	0.0009 -		ISM02.4	PPB	12/19/2017 12:07	SMPL
I6913-01	MBDYX0	30	0.0082 -		ISM02.4	PPB	12/19/2017 12:09	SMPL
I6913-02	MBDYX1	396	0.0972 -		ISM02.4	PPB	12/19/2017 12:12	SMPL
I6913-03	MBDYX7	8	0.0029 -		ISM02.4	PPB	12/19/2017 12:14	SMPL
I6913-04	MBDYX7D	30	0.0082 -		ISM02.4	PPB	12/19/2017 12:16	SMPL
I6913-05	MBDYX7S	4269	1.0389 -		ISM02.4	PPB	12/19/2017 12:19	SMPL
I6913-06	MBDYX8	6	0.0024 -		ISM02.4	PPB	12/19/2017 12:21	SMPL
I6913-07	MBDYY9	20	0.0058 -		ISM02.4	PPB	12/19/2017 12:23	SMPL
I6913-08	MBDYZ7	23	0.0065 -		ISM02.4	PPB	12/19/2017 12:25	SMPL
I6913-09	MBDYX2	17	0.0051 -		ISM02.4	PPB	12/19/2017 12:28	SMPL
I6913-10	MBDYX6	89	0.0226 -		ISM02.4	PPB	12/19/2017 12:30	SMPL
I6913-11	MBDYX4	60	0.0155 -		ISM02.4	PPB	12/19/2017 12:32	SMPL
I6913-12	MBDYT7	66	0.017 -		ISM02.4	PPB	12/19/2017 12:35	SMPL
I6913-13	MBDYW7	78	0.0199 -		ISM02.4	PPB	12/19/2017 12:37	SMPL
I6913-14	MBDYW8	39	0.0104 -		ISM02.4	PPB	12/19/2017 12:39	SMPL
I6913-17	MBDYX3	158	0.0393 -		ISM02.4	PPB	12/19/2017 12:41	SMPL
I6913-18	MBDYX5	120	0.0301 -		ISM02.4	PPB	12/19/2017 12:44	SMPL
I6913-19	MBDYZ8	22	0.0063 -		ISM02.4	PPB	12/19/2017 12:46	SMPL
I6913-20	MBDYT3	50	0.0131 -		ISM02.4	PPB	12/19/2017 12:48	SMPL
I6913-21	MBDYW4	43	0.0114 -		ISM02.4	PPB	12/19/2017 12:50	SMPL
CCV083	CCV083	20402	4.9614 -		ISM02.4	PPB	12/19/2017 12:53	SMPL
CCB083	CCB083	-76	-0.0175 -		ISM02.4	PPB	12/19/2017 12:55	SMPL
PB105132BL	PBW003	30	0.0082 -		ISM02.4	PPB	12/19/2017 12:57	SMPL
I6932-01	MCOAT2	26	0.0073 -		ISM02.4	PPB	12/19/2017 13:00	SMPL
I6932-02	MCOAT3	33	0.009 -		ISM02.4	PPB	12/19/2017 13:02	SMPL
I6932-03	MCOAT5	31	0.0085 -		ISM02.4	PPB	12/19/2017 13:04	SMPL
I6932-04	MCOAT5D	24	0.0068 -		ISM02.4	PPB	12/19/2017 13:06	SMPL
I6932-05	MCOAT5S	4076	0.992 -		ISM02.4	PPB	12/19/2017 13:09	SMPL
CCV084	CCV084	20352	4.9493 -		ISM02.4	PPB	12/19/2017 13:11	SMPL
CCB084	CCB084	-81	-0.0188 -		ISM02.4	PPB	12/19/2017 13:13	SMPL