



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB92245

Ambient Barometric Pressure (mmHg): 760

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	Dilution	Celsius °C	Result °F	Anal Date	Anal Time
1	ICV	81	1	27.0	80.6	12/19/2017	09:25
2	I6983-01		1	>100.0	>212.0	12/19/2017	09:50
3	I6983-01DUP		1	>100.0	>212.0	12/19/2017	10:15

Result = (Celsius * 1.8) + 32

LB92245

WORKLIST(Hardcopy Internal Chain)

WorkList Name : FP121719

WorkList ID : 106721

Date : 12/19/2017 8:36:40 AM

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
12/30/2017	Water	I6983-01	Flash Point	Cool 4 deg C	DEWB01	F42	PHMHB-WCW	12/18/2017	1010A

Date/Time 12/19/17 8:45 AM
Received by: KS
Relinquished by: CP

Date/Time 12/19/17 10:25 AM
Received by: CP
Relinquished by: KS