


7471B

LB 92250
INSTRUMENT ID: CV1

Linear

 μ Abs.:

42611



A= 0.0000e+000

B= 2.3568e-004

C= 2.3375e-002

Rho= 0.9994039

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	32	0.000	32				
0.2	0.200	0.189	-0.011	704	0.0 %	704	0			
2.5	2.500	2.420	-0.080	10168	0.0 %	10168				
5.0	5.000	5.204	0.204	21981	0.0 %	21981				
7.5	7.500	7.290	-0.210	30834	0.0 %	30834				
10.0	10.000	10.066	0.066	42611	0.0 %	42611				

LB92250 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Dec 2017 13:46:54

Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	32	-	0.0000	7471B	PPB	19 Dec 2017 14:20:10	S	Std
0.2 - 1	50.2	704	-	0.2000	7471B	PPB	19 Dec 2017 14:22:28	S	Std
2.5 - 1	52.5	10168	-	2.5000	7471B	PPB	19 Dec 2017 14:24:46	S	Std
5.0 - 1	55	21981	-	5.0000	7471B	PPB	19 Dec 2017 14:27:03	S	Std
7.5 - 1	57.5	30834	-	7.5000	7471B	PPB	19 Dec 2017 14:29:20	S	Std
10.0 - 1	510	42611	-	10.0000	7471B	PPB	19 Dec 2017 14:37:06	S	Std
ICV16 - 1	ICV16	16319	3.8694	-	7471B	PPB	19 Dec 2017 14:39:45	U	SMPL
ICB16 - 1	ICB16	-111	-0.0028	-	7471B	PPB	19 Dec 2017 14:42:00	U	SMPL
CCV85 - 1	CCV85	21322	5.0485	-	7471B	PPB	19 Dec 2017 14:44:16	U	SMPL
CCB85 - 1	CCB85	-107	-0.0018	-	7471B	PPB	19 Dec 2017 14:46:32	U	SMPL
CRA - 1	CRA	681	0.1839	-	7471B	PPB	19 Dec 2017 14:48:48	U	SMPL
HighStd - 1	HighStd	43187	10.2016	-	7471B	PPB	19 Dec 2017 14:51:04	U	SMPL
ChkStd - 1	ChkStd	28898	6.8340	-	7471B	PPB	19 Dec 2017 14:53:21	U	SMPL
PB105146BL - 1	PBS	-164	-0.0153	-	7471B	PPB	19 Dec 2017 14:55:39	U	SMPL
PB105146BS - 1	LCSS	15053	3.5711	-	7471B	PPB	19 Dec 2017 14:57:57	U	SMPL
6680-19 - 1	PATERSON-02-121517	6006	1.4389	-	7471B	PPB	19 Dec 2017 15:00:14	U	SMPL
6920-06 - 1	HL-8C	1311	0.3324	-	7471B	PPB	19 Dec 2017 15:02:31	U	SMPL
6920-09 - 1	HL-7C	330	0.1011	-	7471B	PPB	19 Dec 2017 15:04:49	U	SMPL
6920-18 - 1	HL-5C	16068	3.8103	-	7471B	PPB	19 Dec 2017 15:07:07	U	SMPL
6924-01 - 1	EPE-125-4(0-5)	8385	1.9995	-	7471B	PPB	19 Dec 2017 15:09:22	U	SMPL
CCV86 - 1	CCV86	21294	5.0419	-	7471B	PPB	19 Dec 2017 15:11:38	U	SMPL
CCB86 - 1	CCB86	-104	-0.0011	-	7471B	PPB	19 Dec 2017 15:13:54	U	SMPL
6924-02 - 1	EPE-125-4(5-9)	1951	0.4832	-	7471B	PPB	19 Dec 2017 15:16:10	U	SMPL
6927-01 - 1	R-4297-39456	9763	2.3243	-	7471B	PPB	19 Dec 2017 15:18:25	U	SMPL
6940-01 - 1	VB16206-VB16211	5220	1.2536	-	7471B	PPB	19 Dec 2017 15:20:41	U	SMPL
6949-02 - 1	SSI-4(5-6)	1050	0.2708	-	7471B	PPB	19 Dec 2017 15:22:58	U	SMPL
6949-04 - 1	SSI-3(7-8)	3340	0.8105	-	7471B	PPB	19 Dec 2017 15:25:14	U	SMPL
6949-09 - 1	SSI-5(6-7)	1178	0.3010	-	7471B	PPB	19 Dec 2017 15:27:31	U	SMPL
6949-09DUP - 1	SSI-5(6-7)DUP	1180	0.3015	-	7471B	PPB	19 Dec 2017 15:29:48	U	SMPL
6949-10 - 1	SSI-5(6-7)MS	17092	4.0516	-	7471B	PPB	19 Dec 2017 15:32:06	U	SMPL
6949-11 - 1	SSI-5(6-7)MSD	16716	3.9630	-	7471B	PPB	19 Dec 2017 15:34:24	U	SMPL
6949-16 - 1	SSI-7(13.5-14.5)	713	0.1914	-	7471B	PPB	19 Dec 2017 15:36:40	U	SMPL
CCV87 - 1	CCV87	21382	5.0627	-	7471B	PPB	19 Dec 2017 15:38:56	U	SMPL
CCB87 - 1	CCB87	-103	-0.0009	-	7471B	PPB	19 Dec 2017 15:41:13	U	SMPL
6949-17 - 1	SSI-2(8-9)	151	0.0590	-	7471B	PPB	19 Dec 2017 15:43:29	U	SMPL
6949-21 - 1	SSI-FD02	997	0.2583	-	7471B	PPB	19 Dec 2017 15:45:46	U	SMPL
6957-01 - 1	HALSTED-NW	3402	0.8252	-	7471B	PPB	19 Dec 2017 15:48:02	U	SMPL
6957-02 - 1	HALSTED-SW	20666	4.8939	-	7471B	PPB	19 Dec 2017 15:50:18	U	SMPL
6957-03 - 1	HALSTED-EW	12649	3.0045	-	7471B	PPB	19 Dec 2017 15:52:35	U	SMPL
6957-04 - 1	HALSTED-WW	21074	4.9901	-	7471B	PPB	19 Dec 2017 15:54:52	U	SMPL
6958-01 - 1	DRUM-1-4	85290	20.1244	-	7471B	PPB	19 Dec 2017 15:57:09	U	SMPL
6973-02 - 1	B11(0-10)-COMP	25279	5.9811	-	7471B	PPB	19 Dec 2017 15:59:27	U	SMPL
PB105147BL - 1	PBS	-171	-0.0169	-	7471B	PPB	19 Dec 2017 16:01:48	U	SMPL
PB105147BS - 1	LCSS	15536	3.6849	-	7471B	PPB	19 Dec 2017 16:04:05	U	SMPL
CCV88 - 1	CCV88	22153	5.2444	-	7471B	PPB	19 Dec 2017 16:06:21	U	SMPL
CCB88 - 1	CCB88	-112	-0.0030	-	7471B	PPB	19 Dec 2017 16:08:38	U	SMPL
6962-01 - 1	SSI-8(9-10)	961	0.2499	-	7471B	PPB	19 Dec 2017 16:10:55	U	SMPL
6962-02 - 1	SSI-9(7.5-8.5)	681	0.1839	-	7471B	PPB	19 Dec 2017 16:13:12	U	SMPL
6966-01 - 1	DB-WW-3-0-4	9114	2.1714	-	7471B	PPB	19 Dec 2017 16:15:28	U	SMPL
6966-02 - 1	DB-WW-3-4-10	4271	1.0300	-	7471B	PPB	19 Dec 2017 16:17:44	U	SMPL
6966-02DUP - 1	DB-WW-3-4-10DUP	4743	1.1412	-	7471B	PPB	19 Dec 2017 16:20:01	U	SMPL
6966-03 - 1	I6966-02MS	20218	4.7883	-	7471B	PPB	19 Dec 2017 16:22:18	U	SMPL
6966-04 - 1	I6966-02MSD	19490	4.6168	-	7471B	PPB	19 Dec 2017 16:24:35	U	SMPL
6966-05 - 1	DB-EW-3-0-4	5955	1.4268	-	7471B	PPB	19 Dec 2017 16:26:53	U	SMPL
6966-06 - 1	DB-EW-3-4-10	14139	3.3556	-	7471B	PPB	19 Dec 2017 16:29:11	U	SMPL
6970-01 - 1	S1-0-5	384	0.1139	-	7471B	PPB	19 Dec 2017 16:31:27	U	SMPL
CCV89 - 1	CCV89	21979	5.2034	-	7471B	PPB	19 Dec 2017 16:33:43	U	SMPL
CCB89 - 1	CCB89	-106	-0.0016	-	7471B	PPB	19 Dec 2017 16:36:00	U	SMPL
6970-02 - 1	S2-0-5	1817	0.4516	-	7471B	PPB	19 Dec 2017 16:38:17	U	SMPL
6970-03 - 1	S3-0-5	1196	0.3052	-	7471B	PPB	19 Dec 2017 16:40:33	U	SMPL
6970-04 - 1	S4-0-5	1363	0.3446	-	7471B	PPB	19 Dec 2017 16:42:50	U	SMPL
6970-05 - 1	S5-0-5	1172	0.2996	-	7471B	PPB	19 Dec 2017 16:45:06	U	SMPL
6970-06 - 1	S6-0-5	1433	0.3611	-	7471B	PPB	19 Dec 2017 16:47:23	U	SMPL
6970-07 - 1	S1-10-15	918	0.2397	-	7471B	PPB	19 Dec 2017 16:49:40	U	SMPL
6970-08 - 1	S2-5-10	999	0.2588	-	7471B	PPB	19 Dec 2017 16:51:58	U	SMPL
6970-09 - 1	S3-5-10	1155	0.2956	-	7471B	PPB	19 Dec 2017 16:54:15	U	SMPL
6970-10 - 1	S4-5-10	1283	0.3258	-	7471B	PPB	19 Dec 2017 16:56:33	U	SMPL
6970-11 - 1	S5-5-10	313	0.0971	-	7471B	PPB	19 Dec 2017 16:58:50	U	SMPL
CV90 - 1	CCV90	22157	5.2453	-	7471B	PPB	19 Dec 2017 17:01:07	U	SMPL
CB90 - 1	CCB90	-101	-0.0004	-	7471B	PPB	19 Dec 2017 17:03:23	U	SMPL
6970-12 - 1	S6-5-10	5460	1.3102	-	7471B	PPB	19 Dec 2017 17:05:40	U	SMPL
6975-01 - 1	72-12000	2900	0.7068	-	7471B	PPB	19 Dec 2017 17:07:57	U	SMPL
6975-03 - 1	72-11989	496	0.1403	-	7471B	PPB	19 Dec 2017 17:10:13	U	SMPL
B105148BL - 1	PBS	8	0.0253	-	7471B	PPB	19 Dec 2017 17:12:30	U	SMPL
B105148BS - 1	LCSS	14666	3.4798	-	7471B	PPB	19 Dec 2017 17:14:47	U	SMPL
675-07 - 1	SU-04-121517	628	0.1714	-	7471B	PPB	19 Dec 2017 17:17:04	U	SMPL
677-05 - 1	OR-01-121517-A	2955	0.7198	-	7471B	PPB	19 Dec 2017 17:19:21	U	SMPL
677-07 - 1	OR-01-121517-B	2273	0.5591	-	7471B	PPB	19 Dec 2017 17:21:39	U	SMPL
961-01 - 1	EPE-108-1(15-17)	878	0.2303	-	7471B	PPB	19 Dec 2017 17:23:57	U	SMPL
961-02 - 1	EPE-108-1(20-22)	665	0.1801	-	7471B	PPB	19 Dec 2017 17:26:13	U	SMPL
CV91 - 1	CCV91	19822	4.6950	-	7471B	PPB	19 Dec 2017 17:28:31	U	SMPL

LB92250 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 19 Dec 2017 13:46:54

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
CCB91 - 1	CCB91	-97	0.0005		-7471B	PPB	19 Dec 2017 17:30:48	U	SMPL
I6961-03 - 1	EPE-106-3DS(0-5)	25567	6.0490		-7471B	PPB	19 Dec 2017 17:33:05	U	SMPL
I6961-04 - 1	EPE-108-1(50-52)	-46	0.0125		-7471B	PPB	19 Dec 2017 17:35:22	U	SMPL
I6961-05 - 1	EPE-108-1(60-62)	132	0.0545		-7471B	PPB	19 Dec 2017 17:37:40	U	SMPL
I6961-06 - 1	EPE-108-1(65-67)	1366	0.3453		-7471B	PPB	19 Dec 2017 17:39:57	U	SMPL
I6961-07 - 1	EPE-108-1(70-72)	592	0.1629		-7471B	PPB	19 Dec 2017 17:42:14	U	SMPL
I6961-07DUP - 1	EPE-108-1(70-72)DUP	639	0.1740		-7471B	PPB	19 Dec 2017 17:44:31	U	SMPL
I6961-07MS - 1	EPE-108-1(70-72)MS	16376	3.8829		-7471B	PPB	19 Dec 2017 17:46:49	U	SMPL
I6961-07MSD - 1	EPE-108-1(70-72)MSD	16269	3.8576		-7471B	PPB	19 Dec 2017 17:49:06	U	SMPL
I6961-08 - 1	EPE-108-1(75-77)	470	0.1341		-7471B	PPB	19 Dec 2017 17:51:24	U	SMPL
CCV92 - 1	CCV92	21715	5.1411		-7471B	PPB	19 Dec 2017 17:53:41	U	SMPL
CCB92 - 1	CCB92	-101	-0.0004		-7471B	PPB	19 Dec 2017 17:55:58	U	SMPL
I6949-09LX5 - 1		220	0.0752		-7471B	PPB	19 Dec 2017 17:58:15	U	SMPL
I6949-09A - 1		17969	4.2583		-7471B	PPB	19 Dec 2017 18:00:32	U	SMPL
I6966-02LX5 - 1		765	0.2037		-7471B	PPB	19 Dec 2017 18:02:49	U	SMPL
I6966-02A - 1		20935	4.9573		-7471B	PPB	19 Dec 2017 18:05:06	U	SMPL
I6961-07LX5 - 1		37	0.0321		-7471B	PPB	19 Dec 2017 18:07:23	U	SMPL
I6961-07A - 1		16487	3.9090		-7471B	PPB	19 Dec 2017 18:09:41	U	SMPL
I6958-01DLX5 - 1		17059	4.0438		-7471B	PPB	19 Dec 2017 18:17:52	U	SMPL
CCV93 - 1	CCV93	20806	4.9269		-7471B	PPB	19 Dec 2017 18:20:09	U	SMPL
CCB93 - 1	CCB93	-57	0.0099		-7471B	PPB	19 Dec 2017 18:22:26	U	SMPL